Differences in hemodynamic stability between hemodialysis and hemofiltration: identification of possible pathogenetic factors

Citation for published version (APA):

van Kuijk, W. H. M. (1995). *Differences in hemodynamic stability between hemodialysis and hemofiltration: identification of possible pathogenetic factors*. [Doctoral Thesis, Maastricht University]. Datawyse / Universitaire Pers Maastricht. <https://doi.org/10.26481/dis.19951214wk>

Document status and date:

Published: 01/01/1995

DOI:

[10.26481/dis.19951214wk](https://doi.org/10.26481/dis.19951214wk)

Document Version:

Publisher's PDF, also known as Version of record

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.umlib.nl/taverne-license

Take down policy

If you believe that this document breaches copyright please contact us at:

repository@maastrichtuniversity.nl

providing details and we will investigate your claim.

Differences in hemodynamic stability
between hemodialysis and hemofiltration:

identification of possible pathogenetic factors

© WHM van Kuijk, Maastricht 1995

CIP-DATA KONINKLIJKE BIBLIOTHEEK, DEN HAAG

Kuijk, Willibrordus Hendrikus Maria van

Differences in hemodynamic stability between hemodialysis and hemofiltration : identification of possible pathogenetic factors / Willibrordus Hendrikus Maria van Kuijk. - Maastricht : Universitaire Pers Maastricht. - III. Thesis Rijksuniversiteit Limburg Maastricht. - With ref. - With summary in Dutch.
ISBN 90-5278-208-3
Subject headings: hemodialysis / hemofiltration.

The studies in this thesis as well as printing of this thesis were made possible by financial support of Gambro AB, Sweden

**Differences in hemodynamic stability
between hemodialysis and hemofiltration:
identification of possible pathogenetic factors**

PROEFSCHRIFT

ter verkrijging van de graad van doctor
aan de Rijksuniversiteit Limburg te Maastricht,
op gezag van de Rector Magnificus, Prof. Mr. M.J. Cohen,
volgens het besluit van het College van Dekanen,
in het openbaar te verdedigen
op donderdag 14 december 1995 om 14.00 uur

door

Willibrordus Hendrikus Maria van Kuijk

geboren op 16 november 1965 te Waalwijk

Promotor

Prof. dr. J.P. van Hooff

Co-promotor

Dr. K.M.L. Leunissen

Beoordelingscommissie

Prof. dr. P.J.E.H.M. Kitslaar, voorzitter

Prof. dr. C. Baldamus (University of Cologne, Germany)

Prof. dr. J. Bergström (Karolinska Institute, Stockholm, Sweden)

Mw. Prof. dr. M.P. van Diejen-Visser

Prof. dr. H.A.J. Struijker Boudier

Voor mijn ouders

Contents

Chapter 1:	Introduction	9
Chapter 2:	Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences	33
Chapter 3:	Vascular reactivity during combined ultrafiltration-hemodialysis: influence of dialysate sodium	53
Chapter 4:	Vascular reactivity during combined ultrafiltration-hemodialysis: influence of dialysate derived contaminants	67
Chapter 5:	Influence of changes in ionized calcium on cardiovascular reactivity during hemodialysis	81
Chapter 6:	Critical role of the extracorporeal blood temperature in the hemodynamic response during hemofiltration	95
Chapter 7:	Summary and concluding remarks	109
Chapter 8:	Samenvatting	115
	Dankwoord	121
	Curriculum vitae	125

Chapter 1

Introduction

WHM van Kuijk and KML Leunissen

(part of this chapter will be published in Blood Purification)

General Introduction

One of the main goals of dialysis therapy in patients with end-stage renal failure, is removal of excess fluid that has accumulated in the inter-dialytic period. In this respect hemodynamic instability with symptomatic hypotension will remain one of the most important complications in dialysis therapy especially as the number of elderly, cardiovascular compromised patients on dialysis is still increasing. Although recent data are lacking, the incidence of hypotensive episodes is still reported to be around 20%. Symptomatic hypotension is not only a burden for the patient, it also implicates expensive interventions leading to less effective treatment. Extra treatments are even necessary for certain patients in order to achieve dry-weight. Therefore, it is of the utmost importance to optimize dialysis therapy with respect to blood pressure stability.

Already in the late seventies, Bergström and Quellhorst showed that blood pressure was better maintained when fluid was withdrawn during either isolated ultrafiltration or hemofiltration as compared to combined ultrafiltration-hemodialysis¹⁻³. These observations have been confirmed in clinical trials comparing hemofiltration with acetate buffered combined ultrafiltration-hemodialysis^{4,5}. As compared to bicarbonate buffered combined ultrafiltration-hemodialysis, the hemodynamic advantage of hemofiltration is probably smaller^{6,7}. Identifying factors responsible for these differences in hemodynamic instability is therefore of importance since this provides means for further optimisation of dialysis therapy and fluid removal tolerance. Theoretically, differences in hemodynamic stability could be based on differences in **plasma volume preservation** and **cardiovascular reactivity** or both. These mechanisms are playing a central role in blood pressure regulation during dialysis therapy. The pathophysiology of both mechanisms will now be discussed especially in relation to different dialysis strategies.

Dialysis strategies

The dialysis strategies discussed and used in the present thesis are combined ultrafiltration-hemodialysis (UF+HD), isolated ultrafiltration (i-UF) and hemofiltration (HF). In UF+HD, which is the most widely used form of dialysis therapy, diffusion is the main driving force for molecular transport. Diffusion takes place across a membrane with a predefined cut-off point. The extracorporeal circuit consists of an afferent (arterial) and afferent (venous) line with countercurrent flow of blood and dialysate through a dialyzer. Controlled ultrafiltration is obtained by inducing a net negative pressure across the dialysis membrane towards the dialysate

compartment. With a high small molecular clearance, clearance of middle molecules is relatively low in UF+HD. In i-UF the same extracorporeal circuit is used as in UF+HD. However, in the absence of dialysate, molecular transport and consequently clearance is negligible in i-UF. The ultrafiltration volume is limited to a maximum of 2 liters per hour. In HF, convection is the main driving force for molecular transport. In the absence of dialysate, clearance results from ultrafiltrating large amounts of plasmawater across a membrane (10-30 liter per hour). In addition, a sterile substitution fluid, resembling dialysate, is infused either before (pre-dilution) or after (post-dilution) the hemofilter. The rate of ultrafiltration and substitution are balanced. A main advantage of HF is the larger clearance of middle molecules. Dry-weight is defined as the post-dialytic weight at which the patient has an optimal fluid status.

Plasma Volume Preservation

Pathophysiology

Since fluid is directly withdrawn from the intravascular compartment during dialysis, plasma volume (PV) initially declines and as a result, the intracapillary hydrostatic pressure decreases while the colloid-osmotic pressure (COP) increases. These changes in the capillary Starling-equilibrium induce a refill of plasma volume at the capillary level from the extravascular spaces, i.e. the interstitial and intracellular compartments⁸.

Table 1.1 Factors influencing plasma volume preservation

Positive	Negative
optimal fluid status	hypovolemia
high venous compliance	low venous compliance
high-sodium dialysis	rapid fall in osmolality
bicarbonate dialysis	acetate dialysis
moderate ultrafiltration rate	high ultrafiltration rate

The process of plasma volume preservation (PVP) is influenced by a number of factors^{9,10} (table 1.1). First of all, the patient's fluid status is important. Refill is better with well hydrated extravascular spaces, a phenomenon that is related to the finite compliance of the interstitial compartment. Overhydration leads to an increase in interstitial pressure thereby improving refill¹¹. It has been shown that ultrafiltration below dry-weight resulting in depletion of the extravascular spaces

leads to an exponential decline in plasma volume¹². Therefore, adequate estimation of dry-weight is of the utmost importance¹³. PVP is also influenced by changes in plasma osmolality. Plasma osmolality decreases during dialysis as a result of hypertonic fluid loss, thus impairing refill. The decrease in osmolality can be partially compensated with higher dialysate sodium concentrations (>140 mmol/l)¹⁴. It has been shown that at dialysate sodium concentrations above 145 mmol/l, refill of the intravascular compartment is improved by inducing a transcellular fluid shift to the extracellular compartment¹⁵. These results have been confirmed in a more recent study using conductivity techniques¹⁶. Another factor influencing PVP concerns the used buffer substrate. Acetate impairs PVP at the capillary level by inducing arteriolar vasodilation thus increasing the intracapillary hydrostatic pressure. Bicarbonate has been shown to be more favourable in this respect, especially in patients with compromised left ventricular function^{17,18}.

Recently an inverse relation ($r = -0.66$; $p < 0.025$) between PVP and venous compliance was found by Kooman et al.. The relation is explained by a higher postcapillary pressure for a given level of plasma volume when venous compliance is reduced. As a result, hydrostatic pressure over the capillary wall will be higher and refill from the interstitium will be impaired¹⁹.

It is obvious that the decline in plasma volume is the resultant of (dis)balance between net ultrafiltration and PVP. Plasma volume declines progressively at ultrafiltration rates considerably exceeding the rate of PVP^{11,20}.

Differences between dialysis treatments

According to the first publications on better hemodynamic stability with i-UF, the apparent advantage of i-UF over UF+HD was attributed to differences in PVP that would arise from differences in plasma osmolality²¹. However, Rodriguez et al. found no significant differences in the decrement in PV (serial hematocrit) and the increase in COP, when comparing two hours i-UF with two hours UF+HD (sodium 135 mmol/l)⁸. Fleming et al. measured vascular refilling in ten patients during 30 minutes i-UF and 30 minutes UF+HD (acetate, sodium 140 mmol/l), sequentially performed within one dialysis session at an ultrafiltration rate of 1.2 l/hr. Changes in blood volume were measured with an optical monitor (infrared light absorption) that was clipped on the arterial line. After i-UF, the ultrafiltrate was reinfused, followed by an equilibration period of ± 30 minutes to achieve a comparable fluid status at the start of UF+HD. They also found no differences in vascular refilling between the two treatment modalities although confirming the better tolerance to a decline in blood volume during i-UF²².

Differences in PVP, however, do seem to play a role with regard to differences in blood pressure stability between UF+HD and HF. De Vries et al. have found a smaller decrease in blood volume (BV) (-1.69 vs -2.25% /hr, serial erythrocyte

count) together with a more pronounced fluid shift from the intra to the extra-cellular compartment during HF as compared to UF+HD (bicarbonate) (conductivity measurements). The ultrafiltration rate and dialysate sodium in both treatments were profiled in a similar way; first two hours: $\frac{2}{3}$ of total UF-volume with sodium 138 mmol/l, third and fourth hour: $\frac{1}{3}$ of total UF-volume with sodium 146 mmol/l. The authors explained the difference by a reduced clearance of urea in HF with a concomitant smaller decrease in plasma osmolality²³. Olthof et al. have confirmed these results²⁴.

Cardiovascular Reactivity

Pathophysiology

Due to disbalance in ultrafiltration and PVP, plasma volume normally declines during dialysis therapy and without adequate cardiovascular compensatory mechanisms, blood pressure would decrease. When intravascular volume declines under normal physiologic conditions, arterial blood pressure is maintained by an increase in either total peripheral resistance (TPR) or cardiac output.

A decline in intravascular volume is primarily sensed by low pressure receptors, located in the atria and pulmonary veins. A decrease in arterial blood pressure is sensed by baroreceptors which are mainly located in the carotid sinuses and in the wall of the aortic arch. In the resting state both sets of receptors tonically inhibit the vasomotor centre, located in the medullary part of the autonomic nervous system. Deactivation of either one or both sets of receptors results in a decrease in their inhibitory output which will lead to an increase in sympathetic output from the vasomotor centre to the cardiovascular system. As a result, heart rate and myocardial contractility both increase while splanchnic and limb blood flow decrease. In addition, plasma renin activity increases^{25,26}.

It has been shown in both human and animal studies that as compared to changes in heart rate and myocardial contractility, changes in vascular tone are far more important in maintaining arterial blood pressure during experimental hypovolemia²⁷⁻²⁹. Cardiac factors determine the limits of cardiac function. The actual output and consequently the arterial blood pressure are mainly regulated by the peripheral circulation³⁰.

Changes in TPR are especially important in maintaining arterial blood pressure at the pre-capillary level. However, according to Guyton, changes in arterial blood pressure are particularly buffered by changes in cardiac output, thereby pointing at the main determinant of cardiac output: cardiac filling³⁰. Right ventricular filling is mainly regulated by changes in venous tone. The venous system contains with minimal sympathetic activity a large blood volume reserve of hemodynamically

inactive or 'unstressed' blood which accounts for 60% of the total blood volume. When blood volume declines, 'unstressed' blood can become hemodynamically active by active, sympathetically mediated venous constriction. In addition, 'unstressed' blood is also temporarily mobilised by passive venous constriction. Passive recoil of the veins occurs when, as a result of an increase in pre-capillary resistance, the transmural venous distending pressure decreases. This is also known as the deJager-Kroch phenomenon^{31,32}. Especially in dialysis patients, adequate right ventricular filling is of utmost importance since diastolic ventricular compliance is often reduced in uremic patients with an incidence of left ventricular hypertrophy of about 70% in non-diabetics³³.

Differences between dialysis strategies

Baldamus et al. studied the acute hemodynamic (thermodilution) and sympathetic response in two groups of eight stable hemodialysis patients during i-UF, HF (acetate) and UF+HD with both acetate and bicarbonate buffering. Each patient served as his or her own control and was studied at the same individual pre-treatment fluid status with a sodium concentration of 140 mmol/l in both dialysate and substitution fluid. To compensate for a higher ultrafiltrable fraction of total calcium in comparison to ionized calcium, the calcium concentrations in dialysate and substitution fluid were respectively 1.40 and 1.75 mmol/l. In addition, small molecule clearances were matched. Their results show that although cardiac index (CI) decreased significantly during i-UF and HF, MAP was best preserved during these treatments as a result of a significant increase in TPR which was associated with an increase in plasma noradrenaline while heart rate (HR) remained unchanged. In contrast, the HR associated smaller decrease in CI during UF+HD resulted in a significant decrease in MAP: during UF+HD with bicarbonate buffering TPR remained more or less unchanged while with acetate, TPR tended to decrease. Moreover, no significant changes in plasma noradrenaline were observed during UF+HD irrespective of the buffer used³⁴.

Kooman et al. showed that together with an impaired response in TPR, venous reactivity is also impaired during UF+HD. Using venous-occlusion plethysmography they measured changes in both forearm vascular resistance and venous tone in 12 stable hemodialysis patients during sequential i-UF (45 min.) and UF+HD (45 min., bicarbonate, sodium 141 and calcium 1.75 mmol/l, 37.5°C). The ultrafiltration rate was fixed at 1.0 l/hr while the order of treatments was randomized. Their results show that peripheral resistance and venous tone increased only during i-UF while in several patients venodilation occurred during UF+HD³⁵.

Fox et al. studied in a cross-over format the hemodynamic response during

UF+HD and HF using thoracic electrical bioimpedance. Both treatments were matched with respect to membrane type (AN69), the dialysate and substitution fluid composition and temperature (acetate, sodium 144 and calcium 1.75 mmol/l, 37.0°C). Stroke index (SI) and TPR decreased significantly during UF+HD while CI and HR increased. HF was associated with a comparable decrease in SI. However, the increase in HR during HF was significantly less as compared to UF+HD while both CI and TPR remained stable. Plasma noradrenaline increased comparably during both treatments³⁶.

In general, the results obtained by Baldamus et al., Kooman et al. and Fox et al. are in agreement with the results obtained by others³⁷⁻⁴⁴. Consequently, it may be concluded that ultrafiltration induced decline in plasma volume is well buffered during both i-UF and HF by the following mechanisms:

1. TPR increases in order to maintain arterial blood pressure at the pre-capillary side while temporarily mobilising 'unstressed' blood by passive venous constriction.
2. Venous tone increases in order to maintain cardiac filling and output by the mobilisation of 'unstressed' blood. However, venous tone has not been directly measured yet during HF.

In contrast, during UF+HD with bicarbonate buffering, both arteriolar and venous reactivity are impaired. As a result, heart rate will increase in order to maintain cardiac output as much as possible (table 1.2). The hemodynamic effects of a decline in plasma volume during acetate buffered UF+HD are even amplified due to direct the vasodilatory capacity of acetate.

Table 1.2 Differences between dialysis strategies: hemodynamics

	UF+HD	HF	i-UF
arterial blood pressure	unstable	stable	stable
peripheral resistance	=/↓	↑↑	↑↑
venous tone	=/↓	??	↑↑
catecholamines	=	↑	↑

Disparity in vascular reactivity: pathogenetic factors

The foregoing discussion points towards differences in vascular reactivity as main determinant of differences in blood pressure stability during different treatments modalities while differences in plasma volume preservation seem to play a minor

role. When applying, for example, the results obtained by de Vries et al. to a dialysis treatment of four hours, the difference in PVP between UF+HD and HF seems to be relatively small: a decrease in blood volume of respectively -9.0% and -6.8% ²³. On the other hand, reproducible differences in both arteriolar and venous reactivity have been found in a number of hemodynamic studies when comparing UF+HD with either i-UF or HF³⁴⁻³⁶. However, the mechanisms responsible for these differences in hemodynamics are still poorly understood. Identification of the factors involved will improve our understanding of the pathophysiology of hemodynamic instability during dialysis therapy and may provide us with new tools to prevent the occurrence of this complication.

The following discussion will focus on factors (table 1.3) which might be of importance when studying the interaction between the dialysis treatment and cardiovascular responses. In addition, the role of the autonomic nervous system will be discussed.

Table 1.3 Differences between dialysis strategies: pathogenetic factors

	UF+HD	HF	i-UF
extracorporeal temperature	=/↑	↓	↓↓
osmolality	↓↓	↓	=
bio-incompatibility			
membrane	high	low	high
dialysate/infusate	high	low	
clearance			
small molecules	high	low	
middle molecules	low	high	
total CO ₂	↑↑	↑↑	=/↓
calcium	↑	↑	=

Extracorporeal blood temperature

Blood cools during extracorporeal circulation in i-UF in the absence of dialysate and infusate. Maggiore et al. found a difference of 2.0°C in blood temperature between the arterial and venous line in i-UF⁴⁵. In HF, extracorporeal blood cooling is compensated by heat transfer from the substitution fluid to blood. However, depending upon the infusate temperature and as a result of the relatively large extracorporeal circuit and a low infusate flow (± 100 ml/min in post-dilution HF) compensation is only partial. According to Maggiore, blood cools down during extracorporeal circulation by about 0.7°C in HF⁴⁶. In contrast, cooling of blood in

the extracorporeal circuit during UF+HD is (over)compensated by heat transfer across the dialysis membrane from dialysate (500 ml/min) to blood. Lindholm et al. found that at a dialysate temperature of 37.0°C the blood temperature increased from 35.8°C in the arterial line to 36.9°C in the venous line. A dialysate temperature of 36.0°C resulted in comparable blood temperatures in the arterial and venous lines⁴⁷.

Differences in thermal balance between different dialysis treatments are even more stressed by the observation that only UF+HD is associated with an increase in core body temperature or in the blood temperature measured in the arterial line in the range of 0.7 to 1.0°C during four hours treatment at a dialysate temperature in the range of 37.0 to 38.0°C⁴⁵⁻⁴⁷. Although the observed increase in temperature might be partially explained by the presence of dialysate derived contaminants in UF+HD⁴⁸, Maggiore et al. found an increase in core body temperature of 1.0°C also during HF (sterile substitution fluid) in which blood reentering the patient had a comparable temperature as in UF+HD with a dialysate temperature of 38.0°C⁴⁶.

Most authors who studied the relation between the dialysate temperature and cardiovascular stability during UF+HD found an improvement in hemodynamic stability when dialysate temperature was lowered to about 35.0°C^{47,49}. A better increase in total peripheral resistance⁵⁰⁻⁵² and myocardial contractility⁵³ in association with higher levels of plasma noradrenaline^{52,54} are probably responsible for these observations. Only Maggiore et al. has actually compared UF+HD with both i-UF and HF in relation to the extracorporeal blood temperature measuring arterial blood pressure and heart rate. They found a reversal of hemodynamic stability in both i-UF and HF when blood reentering the patient was warmed to a comparable temperature as in UF+HD. However, at equivalent changes in the extracorporeal blood temperature, both i-UF and HF did afford some advantage over UF+HD with respect to hemodynamic stability; tachycardia was still more pronounced during UF+HD while in contrast to HF, symptomatic hypotension occurred during cold UF+HD^{45,46}. In agreement, Fox et al. also found a qualitatively different cardiovascular response between HF and UF+HD with both treatments matched for dialysate and infusate temperature³⁶. However, the results of both studies are confounded by the fact that in UF+HD acetate was used as a buffer while Fox et al. did not directly measure extracorporeal blood temperatures.

Osmolality

Traditionally, iso-osmolality was thought to be the main factor responsible for better hemodynamic stability during i-UF compared to UF+HD¹. In our opinion and to our knowledge, direct evidence is still lacking whether changes in plasma osmolality influence cardiovascular reactivity during dialysis. It is important to

realise that the decrease in plasma osmolality during UF+HD is the result of a decrease in plasma urea together with changes in the concentration of different electrolytes, especially sodium which is a tonically active substance⁵⁵. Osmolality as a blood pressure stabilising factor should thus always be related to PVP.

Vascular reactivity during UF+HD is probably not influenced by a decrease in plasma osmolality as a result of small molecule clearance. Daugirdas et al. found no differences in hemodynamics in acutely uremic dogs, comparing urea-free and urea-supplemented dialysate (bicarbonate)⁵⁶. Wehle et al. have also shown that in isovolemic HD, adding of urea to dialysate to minimise the fall in plasma osmolality did not prevent a reduction in blood pressure or a decrease in total peripheral resistance⁵⁷. Moreover, different authors have still demonstrated a qualitatively different cardiovascular response between UF+HD and HF with both treatments matched for small molecule (urea) clearance and changes in plasma osmolality^{34,36,58}.

Concerning plasma osmolality in relation to the dialysate sodium concentration, many authors have shown that blood pressure stability during UF+HD is improved with higher dialysate sodium concentrations, especially in acetate buffered dialysis⁵⁹⁻⁶¹. The improved hemodynamic stability that was stated before at higher dialysate sodium concentrations seems mainly the result of a better preservation of plasma volume^{14,16}. In a non-uremic dog model, Daugirdas et al. could not find any adverse effects of hyponatric dialysate (90% of the plasma sodium concentration, bicarbonate) on peripheral resistance as compared to isonatric UF+HD while ultrafiltration was better tolerated during the latter treatment. The hemodynamic data, although not directly measured, suggest that venous capacitance was also unaffected⁶². However, influence of dialysate sodium concentration on vascular reactivity is suggested by the results obtained by Schultze et al.. They found a significant larger increase in plasma PGE₂ while blood pressure was significantly lower after low-sodium UF+HD (126 mmol/l) as compared to high-sodium UF+HD (140 mmol/l). The authors speculate that low-sodium dialysate might contribute to hypotension due to PGE₂ induced vasodilation⁶³.

Henrich et al. studied the blood pressure and heart rate response in six dialysis patients during regular UF+HD, i-UF and iso-osmolar UF+HD as well as during an orthostatic manoeuvre performed directly after these treatments at comparable decrements in body weight (2.1-3.2%). Iso-osmolality was achieved by hypertonic mannitol infusion (25%, 0.16 g/kg/hr) while in another dialysis session isotonic mannitol (5%) was infused in a volume five times the volume used in hypertonic dialysis. Plasma osmolality remained stable only during i-UF and UF+HF with hypertonic mannitol infusion. During the different treatments, no significant changes in supine blood pressure were observed. However, on assumption of the upright posture, blood pressure decreased significantly only after both regular

UF+HD and UF+HD with iso-tonic mannitol infusion and not after i-UF and hypertonic UF+HD. The authors suggest a direct effect of plasma osmolality on vascular reactivity since a comparable increase in hematocrit and plasma proteins during hypertonic and isotonic mannitol UF+HD are indicative of a comparable decline in plasma volume⁶⁴.

Dialysate derived contaminants

Differences in cardiovascular reactivity might also be related to the presence or absence of a yet unidentified vaso-active agent during either treatment. With (1) vascular reactivity being non-physiological in UF+HD, (2) a negligible molecular clearance in i-UF and (3) assuming a common underlying mechanism responsible for the disparity in hemodynamics between UF+HD and both i-UF and HF, the most likely possibility is a vasodilator or a vasoconstrictor-inhibitor produced only during UF+HD and not during i-UF and HF. Alternatively, a small molecular vasoconstrictive agent might be produced during all three treatment modalities that is not cleared during i-UF and to a lesser extent during HF as compared to UF+HD. The finding, however, of a distinct difference in hemodynamics between UF+HD and HF despite matching for small molecule clearances makes the latter possibility less likely³⁴.

In this respect, an interesting theory that has attained much attention since its publication in 1983 is formulated in the well known 'Interleukin Hypothesis'. The authors suggested that hemodynamic instability during UF+HD could be related to the release of Interleukin-1 by blood mononuclear cells after stimulation by C5a and or by pyrogenic materials derived from dialysate. The better hemodynamic stability, as observed during i-UF and HF would then reflect the absence of dialysate (and contaminants) in both treatment modalities and the use of more biocompatible membranes in HF⁶⁵.

Since the publication of the 'Interleukin Hypothesis', the effects of uremia and hemodialysis on the production of especially IL-1 and TNF but also of IL-6 and IL-8 by peripheral mononuclear cells have been studied extensively⁶⁶. It is now generally accepted that monocytes of patients on hemodialysis are in a state of activation, i.e. they have an elevated cytokine content and they are 'primed' to produce cytokines in larger amounts after in-vitro stimulation by lipopolysaccharides (LPS) in comparison with monocytes from normal controls⁶⁷⁻⁷¹. Conflicting data exist, however, whether plasma levels of circulating cytokines are elevated in patients on hemodialysis and whether they increase during the hemodialysis procedure⁷²⁻⁷⁵. Complement activation (C5a) during (cuprophane) hemodialysis probably only results in mRNA synthesis while dialysate derived endotoxins are necessary for cytokine synthesis and release^{76,77}. In-vitro studies

have shown that fragments of LPS are indeed able to pass most dialysis membranes used in daily clinical practice⁷⁸⁻⁸⁰. Studies with radiolabelled LPS even suggest that the complete LPS particle can pass dialysis membranes of different origins⁸¹. Membrane transfer of cytokine inducing substances at lower, clinical more relevant concentrations of endotoxins, could however not be demonstrated⁸².

The relevance of the mentioned observations with respect to hemodynamic instability remains to be established. Both IL-1 and TNF as well as LPS are known to induce hypotension at certain plasma levels when directly injected in humans with both cytokines acting synergistically⁸³. Although the degree of pyrogen exposition during hemodialysis is rather small, low grade monocyte stimulation in relation to dialysate derived contaminants might be responsible for impairing the normal physiologic response to a decline in plasma volume as observed during UF+HD. However, for their biological action, both IL-1 and TNF are dependent on the activity of their specific antagonists, namely the IL-1 receptor antagonist (IL-1Ra) and the soluble TNF receptor (sTNFR), which have both been found elevated in hemodialysis patients^{84,85}. In addition, LPS activity is determined by the activity of both lipopolysaccharide binding protein (LBP) and bactericidal/permeability increasing protein (BPI). BPI has been shown to increase gradually during hemodialysis⁸⁶. Finally, the acute effect of monocyte stimulation on vascular reactivity during hemodialysis is probably also limited by sequestration of activated monocytes in the pulmonary vasculature⁸⁷.

Buffer substrate and acid-base status

Correction of uremic acidosis is one of the main objectives of dialysis therapy. Influx of buffer substrate or changes in acid-base balance may play a role in the impaired hemodynamic response during UF+HD as compared to i-UF. The latter is associated with bicarbonate loss as a result of convective transport. Until the eighties mainly acetate was used as a buffer in dialysate⁸⁸. However, acetate has direct vasodilatory capacities⁸⁹, induces hypoxemia^{90,91} and may impair myocardial contractility⁹². Especially in high efficiency dialysis, when acetate load may exceed the metabolic capacity of muscles and liver, the adverse effects of acetate become apparent⁸⁸. In this respect, the advantage of bicarbonate as a buffer in reducing intradialytic morbidity has been documented in several studies, especially in patients who are prone to developing symptomatic hypotension or with dialysate sodium concentrations of about 130 mmol/l^{68,90,91,93-95}.

Most probably, a decrease in total peripheral resistance during UF+HD, as reported by different authors is explained by the use of acetate. Baldamus et al. found that only in acetate buffered UF+HD, total peripheral resistance tended to decrease during the whole treatment, while in bicarbonate buffered UF+HD total

peripheral resistance first tended to increase after which it tended to decrease back to the baseline level. In addition, acetate buffered UF+HD was associated with a significant larger increase in heart rate as compared to bicarbonate buffered UF+HD³⁴. Thus, although bicarbonate has most probably no direct adverse effect on hemodynamics, the results obtained by both Baldamus et al. and Kooman et al. show that arteriolar and venous reactivity remain impaired during bicarbonate UF+HD as compared to both i-UF and HF^{34,35}. Moreover, HF is still associated with a significant increase in total peripheral resistance even when acetate is used as a buffer in the substitution fluid³⁴.

Calcium

Calcium ions play a critical role in the contractile process of both cardiac and vascular smooth muscle cells as well as in the release of catecholamines from adrenergic nerves^{96,97}. Therefore, changes in ionized calcium are potentially important denominators of the cardiovascular response during different dialysis strategies.

A dialysate calcium concentration in the range of 1.50-1.75 mmol/l results in an increase of the plasma ionized calcium concentration of about 0.20 mmol/l while with a dialysate calcium concentration of 1.25 mmol/l in UF+HD, plasma ionized calcium remains more or less unchanged⁹⁸. The increase has been shown to be dependent on the degree of acidosis correction with a negative correlation between changes in ionized calcium and changes in total CO₂ ($r = -0.70$; $p < 0.001$)⁹⁹.

In a number of studies arterial blood pressure during UF+HD was indeed found to be positively correlated with changes in ionized calcium⁹⁹⁻¹⁰³. In this respect, an increase in left ventricular contractility has been a consistent finding^{100,102,103}. Fellner et al. could not find a relation between the level of plasma ionized calcium and both total vascular resistance and left ventricular end-diastolic dimensions, measured echocardiographically in eight dialysis patients after dialysis with three different dialysate calcium concentrations (0.5, 1.75 and 2.5 mmol/l)¹⁰³. Although they only performed post-dialytic echographies, these results suggest that the peripheral vascular response during UF+HD is not directly related to changes in plasma ionized calcium. In the study of Kooman et al. peripheral vascular tone remained unchanged during UF+HD despite a significant increase in plasma ionized calcium³⁵. These results are in contrast with the observation that under normal physiologic conditions acute hypercalcemia (serum total calcium from 2.2 to 3.2 mmol/l in three hours) results in an increase in systemic vascular resistance⁹⁷. Therefore, measurement of peripheral vascular reactivity during dialysis at different levels of plasma ionized calcium seems necessary.

Autonomic Nervous System

A treatment specific dysfunction of the autonomic nervous system (ANS) has been suggested to play a role in the disparative hemodynamic response during different forms of dialysis therapy. Bergström extensively reviewed the relation between the ANS and blood pressure control in dialysis therapy. It might be concluded on the basis of plasma catecholamine levels that UF+HD is indeed associated with an impairment of the ANS function although conflicting data exist. UF+HD induces a variable response of plasma noradrenaline while in contrast, both i-UF and HF are associated with a more marked increase. However, the noradrenaline response to different autonomic stimuli is probably not impaired by UF+HD¹⁰⁴. These results, however, should be considered with caution since plasma catecholamine levels are only weak denominators of regional sympathetic activity, especially in uremic patients^{105,106}. Factors that may influence efferent sympathetic activity during dialysis are the extracorporeal blood temperature^{52,54}, changes in plasma ionized calcium⁹⁷ and the presence of dialysate derived contaminants⁸³.

Zucchelli et al concluded that UF+HD is associated with an acute dysfunction of the baroreceptors on the basis of a normal cold pressor and diving test after both UF+HD and HF and an abnormal Valsalva ratio and baroreceptor reflex sensitivity index after UF+HD. Moreover lower levels of plasma noradrenaline were found after UF+HD in comparison with HF. Pre-treatment autonomic function tests were comparable to controls¹⁰⁷. At variance with this, Baldamus et al. found a normalisation of baroreflex sensitivity during (isovolemic) HF which remained impaired during (isovolemic) HD¹⁰⁸. So, although both studies point at an alteration of the ANS function at the level of the afferent part of the baroreflex arc, they are not completely conclusive.

A technique that may improve our knowledge about the role of the ANS in the disparative hemodynamic response during different forms of dialysis therapy in relation to different factors is microneurography. With this technique it is possible to measure efferent, post-ganglionic muscle nerve sympathetic activity directly¹⁰⁹. Such measurements have already been performed in the interdialytic period in nephrectomized and non-nephrectomized normotensive and hypertensive uremic patients as well as during dialysis comparing stable with hypotensive prone patients^{110,111}. The latter study showed that acute hypotension during UF+HD is related to an acute decrease in both calf-vascular resistance and peripheral sympathetic nerve activity¹¹¹.

Conclusion

Hemodynamic instability is still an important problem in dialysis therapy. Maintenance of blood pressure during dialysis is related to an optimal preservation of plasma volume as well as to adequate cardiovascular compensating mechanisms. In this respect, it is distinctly established that on the basis of an impaired vascular response, blood pressure is less stable during UF+HD as compared to both i-UF and HF. However, the mechanisms responsible for these differences in vascular reactivity are still poorly understood.

Differences in the extracorporeal blood temperature seem to be at least partially responsible. However, only few comparative studies have yet been performed, while most studies are confounded by the fact that acetate was used as a buffer in UF+HD. Based on measurements of blood pressure, heart rate and prostaglandins it is suggested that the relatively large decline in plasma osmolality could also lead to an impairment of the vascular response during UF+HD. In addition, vascular reactivity could also be influenced by changes in acid-base status. Ionized calcium has been shown to be an important determinant of the blood response during dialysis therapy by influencing myocardial contractility. However, it is unknown whether vascular reactivity is also influenced by changes in ionized calcium. Finally, the clinical relevance of bioincompatibility factors with respect to hemodynamics remains to be established.

Aim of the Study

The aim of the present study is to identify factors that could be responsible for the impaired vascular response during UF+HD as compared to i-UF and HF. The role of the following factors in the vascular response during dialysis therapy was studied under completely standardised conditions:

- extracorporeal blood temperature
- plasma osmolality
- dialysate derived contaminants
- plasma ionized calcium

The first study compares i-UF and UF+HD in relation to differences in extracorporeal blood temperature.

- 1. Examination of the role of the extracorporeal blood temperature in the disparity in vascular reactivity between i-UF and UF+HD (bicarbonate).**
-

To be able to study the role of the other factors separately, vascular reactivity was studied during UF+HD in relation to a single factor:

2. **Establishing the role of changes in plasma osmolality, the presence of dialysate derived contaminants and changes in ionized calcium in the vascular response during UF+HD (bicarbonate).**

Based on the results of these first four studies a comparison was made between UF+HD and HF in relation to the factor that appeared to be the most important:

3. **Study the role of the extracorporeal blood temperature in the disparity in vascular reactivity between UF+HD (bicarbonate) and pre-dilution HF (bicarbonate).**
-

References

- 1 Bergström J, Asaba H, Fürst P, Oulès R: Dialysis, ultrafiltration and blood pressure. *Proc EDTA* 1976; 13: 293-305.
 - 2 Bergström J: Ultrafiltration without dialysis for removal of fluid and solutes in uremia. *Clin Nephrol* 1978; 9: 156-164.
 - 3 Quellhorst E, Doht B, Schuenemann B: Hemofiltration: treatment of renal failure by ultrafiltration and substitution. *J Dial* 1977; 1: 529-543.
 - 4 Baldamus CA, Quellhorst E: Outcome of long-term hemofiltration. *Kidney Int* 1985; 17: S41-S46.
 - 5 Henderson LW: Heterogeneity of the cardiovascular response to hemofiltration. *Kidney Int* 1986; 29: 901-907.
 - 6 Keshaviah P, Berkseth R, Ilstrup K, McMichael C, Collins A: Reduced treatment time: hemodialysis (HD) versus hemofiltration (HF). *Trans Am Soc Artif Intern Organs* 1985; 31: 176-182.
 - 7 Malberti F, Surian M, Colussi G, Minetti L: Acid-base status and dialysis symptoms in different forms of renal replacement therapy. *Kidney Int* 1988; 33: S117-S119.
 - 8 Rodriguez M, Pederson JA, Llach F: Effects of dialysis and ultrafiltration on osmolality, colloid osmotic pressure, and vascular refilling rate. *Kidney Int* 1985; 28: 808-813.
 - 9 Leunissen KML, Noordzij TC, van Hooff JP: Pathophysiologic aspects of plasma volume preservation during dialysis and ultrafiltration. *Contrib Nephrol* 1990; 78: 201-211.
 - 10 de Vries PMJM: Plasma volume changes during hemodialysis. *Sem Dial* 1992; 5: 42-47.
 - 11 Mann H, Stiller S, Schallenberg U, Thömmes A: Optimizing dialysis by variation of ultrafiltration rate and sodium concentration controlled by continuous measurement of circulation blood volume. *Contrib Nephrol* 1989; 74: 182-190.
 - 12 de Vries J-PPM, Kouw PM, van der Meer NJM, Olthof CG, Oe LP, Donker AJM, de Vries PMJM: Non-invasive monitoring of blood volume during hemodialysis: Its relation with post-dialytic dry weight. *Kidney Int* 1993; 44: 851-854.
 - 13 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
 - 14 Fleming SJ, Wilkinson JS, Greenwood RN, Aldridge C, Baker LRI, Cattell WR: Effect of dialysate composition on intercompartmental fluid shift. *Kidney Int* 1987; 32: 267-273.
 - 15 Chanard J, Brunois JP, Toupance O, Roujoulkeh H, Milcent T: Effect of sodium dialysate concentration on the extracellular fluid distribution during hemodialysis. *Contrib Nephrol* 1984; 41: 371-375.
 - 16 de Vries PMJM: Fluid balance during haemodialysis and haemofiltration: The effect of dialysate sodium. *Nephrol Dial Transplant* 1990; 5 suppl 1: 158-161.
-

- 17 Hsu CH, Schwartz RD, Somermeyer MG, Raj A: Bicarbonate hemodialysis: Influence on plasma refilling and hemodynamic stability. *Nephron* 1984; 38: 202-208.
- 18 Leunissen KML, Cheriex EC, Janssen J, Teule GJJ, Mooy JMV, Ramentol M, van Hooff JP: Influence of left ventricular function on changes in plasma volume during acetate and bicarbonate dialysis. *Nephrol Dial Transplant* 1987; 2: 99-103.
- 19 Kooman JP, Wijnen JAG, Draaijer P, van Bortel LMAB, Gladziwa U, Peltenburg HG, Struyker-Boudier HAJ, van Hooff JP, Leunissen KML: Compliance and reactivity of the peripheral venous system in chronic intermittent hemodialysis. *Kidney Int* 1992; 41: 1041-1048.
- 20 Fauchald P: Effects of ultrafiltration on body fluid volumes and transcapillary colloid osmotic gradient in hemodialysis patients. *Contrib Nephrol* 1989; 74: 170-175.
- 21 Keshaviah P, Shapiro FL: A critical examination of dialysis induced hypotension. *Am J Kidney Dis* 1982; 11: 290-301.
- 22 Fleming SJ, Wilkinson JS, Aldridge C, Greenwood RN, Baker LRI, Cattell WR: Blood volume changes during isolated ultrafiltration and combined ultrafiltration-dialysis. *Nephrol Dial Transplant* 1988; 3: 272-276.
- 23 de Vries PMJM, Olthof CG, Solf A, Schuenemann B, Oe PL, Quellhorst E, Schneider H, Donker AJM: Fluid balance during haemodialysis and haemofiltration: The effect of dialysate sodium and a variable ultrafiltration rate. *Nephrol Dial Transplant* 1991; 6: 257-263.
- 24 Olthof CG, de Vries PMJM, Kouw PM, Oe PL, Gerlag PGG, Schneider H, Donker AJM: The recovery of the fluid balance after hemodialysis and hemofiltration. *Clin Nephrol* 1992; 37: 135-139.
- 25 Hirsch AT, Levenson DJ, Cutler SS, Dzau VJ, Creager MA: Regional vascular responses to prolonged lower body negative pressure in normal subjects. *Am J Physiol* 1989; 257: H219-H225.
- 26 Berdeaux A, Duranteau J, Pussard E, Edouard A, Giudicelli J-F: Baroreflex control of regional vascular resistances during simulated orthostatism. *Kidney Int* 1992; 41: S29-S33.
- 27 Hintze TH, Vatner SF: Cardiac dynamics during hemorrhage: relative unimportance of adrenergic inotropic responses. *Circ Res* 1982; 50: 705-713.
- 28 Usioda E, Nuwayhid B, Kleinman G, Tabsh K, Brinkman III CR, Assali NS: The contribution of the β -adrenergic system to the cardiovascular response to hypovolemia. *Am J Obstet Gynecol* 1983; 147: 423-429.
- 29 Sander-Jensen K, Mehlsen J, Stadeager C, Christensen NJ, Fahrenkrug J, Schwartz TW, Warberg J, Bie P: Increase in vagal activity during hypotensive lower-body negative pressure in humans. *Am J Physiol* 1988; 255: R149-R156.
- 30 Guyton AC, Jones CE, Coleman TG (ed.): *Circulatory physiology: cardiac output and its regulation*. Philadelphia London Toronto, W.B. Saunders Company, 1973, 556.

-
- 31 Rothe CF: Venous system: Physiology of capacitance vessels; in Shepherd JT, Abboud FM (ed.): *Handbook of Physiology*. Bethesda, American Physiological Society, 1983, 3, Peripheral circulation and organ blood flow, pp. 397-452
 - 32 Greenway CV, Lautt WW: Blood volume, the venous system, preload, and cardiac output. *Can J Physiol Pharmacol* 1986; 64: 383-387.
 - 33 Foley RN, Parfrey PS, Harnett JD: Left ventricular hypertrophy in dialysis patients. *Sem Dial* 1992; 5: 34-41.
 - 34 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
 - 35 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
 - 36 Fox SD, Henderson LW: Cardiovascular response during hemodialysis and hemofiltration: thermal, membrane, and catecholamine influences. *Blood Purif* 1993; 11: 224-236.
 - 37 Wehle B, Asaba H, Castenfors J, Fürst P, Gunnarsson B, Shaldon S, Bergström J: Hemodynamic changes during sequential ultrafiltration and dialysis. *Kidney Int* 1979; 15: 411-418.
 - 38 Shaldon S, Deschodt G, Beau MC, Claret G, Mion H, Mion C: Vascular stability during high flux haemofiltration (HF). *Proc EDTA* 1979; 16: 695-697.
 - 39 Hampl H, Paepre H, Unger V, Fischer C, Resa I, Kessel M: Hemodynamic changes during hemodialysis, sequential ultrafiltration, and hemofiltration. *Kidney Int* 1980; 18: S83-S88.
 - 40 Chaignon M, Tzuoh Chen W, Tarazi RC, Bravo EL, Nakamoto S: Effect of hemodialysis on blood volume distribution and cardiac output. *Hypert* 1981; 3: 327-332.
 - 41 Kinet J-P, Soyeur D, Bolland N, Saint-Remy M, Collignon P, Godon J-P: Hemodynamic study of hypotension during hemodialysis. *Kidney Int* 1982; 21: 868-876.
 - 42 Baldamus CA, Ernst W, Lysaght MJ, Shaldon S, Koch KM: Hemodynamics in hemofiltration. *Int J Artif Organs* 1983; 6: 27-31.
 - 43 Bradley JR, Evans DB, Gore SM, Cowley AJ: Is dialysis hypotension caused by an abnormality of venous tone? *BMJ* 1988; 296: 1634-1637.
 - 44 Bradley JR, Evans DB, Cowley AJ: Comparison of vascular tone during combined haemodialysis with ultrafiltration and during ultrafiltration followed by haemodialysis: a possible mechanism for dialysis hypotension. *BMJ* 1990; 300: 1312.
 - 45 Maggiore Q, Pizzarelli F, Zoccali C, Sica S, Nicolò F: Influence of blood temperature on vascular stability during hemodialysis and isolated ultrafiltration. *Int J Artif Organs* 1985; 8: 175-178.
-

-
- 46 Maggiore Q, Pizzarelli F, Sisca S, Zoccali C, Parlongo S, Nicolò F, Creazzo G: Blood temperature and vascular stability during hemodialysis and hemofiltration. *Trans Am Soc Artif Intern Organs* 1982; 28: 523-527.
 - 47 Lindholm T, Thysell H, Yamamoto Y, Forsberg B, Gullberg CA: Temperature and vascular stability in hemodialysis. *Nephron* 1985; 39: 130-133.
 - 48 Dinarello CA: Pathogenesis of fever during hemodialysis. *Contrib Nephrol* 1983; 36: 90-99.
 - 49 Marcén R, Quereda C, Orofino L, Lamas S, Teruel JL, Matesanz R, Ortuño J: Hemodialysis with low-temperature dialysate: A long-term experience. *Nephron* 1988; 49: 29-32.
 - 50 Coli U, Landini S, Lucatello S, Fracasso A, Morachiello P, Righetto F, Scanferla F, Onesti G, Bazzato G: Cold as cardiovascular stabilizing factor in hemodialysis: hemodynamic evaluation. *Trans Am Soc Artif Intern Organs* 1983; 29: 71-75.
 - 51 Bazzato G, Coli U, Landini S, Lucatello S, Fracasso A, Morachiello P, Righetto F, Scanferla F: Temperature monitoring in dialysis-induced hypotension. *Kidney Int* 1985; 28: S161-S165.
 - 52 Jost CMT, Agarwal R, Khair-El-Din T, Grayburn PA, Victor RG, Henrich WL: Effects of cooler temperature dialysate on hemodynamic stability in "problem" dialysis patients. *Kidney Int* 1993; 44: 606-612.
 - 53 Levy FL, Grayburn PA, Foulks CJ, Brickner ME, Henrich WL: Improved left ventricular contractility with cool temperature hemodialysis. *Kidney Int* 1992; 41: 961-965.
 - 54 Mahida BH, Dumler F, Zasuwa G, Fleig G, Levin NW: Effect of cooler dialysate on serum catecholamines and blood pressure stability. *Trans Am Soc Artif Intern Organs* 1983; 24: 384-389.
 - 55 Gennari FJ: Serum osmolality: Uses and limitations. *N Engl J Med* 1984; 310: 102-105.
 - 56 Daugirdas JT, Ing TS, Chen W-T, Vestal RE, Hano JE, Al-Kudsi RR: Ultrafiltration hemodynamics in conscious, uremic dogs: Effect of a decreasing plasma urea level. *Mineral Electrolyte Metab* 1985; 11: 79-84.
 - 57 Wehle B, Asaba H, Castenfors J, Gunnarsson B, Bergström J: Influence of dialysate composition on cardiovascular function in isovolemic haemodialysis. *Proc EDTA* 1981; 18: 153-159.
 - 58 Shaldon S: Role of small molecule removal in the control of treatment morbidity with haemodialysis and haemofiltration. *Proc EDTA* 1981; 18: 249-255.
 - 59 Henrich WL, Woodard TD, McPhaul JJ: The chronic efficacy and safety of high sodium dialysate: double-blind, crossover study. *Am J Kidney Dis* 1982; 2: 349-353.
 - 60 Cybulsky AVE, Matni A, Hollomby DJ: Effects of high sodium dialysate during maintenance hemodialysis. *Nephron* 1985; 41: 57-61.
 - 61 Sadowski RH, Allred EN, Jabs K: Sodium modeling ameliorates intradialytic and interdialytic symptoms in young hemodialysis patients. *J Am Soc Nephrol* 1993; 4: 1192-1198.
-

-
- 62 Daugirdas JT, Purandare VV, Ing TS, Chen WT, Popli S, Hano JE, Klok MA: Ultrafiltration hemodynamics in an animal model: effect of a decreasing plasma sodium level. *Trans Am Soc Artif Intern Organs* 1984; 30: 603-609.
 - 63 Schultze G, Maiga M, Neumayer H-H, Wagner K, Keller F, Molzahn M, Nigam S: Prostaglandin E₂ promotes hypotension on low-sodium hemodialysis. *Nephron* 1984; 37: 250-256.
 - 64 Henrich WL, Woodard TD, Blachley JD, Gomez-Sanchez C, Pettinger W, Cronin RE: Role of osmolality in blood pressure stability after dialysis and ultrafiltration. *Kidney Int* 1980; 18: 480-488.
 - 65 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
 - 66 Dinarello CA: Interleukin-1 and tumor necrosis factor and their naturally occurring antagonists during hemodialysis. *Kidney Int* 1992; 42: S68-S77.
 - 67 Blumenstein M, Schmidt B, Ward RA, Ziegler-Heitbrock HWL, Gurland HJ: Altered interleukin-1 production in patients undergoing hemodialysis. *Nephron* 1988; 50: 277-281.
 - 68 Herbelin A, Ureña P, Nguyen AT, Zingraff J, Descamps-Latscha B: Influence of first and long-term dialysis on uremia-associated increased basal production of interleukin-1 and tumor necrosis factor α by circulating monocytes. *Nephrol Dial Transplant* 1991; 6: 349-357.
 - 69 Pertosa G, Tarantino EA, Gesualdo L, Montinaro V, Schena FP: C5b-9 generation and cytokine production in hemodialyzed patients. *Kidney Int* 1993; 43: S221-S225.
 - 70 Ryan JJ, Beynon HLC, Rees AJ, Cassidy MJD: In vitro production of tumor necrosis factor by monocytes cultured from dialysis patients. *Kidney Int* 1993; 43: S226-S229.
 - 71 Haeffner-Cavaillon N, Cavaillon J-M, Ciancioni C, Bacle F, Delons S, Kazatchkine MD: In vivo induction of interleukin-1 during hemodialysis. *Kidney Int* 1989; 35: 1212-1218.
 - 72 Bingel M, Lonnemann G, Koch KM, Dinarello CA, Shaldon S: Plasma interleukin-1 during hemodialysis: the influence of dialysis membranes. *Nephron* 1988; 50: 273-276.
 - 73 Tetta C, Camussi G, Turello E, Salomone M, Aimo G, Priolo G, Segoloni G, Vercellone A: Production of cytokines in hemodialysis. *Blood Purif* 1990; 8: 337-346.
 - 74 Herbelin A, Nguyen AT, Zingraff J, Ureña P, Descamps-Latscha B: Influence of uremia and hemodialysis on circulating interleukin-1 and tumor necrosis factor α . *Kidney Int* 1990; 37: 116-125.
 - 75 Miyazaki T, Niwa T, Sato M, Kambe F, Maeda K, Seo H: Plasma Interleukin 8 levels are increased by hemodialysis. *Blood Purif* 1994; 12: 135-140.
 - 76 Schindler S, Lonneman G, Shaldon S, Koch KM, Dinarello CA: Transcription, not synthesis, of interleukin-1 and tumor necrosis factor by complement. *Kidney Int* 1990; 37: 85-93.
-

-
- 77 Schindler R, Linnenweber S, Schulze M, Oppermann M, Dinarello CA, Shaldon S, Koch KM: Gene expression of interleukin-1 β during hemodialysis. *Kidney Int* 1993; 43: 712-721.
 - 78 Evans RC, Holmes CJ: In vitro study of the transfer of cytokine-inducing substances across selected high-flux hemodialysis membranes. *Blood Purif* 1991; 9: 92-101.
 - 79 Lonnemann G: Dialysate bacteriological quality and the permeability of dialyzer membranes to pyrogens. *Kidney Int* 1993; 43: S195-S200.
 - 80 Ureña P, Herbelin A, Zingraff J, Lair M, Khoa Man N, Descamps-Latscha B, Drücke T: Permeability of cellulosic and non-cellulosic membranes to endotoxin subunits and cytokine production during in-vitro haemodialysis. *Nephrol Dial Transplant* 1992; 7: 16-28.
 - 81 Laude-Sharp M, Caroff M, Simard L, Pusineri C, Kazatchkine MD, Haeflner-Cavaillon N: Induction of IL-1 during hemodialysis: transmembrane passage of intact endotoxins. *Kidney Int* 1990; 38: 1089-1094.
 - 82 Kumano K, Yokota S, Nanbu M, Sakai T: Do cytokine-inducing substances penetrate through dialysis membranes and stimulate monocytes? *Kidney Int* 1993; 43: S205-S208.
 - 83 Dinarello CA: Cytokines: agents provocateurs in hemodialysis? *Kidney Int* 1992; 41: 683-694.
 - 84 Pereira BJG, Shapiro L, King AJ, Falagas ME, Strom JA, Dinarello CA: Plasma levels of IL-1 β , TNF α and their specific inhibitors in undialyzed chronic renal failure, CAPD and hemodialysis patients. *Kidney Int* 1994; 45: 890-896.
 - 85 Leeuwenberg JFM, Mat O, Abramowicz D, Gastaldello R, Tielemans C, Buurman WA: Increased plasma levels of soluble tumor necrosis factor-receptors in uraemic patients: effects of dialysis, type of membrane, and anticoagulation method. *Nephrol Dial Transplant* 1994; 9: 1125-1129.
 - 86 Schindler R, Marra MN, McKelligon BM, Lonnemann G, Schulzeck P, Schulze M, Oppermann M, Shaldon S: Plasma levels of bactericidal/permeability-increasing protein (BPI) and lipopolysaccharide-binding protein (LBP) during hemodialysis. *Clin Nephrol* 1993; 40: 346-351.
 - 87 Engelberts I, Francot GJM, Leunissen KML, Haenen B, Ceska M, van der Linden CJ, Buurman WA: Effect of hemodialysis on peripheral blood monocyte tumor necrosis factor- α , interleukin-6, and interleukin-8 secretion in vitro. *Nephron* 1994; 66: 396-403.
 - 88 Leunissen KML, Kooman JP, van Kuijk WHM, van Hooff JP: Bicarbonate dialysis: a review and future perspectives. *Sem Dial* 1994; 7: 186-191.
 - 89 Liang CS, Lowenstein JM: Metabolic control of the circulation. Effects of acetate and pyruvate. *J Clin Invest* 1978; 62: 1029-1038.
 - 90 Hakim RM, Pontzer M-A, Tilton D, Lazarus JM, Gottlieb MN: Effects of acetate and bicarbonate dialysate in stable chronic dialysis patients. *Kidney Int* 1985; 28: 535-540.
-

-
- 91 Herrero JA, Trobo JJ, Torrente J, Torralbo A, Tornero F, Cruceyra A, Coronel F, Barrientos A: Hemodialysis with acetate, DL-lactate and bicarbonate: a hemodynamic and gasometric study. *Kidney Int* 1994; 46: 1167-1177.
 - 92 Leunissen KML, Hoorntje SJ, Fiers H, Dekkers WT, Mulder AW: Acetate versus bicarbonate hemodialysis in critically ill patients. *Nephron* 1986; 42: 146-151.
 - 93 Henrich WL, Woodard TD, Meyer BD, Chappell TR, Rubin LJ: High sodium bicarbonate and acetate hemodialysis: double-blind crossover comparison of hemodynamic and ventilatory effects. *Kidney Int* 1983; 24: 240-245.
 - 94 Velez RL, Woodard TD, Henrich WL: Acetate and bicarbonate hemodialysis in patients with and without autonomic dysfunction. *Kidney Int* 1984; 26: 59-65.
 - 95 Collins A, Keshaviah P, Berkseth R, Ilstrup K, McMichael C, Ebben J: Short efficient hemodialysis with reduced symptoms. *Kidney Int* 1985; 27: 158.
 - 96 Rubin RP: Neurotransmitter substances and hormones. *Pharmacol Rev* 1970; 22: 389-428.
 - 97 Marone C, Beretta-Picolli C, Weidmann P: Acute hypercalcemic hypertension in man: role of hemodynamics, catecholamines and renin. *Kidney Int* 1980; 20: 92-96.
 - 98 Argilès A, Kerr PG, Canaud B, Flavier JL, Mion C: Calcium kinetics and the long-term effects of lowering dialysate calcium concentration. *Kidney Int* 1993; 43: 630-640.
 - 99 Leunissen KML, van den Berg BW, van Hooff JP: Ionized calcium plays a pivotal role in controlling blood pressure during haemodialysis. *Blood Purif* 1989; 7: 233-239.
 - 100 Henrich WL, Hunt JM, Nixon JV: Increased ionized calcium and left ventricular contractility during hemodialysis. *N Engl J Med* 1984; 310: 19-23.
 - 101 Maynard JC, Cruz C, Kleerekoper M, Levin NW: Blood pressure response to changes in serum ionized calcium during hemodialysis. *Ann Intern Med* 1986; 104: 358-361.
 - 102 Lang RM, Fellner SK, Neumann A, Bushinsky DA, Borow KM: Left ventricular contractility varies directly with blood ionized calcium. *Ann Intern Med* 1988; 108: 524-529.
 - 103 Fellner SK, Lang RM, Neumann A, Spencer KT, Bushinsky DA, Borow KM: Physiological mechanisms for calcium-induced changes in systemic arterial pressure in stable dialysis patients. *Hypert* 1989; 13: 213-218.
 - 104 Bergström J: Catecholamines and control of blood pressure during hemodialysis and hemofiltration. *Kidney Int* 1988; 34: S110-S114.
 - 105 Esler MD, Hasking GJ, Willett IR, Leonard PW, Jennings GL: Noradrenaline release and sympathetic nervous system activity. *J Hypert* 1985; 3: 117-129.
 - 106 Chang PC, Kriek E, van der Krogt JA, van Brummelen P: Does regional norepinephrine spillover represent local sympathetic activity? *Hypert* 1991; 18: 56-66.
-

- 107 Zucchelli P, Santoro A, Sturani A, Degli Esposti E, Chiarini C, Zuccalà A: Effects of hemodialysis and hemofiltration on the autonomic control of circulation. *Trans Am Soc Artif Intern Organs* 1984; 30: 163-167.
 - 108 Baldamus CA, Mantz P, Kachel HG, Koch KM, Schoeppe W: Baroreflex in patients undergoing hemodialysis and hemofiltration. *Contrib Nephrol* 1984; 41: 409-414.
 - 109 Vallbo AB, Hagbarth K-E, Torebjörk HE, Wallin BG: Somatosensory, proprioceptive, and sympathetic activity in human peripheral nerves. *Physiol Rev* 1979; 59: 919-957.
 - 110 Converse RL, Jacobsen TN, Toto RD, Jost CMT, Cosentino F, Fouad-Tarazi F, Victor RG: Sympathetic overactivity in patients with chronic renal failure. *N Engl J Med* 1992; 327: 1912-1918.
 - 111 Converse RL, Jacobsen TN, Jost CMT, Toto RD, Grayburn PA, Obregon TM, Fouad-Tarazi F, Victor RG: Paradoxical withdrawal of reflex vasoconstriction as a cause of hemodialysis-induced hypotension. *J Clin Invest* 1992; 90: 1657-1665.
-

Chapter 2

Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences

WHM van Kuijk, AJ Luik, PW de Leeuw,
JP van Hooff, FHM Nieman, HML Habets,
KML Leunissen

(Parts of this chapter will be published in Nephrology Dialysis and Transplantation)

Abstract

Background

The present study was performed to assess the role of the extracorporeal blood temperature in the disparitive cardiovascular response between isolated ultrafiltration and combined ultrafiltration-hemodialysis.

Methods

In twelve stable dialysis patients (21-77 years), blood pressure and heart rate (Finapres) as well as forearm vascular resistance and venous tone (strain-gauge plethysmography) were measured during one hour isolated ultrafiltration and one hour combined ultrafiltration-hemodialysis (bicarbonate, sodium 141 mmol/l) at a fixed ultrafiltration rate of 0.9 l/hr. In addition, blood pressure and heart rate were measured during the following two dialysis hours as well as during 22 post-dialytic hours with ambulatory blood pressure monitors. Serving as his or her own control, each patient was studied at two different dialysate temperatures: 37.5 and 35.0°C.

Results

At a dialysate temperature of 35.0°C, extracorporeal blood cooling during combined ultrafiltration-hemodialysis was comparable to isolated ultrafiltration. The cardiovascular response in isolated ultrafiltration was characterized by a significant increase in both forearm vascular resistance and venous tone while heart rate even decreased. As a result, blood pressure remained unchanged or even increased. In contrast, during combined ultrafiltration-hemodialysis at a dialysate temperature of 37.5°C, the increase in forearm vascular resistance was only small and insignificant while venous tone decreased significantly. Heart rate tended to increase. Combined ultrafiltration-hemodialysis at a dialysate temperature of 35.0°C was also associated with a small increase in forearm vascular resistance. However, venous tone remained stable while heart rate decreased. Arterial blood pressure was significantly higher at a dialysate temperature of 35.0°C only during the last two dialysis hours. There was no rebound drop in blood pressure in the post-dialytic period after cold dialysis.

Conclusion

We conclude that differences in cardiovascular reactivity between isolated ultrafiltration and combined ultrafiltration-hemodialysis are only partially explained by differences in the extracorporeal blood temperature. In addition, especially venous reactivity is improved by lowering the dialysate temperature. Cold dialysis is a safe method to improve hemodynamic stability in combined ultrafiltration-hemodialysis without adverse effects on post-dialytic blood pressure.

Introduction

Symptomatic hypotension is still the most important complication in dialysis therapy and although our insight in the underlying mechanisms has increased substantially during the past two decades, many controversies still exist. One important problem that remains unresolved are the differences in cardiovascular reactivity during different forms of dialysis therapy^{1,2}.

Under normal conditions, the cardiovascular response to a decline in plasma volume mainly includes an increase in both total peripheral resistance and venous tone which results in a centralisation of blood volume, thus maintaining adequate cardiac filling and output^{3,4}. It is well recognised that these vascular responses are well preserved when fluid is withdrawn during either hemofiltration or isolated ultrafiltration while during combined ultrafiltration-hemodialysis vascular reactivity is impaired^{5,6}. However, as only few comparative studies have yet been performed, it is unknown which factors are actually responsible for impaired vascular reactivity during combined ultrafiltration-hemodialysis in comparison to hemofiltration and isolated ultrafiltration.

It has been shown by many investigators that blood pressure stability during combined ultrafiltration-hemodialysis can be improved by using colder dialysate ($\pm 35.0^{\circ}\text{C}$)^{7,8} which is probably mediated through a better increase in total peripheral resistance⁹⁻¹¹ and cardiac contractility¹² in association with higher levels of plasma noradrenaline^{11,13}. However, only Maggiore and co-workers studied blood pressure stability in relation to the extracorporeal blood temperature, comparing combined ultrafiltration-hemodialysis with both hemofiltration and isolated ultrafiltration. Measuring mean arterial blood pressure and heart rate, they found that changes in the two parameters were almost comparable between treatments at equivalent changes in the extracorporeal blood temperature and concluded that differences in cardiovascular stability between treatments are mainly due to opposite changes in the extracorporeal blood temperature^{14,15}.

The present study was performed to further assess the relative role of the extracorporeal blood temperature in the differences in cardiovascular stability between combined ultrafiltration-hemodialysis and isolated ultrafiltration. In order to make a more pathophysiological comparison possible, changes in blood pressure and heart rate as well as changes in forearm vascular resistance and venous tone were measured during isolated ultrafiltration and during both warm and cold combined ultrafiltration-hemodialysis. In addition, blood pressure was also measured in the post-dialytic period to study whether cold dialysis resulted in rebound hypotension.

Subjects and methods

Subjects and dialysis

Twelve patients (3 women, 9 men) on chronic intermittent hemodialysis signed an informed consent for participation in the study that had been approved by the Ethics Committee of the Maastricht University Hospital. The patient group had a mean age of 51 years (range 21-77) and the mean time on hemodialysis was 3.0 years (range 0.1-21). They were all stable patients who rarely suffered from intradialytic hypotension. Exclusion criteria were severe coronary (NYHA II or more) or valvular heart disease, compromised left ventricular function (ejection fraction 30% or less) and diabetes mellitus. By adjusting fluid intake, patients were instructed to achieve a predialysis weight at two kilograms above their dry-weight which was estimated by echography of the inferior caval vein¹⁶. All vasoactive medication was stopped 48 hours before the study which resulted in hypertension in patient number 11. In addition, patients were not allowed to smoke or drink caffeine containing beverages during 12 hours preceding the study.

All patients were studied on the regular day of his or her dialysis schedule after an overnight fast. The first two dialysis hours consisted of one hour isolated ultrafiltration (i-UF) and one hour combined ultrafiltration-hemodialysis (UF+HD) (protocol A) followed by two hours UF+HD (protocol B). Each patient served as his or her own control and was studied once at a dialysate temperature of 37.5 and once at a dialysate temperature of 35.0°C, respectively warm UF+HD and cold UF+HD. The determination of the dialysate temperature in cold UF+HD was based on preliminary experiments. At a dialysate temperature of 35.0°C, blood in the venous line had the same temperature during UF+HD as measured during i-UF, indicating a comparable degree of extracorporeal blood cooling. To assure that differences in cardiovascular reactivity between i-UF and UF+HD would not arise from large differences in plasma volume, patients were randomly selected to start with either UF (UF-UF+HD, n=6) or UF+HD (UF+HD-UF, n=6).

Dialysis was performed with a Gambro AK-100 module using hemophane membranes (GFS-16; Gambro, Sweden). The composition of dialysate was: sodium 141 mmol/l, potassium 2.0 mmol/l, bicarbonate 34 mmol/l, acetate 3 mmol/l, calcium 1.75 mmol/l, magnesium 0.5 mmol/l and chloride 108 mmol/l. Blood flow was 250 ml/min and dialysate flow was 500 ml/min.

Methods

The study started with the insertion of the dialysis needles after which patients were in supine rest for 30 minutes. All studies were performed in a temperature controlled room at a room temperature of about 23°C. The luminal blood

temperature was measured in the arterial and venous line with needle thermometers (Hewlett-Packard 78214C temperature monitor). Connectors with a built-in adaptor were placed in the arterial and venous line under sterile conditions at a distance of about 15 centimetres from the patient. There was no direct contact between blood and thermometer.

Protocol A

Protocol A was performed during the first two dialysis hours consisting of sequential i-UF and UF+HD at a fixed ultrafiltration rate of 0.9 l/hr. All hemodynamic measurements were performed before (pre-treatment) and every 15 minutes during dialysis.

Arterial blood pressure (BP) and heart rate (HR) were measured with the Finapres method (Finapres, Ohmeda 2300, Lameris, The Netherlands). The mean values of 3 minutes were calculated. With the Finapres device, arterial blood pressure and heart rate are measured beat to beat at zero transmural pressure by the use of a small finger cuff that is equipped with an infrared photoplethysmograph¹⁷. The Finapres cuff was applied to the third finger.

Vascular reactivity was studied at the non-fistula arm that was positioned just above heart level using strain-gauge plethysmography as described by Whitney (Periflow; Janssen Scientific Instruments, Beerse, Belgium)¹⁸. Changes in forearm vascular and sympathetic tone have been shown to correlate well with the systemic hemodynamic and sympathetic response during unloading of cardiopulmonary or baroreceptors¹⁹⁻²¹. An inflatable cuff was applied to the upper arm while the mercury-filled strain gauge was positioned at the thickest part of the forearm. In addition, an antecubital vein was cannulated (Venflon, 1 mm $\varnothing$) for the recording of direct intravenous pressure (Hewlett-Packard 78205C pressure monitor). The pressure dome was positioned at the level of the right atrium approximately 5 centimetres under the sternal angle.

Changes in venous tone (VT) (active venous constriction) were estimated by recording the pressure/volume ratio (mmHg/ml/100ml) at a cuff pressure of 40 mmHg. Pressure was applied until both intravenous pressure and arm volume had reached a plateau. During venous occlusion and after obtaining a plateau, the intravenous distending pressure is determined by the cuff pressure. Therefore, changes in venous tone are primarily related to volume changes. In addition, at a transmural pressure of about 40 mmHg, changes in volume only reflect active venous constriction whereas the contribution of passive venous constriction as a result of elastic recoil of the venous wall is negligible²². ▲ Pressure and ▲ volume were estimated directly after the deflation of the upper arm cuff to minimize the influence of capillary filtration on assessing ▲ volume. The coefficient of variation

of consecutive measurements is 11.9%⁶.

The measurement of forearm blood flow is based on the measurement of the initial volume change during three heart beats after initiating venous occlusion. Measurement of forearm blood flow by venous occlusion plethysmography has been shown to correlate well with electromagnetic flow measurements²³. The coefficient of variation of consecutive flow measurements is 14%²⁴. To measure forearm blood flow (FBF), the upper arm cuff was rapidly inflated, 5 msec after the R-top of the electrocardiogram, at a cuff pressure of 50 mmHg. The inflation/deflation ratio was 3:2 heart beats. Due to the presence of the Finapres cuff, the hand circulation was not occluded. Forearm blood flow was measured during four minutes and estimated with a computerized integrator. The mean value of the last two minutes was calculated. Forearm vascular resistance (FVR) was calculated by dividing mean arterial pressure by forearm blood flow (mmHg/ml/100ml/sec).

Prior to dialysis and after both 1 hour i-UF and 1 hour UF+HD, blood was withdrawn for the estimation of: sodium, potassium, ionized calcium, total CO₂ and osmolality (Wescor, vapour pressure osmometer). In addition, blood was withdrawn in ice-chilled tubes and directly centrifuged at a temperature of 4°C for the estimation of plasma catecholamines (HPLC, fluorescence detection)²⁵. Changes in plasma volume were calculated with the serial hematocrit (Hct) method using the formula²⁶:

$$\Delta \text{plasma volume} = \frac{100}{(100 - \text{Hct}_{\text{before}})} \times \frac{100 (\text{Hct}_{\text{before}} - \text{Hct}_{\text{after}})}{\text{Hct}_{\text{after}}}$$

Protocol B

Nine patients participated in protocol B that was performed during the last two dialysis hours as well as during 22 post-dialysis hours. The dialysate temperature was maintained at the initial level while patients were ultrafiltrated to their dry-weight. Arterial BP and HR were measured non-invasively with Spacelabs 90207 oscillometric ambulatory blood pressure monitors²⁷. BP was measured every 15 minutes during daytime (7 a.m.-11 p.m.) and every 30 minutes during nighttime (11 p.m.-7 a.m.).

Statistical analysis

Changes in hemodynamic parameters within each treatment as well as differences between treatments were analysed by repeated measurements MANOVA (SPSS-PC version)²⁸. If the sphericity of the variance-covariance matrix appeared to be violated, degrees of freedom in the univariate MANOVA tests were corrected by

the Greenhouse-Geisser epsilon to avoid Type 1 error in testing the F-ratio. All laboratory parameters were analysed by Friedmans ANOVA and if appropriate, by the Wilcoxon's signed-rank test. All values are expressed as mean $\pm$ SD.

Results

Blood temperature

The uncorrected blood temperatures are presented in figure 2.1a-d. During i-UF, blood cooled down during circulation through the extracorporeal circuit by about 1.7°C. In contrast, blood had a comparable temperature in the arterial and venous line during warm UF+HD. Cold UF+HD resulted in a comparable degree of blood cooling in the extracorporeal circuit as compared to i-UF. Only during warm UF+HD preceded by i-UF, the temperature of blood in the arterial line increased from $35.7 \pm 0.6^\circ\text{C}$ to $36.0 \pm 0.5^\circ\text{C}$ ($p < 0.05$). During three consecutive hours of warm and cold UF+HD, the temperature of blood in the arterial line increased by $0.7 \pm 0.1^\circ\text{C}$ ($p < 0.05$) and $0.4 \pm 0.2^\circ\text{C}$ ($p < 0.05$) respectively while the temperature of blood in the venous line remained unchanged.

Figure 2.1a Arterial (—) and venous (---) blood temperature in UF-UF+HD, 37.5°C; *= $p < 0.05$, arterial blood temperature as compared to $t = 0$ min

Figure 2.1b Arterial (—) and venous (- - -) blood temperature in UF-UF+HD, 35.0°C

Figure 2.1c Arterial (—) and venous (- - -) blood temperature in UF+HD-UF, 37.5°C; * $p < 0.05$, arterial blood temperature as compared to $t = 0$ min

Figure 2.1d Arterial (—) and venous (- - -) blood temperature in UF+HD-UF, 35.0°C

Protocol A

Isolated ultrafiltration followed by combined ultrafiltration-hemodialysis

The hemodynamic data are presented in figure 2.2a+b and table 2.1. The predialysis weights at the start of warm and cold dialysis were 71.3 ± 11.4 and 71.4 ± 10.9 kg respectively ($p=0.75$).

I-UF was associated with an increase in both FVR and VT as well as in MAP while HR decreased. During warm UF+HD, FVR remained unchanged while VT decreased significantly even below the pre-treatment level. In addition, HR increased significantly. During cold UF+HD, both VT and HR remained unchanged with both parameters being significantly different as compared to warm UF+HD. As compared to i-UF, changes in FVR, VT, MAP and HR were significantly different only during warm UF+HD.

Figure 2.2a FVR during UF-UF+HD, 37.5 (—) and 35.0°C (- - -)

Figure 2.2b VT during UF-UF+HD, 37.5 (—) and 35.0°C (- - -)

Table 2.1 Changes in hemodynamic parameters during UF-UF+HD

	DIALYS TEMP: 37.5 °C			DIALYS TEMP: 35.0 °C			w-HD vs c-HD
	UF	UF+HD	p	UF	UF+HD	p	p
MAP	+12 (7) ^a	-4 (7)	0.006	+9 (8) ^a	-2 (9)	0.07	0.62
HR	-5 (5) ^b	+8 (8) ^c	0.001	-3 (5)	-2 (5)	0.81	0.026
FVR	+1488 (1094) ^b	-280 (1036)	0.013	+1527 (1018) ^a	-38 (1421)	0.09	0.46
VT	+8.2 (7.0) ^b	-9.5 (8.2) ^d	0.02	+7.9 (9.6)	-2.7 (4.8)	0.08	0.037

MAP (mmHg); HR (beats/min); FVR (mmHg/ml/100ml/sec); VT (mmHg/ml/100ml); ^a p < 0.01, change vs pre-treatment value; ^b p < 0.05, change vs pre-treatment value; ^c p < 0.01, change vs end-UF value; ^d p < 0.05, change vs end-UF value

Table 2.2 Laboratory parameters during UF-UF+HD

	DIALYS TEMP: 37.5 °C			DIALYS TEMP: 35.0 °C		
	Pre	UF	UF+HD	Pre	UF	UF+HD
delta PV		-11.9 (7.8)	-5.5 (6.9)		-12.6 (3.9)	-6.2 (3.0) ^b
Sodium	136.7 (4.5)	137.5 (4.7)	139.3 (3.4) ^a	137.2 (3.1)	137.3 (2.9)	139.5 (2.4) ^{a,b}
Osmolality	299.2 (13.5)	294.7 (8.1)	286.8 (11.2) ^{a,b}	299.0 (8.9)	306.5 (13.2)	290.2 (13.1) ^b
total CO ₂	23.8 (1.3)	22.8 (1.8) ^a	26.7 (1.7) ^{a,b}	23.0 (1.7)	22.9 (1.5)	26.1 (1.5) ^{a,b}
ionized Ca	1.23 (0.12)	1.25 (0.15)	1.36 (0.11) ^{a,b}	1.17 (0.14)	1.21 (0.14)	1.30 (0.13) ^{a,b}
Noradrenaline	1.90 (0.86)	2.81 (0.91) ^a	2.10 (0.79) ^b	1.70 (0.79)	2.51 (0.88) ^a	2.25 (0.68) ^a
Adrenaline	0.20 (0.11)	0.24 (0.09)	0.22 (0.10)	0.21 (0.11)	0.24 (0.13)	0.21 (0.12)

PV (%); Sodium, total CO₂ and ionized calcium (mmol/l); Osmolality (mOsm/kg);

Catecholamines (nmol/l); ^a p < 0.05, compared to pre-treatment value; ^b p < 0.05, compared to UF

The laboratory data are presented in table 2.2. Changes in plasma volume were comparable between the different treatment modalities. Only in i-UF followed by cold UF+HD, the decline in plasma volume was significantly larger during i-UF. Probably as a result of a higher pre-treatment level, plasma ionized calcium was significantly higher after warm UF+HD as compared to cold UF+HD. Plasma noradrenaline increased significantly during i-UF. Although there was no significant difference in plasma noradrenaline between warm and cold UF+HD, plasma noradrenaline remained elevated only during cold UF+HD. Plasma adrenaline remained unchanged during the different treatment modes.

Table 2.3 Changes in hemodynamic parameters during UF+HD-UF

	DIALYS TEMP: 37.5°C			DIALYS TEMP: 35.0°C			w-HD vs c-HD
	UF+HD	UF	p	UF+HD	UF	p	p
MAP	-1 (4)	+5 (9)	0.1	+0 (16)	+4 (7)	0.38	0.87
HR	+2 (4)	-3 (4)	0.37	-5 (2) ^a	+1 (4)	0.84	0.001
FVR	+511 (860)	+962 (486)	0.75	+921 (885)	+159 (625)	0.68	0.33
VT	-3.1 (2.6) ^b	+5.2 (2.7) ^d	0.30	+0.51 (3.4)	-0.1 (4.2)	0.48	0.031

MAP (mmHg); HR (beats /min); FVR (mmHg/ml/100ml/sec); VT (mmHg/ml/100ml); ^a p < 0.01, change vs pre-treatment value; ^b p < 0.05, change vs pre-treatment value; ^c p < 0.01, change vs end-UF+HD value; ^d p < 0.05, change vs end-UF+HD value

Table 2.4 Laboratory parameters during UF+HD-UF

	DIALYS TEMP: 37.5°C			DIALYS TEMP: 35.0°C		
	Pre	UF+HD	UF	Pre	UF+HD	UF
delta PV		-9.5 (4.9)	-11.8 (2.8)		-8.7 (2.4)	-12.4 (4.4)
Sodium	136.5 (1.5)	139.0 (3.0)	138.8 (1.2) ^a	138.2 (1.2)	140.0 (1.3) ^a	139.8 (1.0) ^a
Osmolality	298.3 (9.6)	288.7 (9.7) ^a	293.2 (8.0) ^a	307.0 (14.3)	298.3 (12.5) ^a	294.7 (12.2) ^a
total CO ₂	23.7 (4.5)	26.7 (2.5) ^a	25.2 (3.5) ^{a,b}	23.9 (3.3)	27.0 (2.6) ^a	25.6 (2.8) ^{a,b}
ionized Ca	1.19 (0.11)	1.29 (0.09) ^a	1.28 (0.10) ^a	1.19 (0.11)	1.30 (0.12) ^a	1.30 (0.12) ^a
Noradrenaline	1.05 (0.36)	1.10 (0.27)	1.32 (0.42)	1.03 (0.45)	1.11 (0.40)	1.58 (0.48) ^{a,b}
Adrenaline	0.19 (0.15)	0.18 (0.10)	0.18 (0.11)	0.18 (0.10)	0.14 (0.07)	0.20 (0.08) ^b

PV (%); Sodium, total CO₂ and ionized calcium (mmol/l); Osmolality (mOsm/kg);

Catecholamines (nmol/l); ^a p < 0.05, compared to pre-treatment value; ^b p < 0.05, compared to UF+HD

Combined ultrafiltration-hemodialysis followed by isolated ultrafiltration

The hemodynamic data are presented in table 2.3. and figure 2.3a+b. The predialysis weights were comparable at the start of warm and cold dialysis: 78.3±15.2 and 78.2±15.8 kg (p=1.00).

Warm UF+HD was associated with a significant decrease in VT while FVR tended to increase (p>0.05). Both MAP and HR remained unchanged. Cold UF+HD was associated with significantly higher VT while HR decreased significantly. After both warm and cold UF+HD, changes in the different hemodynamic parameters during i-UF did not reach statistical significance. Only VT increased significantly during i-UF after warm UF+HD. There were no significant differences between i-UF and both warm and cold UF+HD.

The laboratory data are presented in table 2.4. Changes in plasma volume were comparable between the different treatment modes. There were no significant differences in the different laboratory parameters between warm and cold UF+HD. Both plasma noradrenaline and adrenaline increased significantly only during i-UF after cold UF+HD.

Figure 2.3a FVR during UF-UF+HD, 37.5 (—) and 35.0°C (---)

Figure 2.3b VT during UF+HD-UF, 37.5 (—) and 35.0°C (---)

Sub-analysis: isolated ultrafiltration (UF-UF+HD) compared to warm and cold combined ultrafiltration-hemodialysis (UF+HD-UF)

A sub-analysis of the hemodynamic data was performed comparing changes from baseline during i-UF (first hour in UF-UF+HD) with changes from baseline during warm and cold UF+HD (first hour in UF+HD-UF). In i-UF, the mean values of the hours preceding warm and cold UF+HD were used for analysis.

There were no statistically significant differences in MAP and FVR between i-UF and both warm and cold UF+HD. However, changes in HR and VT were significantly different comparing i-UF with warm UF+HD (respectively -4 ± 5 vs $+2 \pm 4$ beats/min, $p < 0.05$ and $+8.0 \pm 8.2$ vs -3.1 ± 2.6 mmHg/ml/100ml, $p < 0.01$). Furthermore, while changes in HR were comparable (respectively -4 ± 5 vs -5 ± 2 beats/min, $p > 0.05$), changes in VT were significantly different between UF and cold UF+HD (respectively $+8.0 \pm 8.2$ vs $+0.5 \pm 3.4$ mmHg/ml/100ml, $p < 0.05$).

Protocol B

The predialysis weights in warm and cold UF+HD in this sub-group of nine patients were respectively 75.5 ± 11.8 and 75.2 ± 12.1 kg (ns). The ultrafiltration rates were respectively 0.4 ± 0.3 and 0.3 ± 0.2 l/hr (ns) while the post-dialysis weights were respectively 73.0 ± 11.7 and 73.2 ± 11.7 kg (ns). MAP was significantly higher in cold UF+HD during both dialysis hours (mean difference $+8 \pm 8$ mmHg, $p < 0.05$) (figure 2.4a). While differences in diastolic BP did not reach statistical significance systolic BP was significantly higher during the last dialysis hour ($+11 \pm 12$ mmHg, $p < 0.05$). In addition, HR was significantly lower in cold UF+HD during the first dialysis hour (-10 ± 9 beats/min, $p < 0.05$) (figure 2.4b). Although MAP tended to be higher during the first four post-dialysis hours, there were no significant differences in either BP or HR in the post-dialytic period between warm and cold UF+HD. None of the patients developed reactive hypotension after cold UF+HD.

Figure 2.4a Mean arterial BP during and after UF+HD, 37.5 (—) and 35.0°C (- - -) ; *= $p < 0.05$

Figure 2.4b Heart rate during and after UF+HD, 37.5°C (—) and 35.0°C (- - -)

* $p < 0.05$

Discussion

In the present study, vascular reactivity was measured during both i-UF and warm UF+HD (37.5°C) as well as during cold UF+HD (35.0°C) to study the role of the extracorporeal blood temperature in the disparate hemodynamic response between isolated ultrafiltration and combined ultrafiltration-hemodialysis. In addition, the effect of cold dialysis on post-dialytic blood pressure was studied.

The main conclusion in the present study is that differences in the extracorporeal blood temperature are only partially responsible for differences in vascular reactivity between i-UF and UF+HD. In addition, especially venous reactivity is improved during UF+HD when a comparable degree of blood cooling as during i-UF is achieved. Finally, cold dialysis does not result in rebound hypotension in the post-dialytic period.

During i-UF blood cooled down in the extracorporeal circuit by about 1.7°C. In contrast, UF+HD at a dialysate temperature of 37.5°C blood was associated with a comparable temperature of blood in the arterial and venous line, indicating that extracorporeal blood cooling was exactly counterbalanced by heat transfer from dialysate to blood. At a dialysate temperature of 35.0°C, extracorporeal blood cooling was comparable between UF+HD and i-UF.

The increase in temperature of blood in the arterial line by 0.3°C during one hour warm UF+HD (in UF-UF+HD) suggests that in protocol A, warm UF+HD was already associated with net heat gain. Three hours UF+HD resulted in an even larger increase in blood temperature, especially at a dialysate temperature of 37.5°C. These results are in agreement with studies in which an increase in core body temperature or arterial blood temperature in the range of 0.7 to 1.0°C was

demonstrated after four hours UF+HD at dialysate temperatures ranging between 37.0 and 38.0°C^{7,14,15}. In our experiments, however, heat gain is probably not only explained by net heat load at the dialyser level as even cold UF+HD was associated with a small but significant increase in the arterial blood temperature. Possible mechanisms that might be important in this respect are a decline in heat loss due to constriction of the skin vascular bed or an increase in internal heat generation due to the induction of a pyrogenic reaction, reflecting an interaction between blood and the non-sterile dialysate²⁹.

With respect to the cardiovascular and laboratory data, both groups (UF-UF+HD and UF+HD-UF) were analyzed separately as statistical analysis revealed a significant effect of the specific order on some cardiovascular data. Only in the sub-analysis, a between-group comparison was made. In i-UF followed by warm UF+HD, changes in all four hemodynamic parameters were significantly different between the two treatments showing an adequate increase in FVR and VT during i-UF with vascular reactivity being impaired during UF+HD. However, at a dialysate temperature of 35.0°C these differences did not reach statistical significance. In UF+HD followed by i-UF, differences in hemodynamics between the two treatments were less pronounced irrespective of the dialysate temperature. The data suggest that vascular reactivity was disturbed during i-UF when preceded by UF+HD, especially after cold UF+HD. Although this might be a physiologic adaptation to a more stable circulation during cold UF+HD, vascular reactivity during i-UF might also be disturbed as a result of ongoing effects of the first UF+HD hour. Possible factors in this respect are the decrease in plasma osmolality³⁰ or exposition to dialysate derived contaminants³¹. In order to exclude order effects in comparing i-UF with UF+HD, a sub-analysis was performed showing a statistically significant difference in VT between the two treatments at both dialysate temperatures. Differences in FVR were not significant.

The use of colder dialysate mainly affected venous reactivity. In both treatment sequences, VT was significantly higher during cold UF+HD as compared to warm UF+HD. Furthermore, in contrast to warm UF+HD, HR decreased during cold UF+HD which is probably related to a better maintenance of VT, improving venous return, together with an increase in left ventricular contractility during cold UF+HD as observed by Levy et al.¹². A delayed decrease in plasma α -ANP during cold UF+HD as compared to warm UF+HD as recently reported by Hegbrant et al. is also in agreement with our observations³². We found no significant difference in FVR during UF+HD at the two dialysate temperatures although in both treatment sequences, mean FVR did tend to be higher during cold UF+HD.

In protocol A, cold UF+HD was not associated with a significantly higher blood pressure as compared to warm UF+HD. However, when patients were further ultrafiltrated to their dry-weight in protocol B arterial blood pressure was significantly higher during cold dialysis while HR continued to be lower as observed in protocol A. Special attention was given to the post-dialytic period as rewarming of patients after cold dialysis might lead to rebound hypotension. However, none of the patients developed post-dialytic hypotension while blood pressure even tended to be somewhat higher after cold UF+HD up to four hours after dialysis.

Plasma noradrenaline increased significantly only during i-UF in our experiments which might indicate that UF+HD is associated with an impairment of sympathetic reactivity. Furthermore, our results suggest that efferent sympathetic activity was unaffected by the extracorporeal blood temperature as plasma noradrenaline was not significantly different comparing warm with cold UF+HD, which is at variance with the results of prior studies^{13,15}. Moreover, these results make it less likely that improved venous reactivity during cold UF+HD is explained by an increase in the overall sympathetic output from the vasomotor centre^{33,34}. However, we should bear in mind that plasma noradrenaline is only a weak indicator of sympathetic activity³⁵. A possible explanation for improved venous reactivity in cold UF+HD might be a more specific decrease in venous capacitance of the skin vascular bed³⁴. In addition, it has been shown that the vascular response to adrenergic stimulation increases when the peripheral blood temperature decreases³⁶.

Overall, our results confirm the results obtained by other investigators^{5,6} showing that both vascular and sympathetic reactivity are impaired during combined ultrafiltration-hemodialysis as compared to isolated ultrafiltration. In addition, they show that at equivalent changes in the extracorporeal blood temperature, especially venous reactivity is still better during isolated ultrafiltration as compared to combined cold ultrafiltration-hemodialysis. Therefore, we conclude that disparity in vascular reactivity between isolated ultrafiltration and combined ultrafiltration-hemodialysis is only partially explained by differences in the extracorporeal blood temperature. Possible factors that might contribute to impaired vascular reactivity during combined ultrafiltration-hemodialysis and that cannot be excluded in the present study are a decrease in plasma osmolality³⁰ or the induction of cytokine production by monocytes in the presence of dialysate derived contaminants^{31,37}. Differences in plasma ionized calcium are probably not responsible for disparity in vascular reactivity as ionized calcium increased only during UF+HD in the present study. Finally we conclude, that lowering the dialysate temperature to 35.0°C does not lead to rebound hypotension in the post-dialytic period. Thus, lowering the dialysate temperature is in this respect a safe method to improve hemodynamic stability in the clinical situation.

References

- 1 Henrich WL: Hemodynamic instability during hemodialysis. *Kidney Int* 1986; 30: 605-612.
 - 2 Daugirdas JT: Dialysis hypotension: A hemodynamic analysis. *Kidney Int* 1991; 39: 233-246.
 - 3 Rothe CF: Venous system: Physiology of capacitance vessels; in Shepherd JT, Abboud FM (ed.): *Handbook of Physiology*. Bethesda, American Physiological Society, 1983, 3, Peripheral circulation and organ blood flow, pp. 397-452
 - 4 Greenway CV, Lauth WW: Blood volume, the venous system, preload, and cardiac output. *Can J Physiol Pharmacol* 1986; 64: 383-387.
 - 5 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
 - 6 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
 - 7 Lindholm T, Thysell H, Yamamoto Y, Forsberg B, Gullberg CA: Temperature and vascular stability in hemodialysis. *Nephron* 1985; 39: 130-133.
 - 8 Marcén R, Quereda C, Orofino L, Lamas S, Teruel JL, Matesanz R, Ortuño J: Hemodialysis with low-temperature dialysate: A long-term experience. *Nephron* 1988; 49: 29-32.
 - 9 Coli U, Landini S, Lucatello S, Fracasso A, Morachiello P, Righetto F, Scanferla F, Onesti G, Bazzato G: Cold as cardiovascular stabilizing factor in hemodialysis: hemodynamic evaluation. *Trans Am Soc Artif Intern Organs* 1983; 29: 71-75.
 - 10 Bazzato G, Coli U, Landini S, Lucatello S, Fracasso A, Morachiello P, Righetto F, Scanferla F: Temperature monitoring in dialysis-induced hypotension. *Kidney Int* 1985; 28: S161-S165.
 - 11 Jost CMT, Agarwal R, Khair-El-Din T, Grayburn PA, Victor RG, Henrich WL: Effects of cooler temperature dialysate on hemodynamic stability in "problem" dialysis patients. *Kidney Int* 1993; 44: 606-612.
 - 12 Levy FL, Grayburn PA, Foulks CJ, Brickner ME, Henrich WL: Improved left ventricular contractility with cool temperature hemodialysis. *Kidney Int* 1992; 41: 961-965.
 - 13 Mahida BH, Dumler F, Zasuwa G, Fleig G, Levin NW: Effect of cooler dialysate on serum catecholamines and blood pressure stability. *Trans Am Soc Artif Intern Organs* 1983; 24: 384-389.
 - 14 Maggiore Q, Pizzarelli F, Sisca S, Zoccali C, Parlongo S, Nicolò F, Creazzo G: Blood temperature and vascular stability during hemodialysis and hemofiltration. *Trans Am Soc Artif Intern Organs* 1982; 28: 523-527.
-

- 15 Maggiore Q, Pizzarelli F, Zoccali C, Sisca S, Nicolò F: Influence of blood temperature on vascular stability during hemodialysis and isolated ultrafiltration. *Int J Artif Organs* 1985; 8: 175-178.
- 16 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
- 17 Penàz J: Photoelectric measurement of blood pressure, volume and flow in the finger. *Digest 10th Int Conf Med Biol Engng* 104, Dresden 1973.
- 18 Whitney RJ: The measurement of volume changes in human limbs. *J Physiol* 1953; 121: 1-27.
- 19 Berdeaux A, Duranteau J, Pussard E, Edouard A, Giudicelli J-F: Baroreflex control of regional vascular resistances during simulated orthostatism. *Kidney Int* 1992; 41: S29-S33.
- 20 Tripathi A, Mack G, Nadel ER: Peripheral vascular reflexes elicited during lower body negative pressure. *Aviat Space Environ Med* 1989; 60: 1187-1193.
- 21 Baily RG, Prophet SA, Shenberger JS, Zelis R, Sinoway LI: Direct neurohumoral evidence for isolated sympathetic nervous system activation to skeletal muscle in response to cardiopulmonary baroreceptor unloading. *Circ Res* 1990; 66: 1720-1728.
- 22 Öberg B: The relationship between active constriction and passive recoil at various distending pressures. *Acta Physiol Scand* 1967; 71: 233-247.
- 23 Jageneau AH: Noninvasive diagnosis in the management of cardiovascular diseases. Amsterdam, New York, London, North Holland Biomedical Press, 1981, 169-229.
- 24 Houben AJHM (1993): Early (micro)circulatory haemodynamic changes in type I diabetes mellitus. Ph.D. Dissertation, University of Limburg, Maastricht, The Netherlands. 144 p.
- 25 van der Hoorn FA, Boomsma F, Man in 't Veld AJ, Schalkenkamp MA: Determination of catecholamines in human plasma by high-performance liquid chromatography: comparison between a new method with fluorescence detection and an established method with electrochemical detection. *J Chromatogr* 1989; 487: 17-28.
- 26 van Beaumont W: Evaluation of hemoconcentration from hematocrit measurements. *J of Appl Physiol* 1972; 32: 712-713.
- 27 O'Brien E, Mee F, Atkins N, O'Mally K: Accuracy of the Spacelabs 90207 determined by the British Hypertension Society protocol. *J Hypert* 1991; 9: 573-574.
- 28 Stevens J: Chapter 13.7: Should we use the univariate or the multivariate approach?; in: *Applied multivariate statistics for the social sciences*. Hillsdale, New Jersey, Lawrence Erlbaum Associates, Publishers, 1992, pp. 454-456.
- 29 Dinarello CA: Pathogenesis of fever during hemodialysis. *Contrib Nephrol* 1983; 36: 90-99.
- 30 Henrich WL, Woodard TD, Blachley JD, Gomez-Sanchez C, Pettinger W, Cronin RE: Role of osmolality in blood pressure stability after dialysis and ultrafiltration. *Kidney Int* 1980; 18: 480-488.

-
- 31 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
 - 32 Hegbrant J, Thysell H, Mårtensson L, Nielsen AL, Lindberg FL: Delayed decrease in plasma levels of atrial natriuretic peptide during cold hemodialysis. *Nephron* 1994; 68: 427-432.
 - 33 Johnson JM: Nonthermoregulatory control of human skin blood flow. *J of Appl Physiol* 1986; 61: 1613-1622.
 - 34 Rowell BR: Cardiovascular adjustment to thermal stress; in Shepherd JT, Abboud FM (ed.): *Handbook of Physiology*, sect 2, Cardiovascular system, vol 3, Peripheral circulation and organ blood flow. Bethesda, American Physiological Society, 1983, 3, pp. 967-1023
 - 35 Esler MD, Hasking GJ, Willett IR, Leonard PW, Jennings GL: Noradrenaline release and sympathetic nervous system activity. *J Hypert* 1985; 3: 117-129.
 - 36 Vanhoutte PM: Physical factors of regulation; in Bohr DF, Somlyo AP (ed.): *Handbook of Physiology*, sect 2, Cardiovascular System, vol 2, Vascular smooth muscle. Bethesda, American Physiological Society, 1980, 2, pp. 443-474
 - 37 Beasley D, Brenner BM: Role of nitric oxide in hemodialysis hypotension. *Kidney Int* 1992; 42: S96-S100.
-

Chapter 3

Vascular reactivity during combined ultrafiltration- hemodialysis: influence of dialysate sodium

WHM van Kuijk, JJM Wirtz, W Grave, F de Heer,
PPCA Menheere, JP van Hooff, KML Leunissen

(accepted for publication Nephrology Dialysis and Transplantation)

Abstract

Background

It is well known that vascular reactivity is impaired during combined ultrafiltration-hemodialysis as compared to isolated ultrafiltration and hemofiltration which might be related to differences in plasma osmolality. Therefore, vascular reactivity was studied during combined ultrafiltration-hemodialysis in relation to sodium related differences in plasma osmolality/tonicity.

Methods

With each patient serving as his or her own control, nine stable dialysis patients (23-71 years) were studied during two hours of combined ultrafiltration-hemodialysis (bicarbonate, UF-rate 1.0 l/hr) at two different dialysate sodium concentrations: 134 and 144 mmol/l. Before dialysis as well as every twenty minutes during dialysis, blood pressure (Dinamap), heart rate (EKG) and forearm vascular resistance and venous tone (strain-gauge plethysmography) were measured. Relative blood volume was monitored continuously with an optical reflection method (Haemoguard 2000) while before and after dialysis blood was obtained for the estimation of plasma prostaglandin E_2 (PGE_2).

Results

High-sodium dialysis resulted in a significantly higher post-dialysis plasma sodium concentration (139.9 vs 135.0 mmol/l; $p < 0.01$) while the decrease in relative blood volume was significantly smaller as compared to low-sodium dialysis (-8.4 vs -18.4%; $p < 0.01$). There were no significant differences in the different hemodynamic parameters between the two treatment modalities. Both high- and low-sodium dialysis were associated with a significant increase in forearm vascular resistance while venous tone remained unchanged. Although there was no significant difference in plasma PGE_2 between the two treatment modalities, PGE_2 increased significantly only during low-sodium dialysis. We found no relation between changes in PGE_2 and vascular reactivity.

Conclusions

We conclude that vascular reactivity during combined ultrafiltration-hemodialysis is not directly influenced by sodium related changes in plasma tonicity. Although not studied directly, the reported improved hemodynamic stability with high-sodium dialysis is probably only mediated through a better preservation of plasma volume. Finally, an increase in plasma PGE_2 as observed during low-sodium dialysis does not lead to a decrease in vascular tone.

Introduction

Hemodynamic instability with symptomatic hypotension is still the most important complication in dialysis therapy. During ultrafiltration, maintenance of blood pressure is related to both optimal preservation of plasma volume as well as to adequate cardiovascular compensating mechanisms to buffer the decline in plasma volume^{1,2}. In this respect it is well recognised that vascular reactivity is impaired during combined ultrafiltration-hemodialysis as compared to both isolated ultrafiltration and hemofiltration³⁻⁵. However, the mechanisms responsible for this impaired vascular response are still not well understood. In a recent study and in agreement with Maggiore et al. we concluded that impaired vascular reactivity during combined ultrafiltration-hemodialysis as compared to isolated ultrafiltration is only partially explained by differences in the extracorporeal blood temperature which indicates that additional factors should play a role^{6,7}.

Traditionally, the relatively large decrease in plasma osmolality during combined ultrafiltration-hemodialysis as result of hypertonic fluid loss is considered to disturb the hemodynamic response during this treatment^{8,9}. It has been shown that counterbalancing changes in osmolality by the use of higher dialysate sodium concentrations or by the infusion of hypertonic mannitol result in a better hemodynamic stability¹⁰⁻¹². Although this improvement is at least partially explained on the basis of a better preservation of plasma volume, results obtained by Henrich et al. suggest a direct effect of plasma osmolality on vascular reactivity¹³. Schultze et al. suggested that increased levels of plasma PGE₂ may contribute to hemodynamic instability in low-sodium dialysis by inducing vasodilation¹⁴. However, the conclusions in both studies are based on measurements of blood pressure and heart rate only.

In the present study therefore, peripheral vascular reactivity was measured during high- and low-sodium dialysis to study the role of sodium related changes in plasma osmolality in the hemodynamic and response during combined ultrafiltration-hemodialysis.

Subjects and methods

Subjects and dialysis

Nine patients on chronic intermittent hemodialysis signed an informed consent for participation in the study that had been approved by the Ethics Committee of the St Laurentius Hospital. All patients were hemodynamically stable patients who rarely suffered from intradialytic hypotension. Exclusion criteria were severe coronary (NYHA II or more) or valvular heart disease, compromised left ventricular

function (ejection fraction 30% or less) and diabetes mellitus.

One woman and eight men were studied with a mean age of 46 years (range 23-71) and a mean time on hemodialysis of 46 months (range 12-53). Patients served as their own control and were studied according to their individual dialysis schedule. By adjusting fluid intake, patients were instructed to achieve a pre-dialysis weight of 2.0 kilograms above their dry-weight that was estimated by echography of the inferior caval vein¹⁵. All vasoactive medication was stopped 48 hours before the study while patients were not allowed to smoke or drink caffeine containing beverages during 12 hours preceding the study. None of the patients used non-steroidal anti-inflammatory drugs.

Measurements were performed during two dialysis treatments consisting of two hours combined ultrafiltration-hemodialysis that only differed in the dialysate sodium concentration: 134 and 144 mmol/l (by measurement of conductivity). The order of the dialysate sodium concentration was randomised while the ultrafiltration rate was 1.0 l/hr in all treatments. Dialysis was performed with a Gambro AK-100 module using hemophane membranes (GFS-16; Gambro, Sweden). The composition of dialysate was: sodium 134 or 144 mmol/l, chloride 112 or 104 mmol/l, potassium 2.0 mmol/l, bicarbonate 34 mmol/l, acetate 3 mmol/l, calcium 1.75 mmol/l and magnesium 0.5 mmol/l. The dialysate temperature was 37.5°C. Blood flow was 250 ml/min and dialysate flow was 500 ml/min.

Methods

Arterial blood pressure (BP) was measured with an automatic blood pressure monitor (Dinamap 1486 SX, Critikon Inc., USA). The mean value of four consecutive measurements was calculated. Heart rate (HR) was obtained from an electrocardiogram. The mean value of two minutes was calculated. Forearm vascular resistance (FVR) and venous tone (VT) were measured as described in chapter 2. The hand circulation was occluded from one minute prior to as well as during flow measurements. All measurements were performed before and every 20 minutes during dialysis.

Changes in relative blood volume (BV) were measured continuously and non-invasively by means of an optical reflection method which operates at a wavelength of 950 nm. The optical sensor was clipped on the arterial line (Haemoguard 2000, Sanofi Sante). The amount of reflected light has been shown to correlate well with the erythrocyte concentration ($r=0.94$; $p<0.001$)¹⁶. The baseline value was obtained after two minutes of extracorporeal circulation at a blood flow of 250 ml/min without ultrafiltration to exclude the influence of saline (recirculation) present in the extracorporeal circuit at the start of dialysis.

Before as well as after two hours UF+HD, blood was withdrawn for the estimation of: sodium (selective electrode), potassium, ionized calcium, total CO₂, urea, osmolality (Wescor, vapour pressure osmometer) and colloid osmotic pressure. In addition, blood was withdrawn in ice-chilled tubes and directly centrifuged at a temperature of 4°C for the estimation of plasma catecholamines (HPLC, fluorescence detection).

Prostaglandin E₂

For the estimation of plasma Prostaglandin E₂ (PGE₂) blood was withdrawn in ice-chilled tubes (EDTA) and directly centrifuged (15 min at 2000 g) at a temperature of 4°C for the estimation of PGE₂. All samples were stored at -70°C until analysis. PGE₂ was determined by RIA using a commercially available PGE₂ assay system (Amersham International PLC, Amersham, UK). In brief, the assay is based upon competition between the methyl oximate derivative of unlabelled PGE₂ and ¹²⁵I labelled methyl oximate derivative of PGE₂ for a limited quantity of binding sites on a rabbit antibody specific for the methyl oximate derivative of PGE₂. Separation of the bound and free derivative of PGE₂ was achieved by reaction of the bound fraction with a donkey antirabbit antibody bound to magnetizable polymer particles. After application of a strong magnetic field the separation was effected by decantation of the supernatant after which the remaining radioactivity was counted. The activity of the standards provided in the kit was plotted against their assigned values. The concentrations of the unknown samples were determined by interpolation from the standard curve.

Crossreactivity (%) at 50% displacement of the tracer was 100% for PGE₂ and about 60% for 8-iso-PGE₂. PGE₂ and 20 OH PGE₂ cross react for about 5%. All other substances known to interfere cross reacted to a percentage below 0.01%. In our laboratory, the sensitivity of the assay was 2.7 pg/ml. The precision of the assay was better than 7% at the level of 25 pg/ml, better than 4% at the level of 100 pg/ml and better than 3% at the level of 400 pg/ml.

Statistical analysis

Changes in hemodynamic parameters within each treatment as well as differences between treatments were analyzed by repeated measurements MANOVA (SPSS-PC version). If the sphericity of the variance-covariance matrix appeared to be violated, degrees of freedom in the univariate MANOVA tests were corrected by the Greenhouse-Geisser epsilon to avoid Type 1 error in testing the F-ratio. All laboratory parameters were analyzed by Friedman's ANOVA and if appropriate by Wilcoxon's signed-rank test. All values are expressed as mean ± SD.

Results

The mean pre-dialysis weights at the start of high- and low-sodium UF+HD were respectively 76.7 ± 11.1 and 76.6 ± 11.3 kg (ns). The laboratory data are shown in table 3.1. In both low- and high-sodium UF+HD, the dialysate sodium concentration was respectively hypo- and hypernatric in comparison to the pre-dialysis plasma sodium level in eight patients while the concentration was isotonic in one patient in both treatment modalities. During high-sodium UF+HD, plasma sodium tended to increase ($p=0.076$) while during low-sodium UF+HD plasma sodium decreased significantly. As a result, plasma sodium was significantly lower after low-sodium UF+HD as compared to high-sodium UF+HD ($p<0.01$). With comparable changes in urea, there was no significant difference in plasma osmolality between high- and low-sodium UF+HD. The increase in colloid osmotic pressure was significantly larger during low-sodium dialysis. Changes in both plasma noradrenaline and adrenaline did not reach statistical significance during both treatment modalities.

Table 3.1 Laboratory data

	Sodium 134 mmol/l		Sodium 144 mmol/l		changes
	pre	post	pre	post	134 vs 144
Sodium	137.1 (2.0)	135.0 (2.7) ^a	138.4 (2.7)	139.9 (2.0)	0.017
Potassium	5.3 (0.8)	3.9 (0.5) ^b	5.2 (0.9)	4.0 (0.8) ^b	ns
ionized Calcium*	1.29 (0.07)	1.39 (0.06) ^a	1.30 (0.07)	1.46 (0.04) ^a	ns
total CO ₂	25.1 (2.8)	29.8 (1.6) ^b	26.0 (2.5)	29.4 (1.7) ^b	ns
Urea	25.5 (6.4)	14.4 (4.9) ^b	26.0 (6.7)	14.5 (4.3) ^b	ns
Osmolality	298.9 (14.0)	285.0 (9.1) ^a	302.9 (12.4)	293.7 (10.9) ^a	ns
COP	3.1 (0.3)	3.8 (0.6) ^b	3.0 (0.3)	3.4 (0.4) ^b	0.018
Noradrenaline	1.36 (0.56)	1.47 (0.59)	1.53 (0.66)	1.49 (0.64)	ns
Adrenaline	0.09 (0.06)	0.10 (0.03)	0.11 (0.06)	0.10 (0.03)	ns
PGE ₂	32.2 (25.0)	76.9 (91.1) ^a	20.6 (8.6)	31.6 (29.7)	ns

Sodium, Potassium, ionized Calcium, total CO₂, Urea (mmol/l); Osmolality (mOsm/kg); COP (kPa); Catecholamines (nmol/l); PGE₂ (pg/ml); post vs pre, ^a= $p<0.05$ and ^b= $p<0.01$; * $n=5$.

The individual changes in plasma PGE₂ are shown in figure 3.1. One blood sample of a patient obtained after low-sodium was lost. Therefore, data of eight patients were analyzed. Plasma PGE₂ increased significantly only during low-sodium UF+HD while during high-sodium UF+HD the increase did not reach statistical significance ($p=0.070$).

Despite a substantially larger mean increase in PGE₂ during low-sodium UF+HD, the difference between the two treatment modalities was not statistically significant.

Figure 3.1 Individual changes in plasma PGE₂ during low (left) and high (right) sodium UF+HD

Changes in relative blood volume are shown in figure 3.2. Blood volume decreased to a mean value of $92 \pm 4\%$ and $82 \pm 5\%$ in respectively high- and low-sodium UF+HD. The difference was statistically significant ($p < 0.01$).

Figure 3.2 Changes in relative blood volume during low (—) and high (---) sodium UF+HD

The hemodynamic data are shown in table 3.2 and figure 3.3a+b. During both treatments, FVR increased significantly while VT remained unchanged. There were no significant differences in FVR and VT between high- and low-sodium UF+HD.

For the group as a whole, there were no significant changes in both arterial blood pressure and heart rate during the two treatment modalities. In two patients, systolic blood pressure decreased by about 20 mmHg only during low-sodium UF+HD and remained stable during high-sodium UF+HD. One of these two patients became symptomatically hypotensive during the third dialysis hour after attaining the upright posture while seated.

There was no significant correlation between changes in plasma PGE₂ and changes in plasma sodium, FVR, VT or arterial blood pressure.

Table 3.2 Hemodynamic data

	Sodium 134 mmol/l		Sodium 144 mmol/l		changes
	pre	post	pre	post	134 vs 144
Systolic BP	148 (30)	145 (27)	146 (26)	150 (31)	ns
Diastolic BP	85 (13)	88 (13)	85 (11)	88 (17)	ns
Mean Arterial BP	107 (19)	109 (18)	107 (16)	111 (23)	ns
Heart Rate	69 (15)	69 (12)	71 (14)	69 (13)	ns
FVR	2113 (741)	3380 (874) ^a	2057 (677)	3363 (1033) ^a	ns
VT	11.9 (2.9)	12.9 (3.0)	12.4 (3.2)	12.5 (2.8)	ns

BP (mmHg); HR (beats/min); FVR (mmHg/ml/100ml/sec); VT (mmHg/ml/100ml); post vs pre, ^a=p<0.01

Figure 3.3a FVR during low (—) and high (---) sodium UF+HD, * = p<0.01 as compared to pre-treatment value

Figure 3.3b VT during low (—) and high (- - -) sodium UF+HD

Discussion

Peripheral vascular reactivity was studied during high- and low-sodium dialysis to investigate whether vascular reactivity during UF+HD is related to changes in plasma osmolality. Our results show that both arteriolar and venous reactivity are not influenced by sodium related changes in plasma osmolality, or better plasma tonicity. In addition, they show that improved hemodynamic stability in high-sodium dialysis as reported in previous studies is probably only mediated by an improvement in plasma volume preservation.

A model of two hours was used to study the vascular responses as Koومان et al. from our group found clear differences in both FVR and VT between isolated ultrafiltration and UF+HD within one hour⁵. In addition, the model offers the possibility to study patients at a high UF-rate of 1 l/hr under completely standardised conditions. Finally, changes in both plasma sodium and osmolality are most pronounced during the first two treatment hours.

Changes in plasma osmolality have been suggested to play a role in the impaired hemodynamic response during UF+HD. In this respect, only loss of tonically active substances seems to be important as adding of urea to dialysate did not improve the hemodynamic response in both humans and dogs^{17,18}. Therefore, vascular reactivity was studied in relation to an active osmol, i.e. sodium. Two hours UF+HD at a dialysate sodium concentration of respectively 134 and 144 mmol/l resulted in a significantly different plasma sodium concentration. As changes in plasma osmolality are mainly urea related, there was no significant difference in plasma osmolality after high- and low-sodium UF+HD. However, with comparable pre-dialysis weights and the same ultrafiltration rate in both

treatment modalities the significantly larger decrease in relative blood volume during low-sodium UF+HD points to an important difference in plasma tonicity. In previous studies, it has been shown that with higher dialysate sodium concentrations refill of the intravascular compartment is improved by the induction of an osmotic gradient both over the cellular as well as over the capillary wall^{19,20}.

High- and low-sodium UF+HD were associated with a significant increase in FVR while VT remained unchanged. Moreover, there were no significant differences in these parameters between the two treatments which indicates that vascular reactivity is not directly influenced by sodium related changes in plasma tonicity. Although not significantly different, the data presented in figure 3.3a even suggest that during low-sodium UF+HD, the larger decline in blood volume was counterbalanced by a somewhat higher level of peripheral resistance.

Arterial blood pressure and heart rate were comparable between high- and low-sodium UF+HD. For the group as a whole one might expect a larger difference in blood pressure between the two treatments as the decrease in blood volume was substantially larger during low-sodium UF+HD while the increase in vascular tone was comparable to high-sodium UF+HD. These observations suggest that in these patients, selected for their relatively good cardiac function, blood pressure was adequately maintained on the basis of an increase in peripheral resistance and probably also in myocardial contractility as a result of an increase in plasma ionized calcium²¹, despite a lower level of venous return. In only two patients, systolic blood pressure decreased substantially during low-sodium UF+HD despite a relatively small decrease in blood volume (-13% and -14%). This could be related to differences in the individual cardiac status as these two patients had the lowest ejection fraction of the studied patients.

Our results are in contrast with those obtained by Henrich et al. who studied the blood pressure response during both regular dialysis and isolated ultrafiltration as well as during dialysis with either isotonic (5%) or hypertonic (25%) mannitol infusion. Plasma osmolality remained stable only during isolated ultrafiltration and dialysis with hypertonic mannitol infusion while blood pressure was best preserved during these treatments as well as during an orthostatic manoeuvre after these treatments. As the increase in both hematocrit and plasma proteins was comparable between dialysis with either isotonic or hypertonic mannitol infusion, compatible with comparable changes in plasma volume during these two treatments, the authors suggested a direct relation between changes in plasma osmolality and vascular reactivity¹³. However, vascular reactivity was not directly monitored in their study while it is our impression that differences in extracellular volume distribution are not conclusively excluded since data on changes in plasma volume are not presented for all treatment modalities. Moreover, at variance with our study plasma osmolality was completely counterbalanced with hypertonic mannitol

infusion.

Release of eicosanoids during dialysis has been related to the membrane type^{22,23} as well as to the use of low-sodium dialysate¹⁴. Schultze et al. found a larger decrease in mean arterial blood pressure in association with higher levels of plasma PGE₂ during low-sodium dialysis (126 mmol/l) as compared to high-sodium dialysis (140 mmol/l). They hypothesized that hemodynamic instability during low-sodium dialysis could be related to the release of vasodilating prostaglandins¹⁴. In the present study, plasma PGE₂ increased significantly also only during low-sodium UF+HD while in agreement with Schultze et al. the data shown in figure 3.1 suggest a responder non-responder phenomenon. However, there were no significant differences in plasma PGE₂ between high- and low-sodium UF+HD in the present study while in addition there was no relation between either the vascular or blood pressure response and changes in plasma PGE₂. It is our opinion therefore, that with respect to hemodynamics the increase in plasma PGE₂ during low-sodium dialysis seems to be an epi-phenomenon.

In conclusion, vascular reactivity during combined ultrafiltration-hemodialysis is not related to sodium related changes in plasma tonicity while improved hemodynamic stability with high-sodium dialysate is primarily related to a better preservation of plasma volume. This indicates that in contrast to differences in the extracorporeal blood temperature, changes in plasma osmolality probably do not play an important role in the impaired vascular response during combined ultrafiltration-hemodialysis as compared to isolated ultrafiltration and hemofiltration. Possible factors that might be important in this respect are differences in ionized calcium or biocompatibility^{24,25}.

References

- 1 Daugirdas JT: Dialysis hypotension: A hemodynamic analysis. *Kidney Int* 1991; 39: 233-246.
- 2 Rothe CF: Venous system: Physiology of capacitance vessels; in Shepherd JT, Abboud FM (ed.): *Handbook of Physiology*. Bethesda, American Physiological Society, 1983, 3, Peripheral circulation and organ blood flow, pp. 397-452
- 3 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
- 4 Bradley JR, Evans DB, Cowley AJ: Comparison of vascular tone during combined haemodialysis with ultrafiltration and during ultrafiltration followed by haemodialysis: a possible mechanism for dialysis hypotension. *BMJ* 1990; 300: 1312.
- 5 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
- 6 Maggiore Q, Pizzarelli F, Zoccali C, Sica S, Nicolò F: Influence of blood temperature on vascular stability during hemodialysis and isolated ultrafiltration. *Int J Artif Organs* 1985; 8: 175-178.
- 7 van Kuijk WHM, Luik AJ, de Leeuw PW, van Hooff JP, Nieman FHM, Habets HML, Leunissen KML: Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences. *Nephrol Dial Transplant* 1995; (in press).
- 8 Bergström J: Ultrafiltration without dialysis for removal of fluid and solutes in uremia. *Clin Nephrol* 1978; 9: 156-164.
- 9 Palmer BF: The effect of dialysate composition on systemic hemodynamics. *Sem Dial* 1992; 5: 54-60.
- 10 Henrich WL, Woodard TD, McPhaul JJ: The chronic efficacy and safety of high sodium dialysate: double-blind, crossover study. *Am J Kidney Dis* 1982; 2: 349-353.
- 11 Cybulsky AVE, Matni A, Hollomby DJ: Effects of high sodium dialysate during maintenance hemodialysis. *Nephron* 1985; 41: 57-61.
- 12 Aguilera D, Diab N, Faivre J-M: Influence of sodium dialysate variation on hemodynamic stability. *Kidney Int* 1988; 34: S187-S189.
- 13 Henrich WL, Woodard TD, Blachley JD, Gomez-Sanchez C, Pettinger W, Cronin RE: Role of osmolality in blood pressure stability after dialysis and ultrafiltration. *Kidney Int* 1980; 18: 480-488.
- 14 Schultze G, Maiga M, Neumayer H-H, Wagner K, Keller F, Molzahn M, Nigam S: Prostaglandin E₂ promotes hypotension on low-sodium hemodialysis. *Nephron* 1984; 37: 250-256.

-
- 15 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
 - 16 de Vries JPPM, Olthof CG, Viosser V, Kouw PM, van Es A, Donker AJM, de Vries PMJM: Continuous measurement of blood volume during hemodialysis by an optical method. *ASAIO Journal* 1992; 38: M181-M185.
 - 17 Wehle B, Asaba H, Castenfors J, Gunnarsson B, Bergström J: Influence of dialysate composition on cardiovascular function in isovolemic haemodialysis. *Proc EDTA* 1981; 18: 153-159.
 - 18 Daugirdas JT, Ing TS, Chen W-T, Vestal RE, Hano JE, Al-Kudsi RR: Ultrafiltration hemodynamics in conscious, uremic dogs: Effect of a decreasing plasma urea level. *Mineral Electrolyte Metab* 1985; 11: 79-84.
 - 19 de Vries PMJM, Olthof CG, Solf A, Schuenemann B, Oe PL, Quellhorst E, Schneider H, Donker AJM: Fluid balance during haemodialysis and haemofiltration: The effect of dialysate sodium and a variable ultrafiltration rate. *Nephrol Dial Transplant* 1991; 6: 257-263.
 - 20 Kouw PM, Olthof CG, Gruteke P: Influence of high and low sodium dialysis on blood volume preservation. *Nephrol Dial Transplant* 1991; 6: 876-880.
 - 21 Henrich WL, Hunt JM, Nixon JV: Increased ionized calcium and left ventricular contractility during hemodialysis. *N Engl J Med* 1984; 310: 19-23.
 - 22 Schultze G, Wagner K, Neumayer HH, Fitzner R, Molzahn M: Effect of dialyzer membranes on in vitro generation of eicosanoids. *Int J Artif Organs* 1987; 10: 275-278.
 - 23 Mahiout A, Jörres A, Hiss R, Meinhold H, Kessel M: Effects of blood-dialyzer interaction on prostaglandins in uremic patients and in healthy man. *Nephrol Dial Transplant* 1987; 2: 546-550.
 - 24 Leunissen KML, van den Berg BW, van Hooff JP: Ionized calcium plays a pivotal role in controlling blood pressure during haemodialysis. *Blood Purif* 1989; 7: 233-239.
 - 25 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
-

Chapter 4

Vascular reactivity during combined ultrafiltration- hemodialysis: influence of dialysate derived contaminants

WHM van Kuijk, WA Buurman, PGG Gerlag,
KML Leunissen

(submitted for publication)

Abstract

Background

Hemodynamic instability and impaired vascular reactivity during combined ultrafiltration-hemodialysis have been suggested to be related to bioincompatibility factors such as dialysate derived contaminants or the dialyzer. In the present study it was investigated whether vascular reactivity could be improved by the use of sterile dialysate.

Methods

Forearm vascular resistance and venous tone (strain-gauge plethysmography) as well as arterial blood pressure (Dinamap) and heart rate (EKG) were measured in ten stable dialysis patients (28-71 years) during two hours combined ultrafiltration-hemodialysis (bicarbonate, UF-rate 1.0 l/hr). In addition, a dialysate sample was obtained for culture and LAL testing while blood was withdrawn for the estimation of plasma Bactericidal/Permeability Increasing Factor (BPI, ELISA) and the soluble TNF receptor p75 (sTNF-Rp75, ELIBA). Patients served as their own control, comparing dialysis with non-sterile and sterile dialysate.

Results

We observed no bacterial growth in sterile dialysate while all samples were positive in culture in non-sterile dialysis (*Pseudomonas*). All LAL tests were negative. BPI tended to increase during non-sterile dialysis ($p=0.063$) and remained unchanged during sterile dialysis. In both treatments TNF-Rp75 increased significantly ($p<0.01$). There were no significant differences in hemodynamic parameters between the treatment modalities. Despite use of sterile dialysate, forearm vascular resistance remained unchanged while venous tone decreased significantly.

Conclusion

We conclude that vascular reactivity during combined ultrafiltration-hemodialysis is not improved by the use of sterile dialysate.

Introduction

One of the main goals of dialysis therapy is withdrawal of excess fluid by ultrafiltration which might result in hemodynamic instability with symptomatic hypotension under certain conditions. When blood volume decreases, blood pressure stability is especially dependent on an adequate increase in peripheral vascular tone. In this respect, fluid withdrawal during isolated ultrafiltration and hemofiltration has been shown to result in an adequate increase in peripheral vascular tone while in contrast, peripheral vascular reactivity is impaired during combined ultrafiltration-hemodialysis^{1,2}. A number of factors such as differences in plasma osmolality, acid-base status or the extracorporeal blood temperature have been suggested to be responsible for these differences in vascular reactivity of which the latter seems to be most important³⁻⁶.

Henderson and colleagues hypothesized in their 'Interleukin Hypothesis' that hemodynamic instability during combined ultrafiltration-hemodialysis could be related to the release of cytokines by blood mononuclear cells after stimulation by pyrogenic materials derived from dialysate or as a result of complement activation due to bioincompatibility of the dialyzer. Better hemodynamic stability as observed during isolated ultrafiltration and hemofiltration would then reflect the absence of dialysate (and contaminants) in both treatment modalities and the use of more biocompatible membranes in HF⁷. In this respect, dialysate contaminants are now considered to be the most important bioincompatibility factor^{8,9} while most dialysis membranes have been shown to be permeable for pyrogens derived from dialysate^{10,11}. Conflicting data exist whether plasma cytokine levels increase during dialysis¹²⁻¹⁴. However, monocytes have been shown to be stimulated by a single dialysis treatment in which dialysate derived pyrogens play an essential role^{9,15}. Concerning hemodynamic instability, Beasley et al. have proposed a possible role for cytokine-induced activation of nitric oxide synthesis within vascular smooth muscle cells¹⁶.

However, the clinical importance of dialysate derived contaminants with respect to hemodynamics remains to be established. In the present study therefore, we investigated whether impaired vascular reactivity as observed during combined ultrafiltration-hemodialysis could be improved by the use of pyrogen free dialysate.

Subjects and methods

Subjects and dialysis

All patients signed an informed consent for participation in the study that had been approved by the Ethics Committee of the St. Joseph Hospital Veldhoven. Two women and eight men on intermittent hemodialysis were included with a mean age of 51 years (range 28-71) and a mean time on hemodialysis of 99 months (range 15-260). All patients were hemodynamically stable patients who rarely suffered from intradialytic hypotension. Exclusion criteria were severe coronary (NYHA II or more) or valvular heart disease, compromised left ventricular function (ejection fraction 30% or less), diabetes mellitus and a donor kidney in situ. Patients were studied on the regular day of his or her dialysis schedule. By adjusting fluid intake, patients were instructed to achieve a pre-dialysis weight at about 2.0 kilograms above their dry-weight that was estimated by echography of the inferior caval vein¹⁷. All vasoactive medication was stopped 48 hours before the study and the patients were not allowed to smoke or drink caffeine containing beverages during 12 hours preceding the study.

Each patient served as his or her own control and was studied during two standardised dialysis treatments of two hours combined ultrafiltration-hemodialysis (UF+HD) at an ultrafiltration rate of 1 l/hr using either non-sterile (NS UF+HD) or sterile dialysate (S UF+HD). The order of treatments was randomised. Non-sterile dialysate was routinely prepared from a liquid bicarbonate component while overnight the dialysis monitor (AK-100, Gambro, Sweden) was sterilised by heat disinfection. Sterile dialysate was prepared from a Bicart module while two 1.4 m² Polyamide (U-7000) membranes were used for filtration of dialysate (AK-100 Ultra, Gambro, Sweden). Overnight, the monitor was sterilised with peracetic acid (Dialox, Seppic, France). In all treatments, a hemophane dialyzer (GFS-16, Gambro, Sweden) was used. The composition of dialysate was: sodium 141 mmol/l, potassium 2.0 mmol/l, calcium 1.75 mmol/l, bicarbonate 34 mmol/l, acetate 3 mmol/l, magnesium 0.5 mmol/l and chloride 108 mmol/l. The dialysate temperature was 37.5°C. Blood flow was 250 ml/min and dialysate flow was 500 ml/min.

Methods

Before each dialysis treatment a sample of 250 ml dialysate was obtained under aseptical conditions and directly transported to the laboratory for immediate microbiological analysis. Each sample was filtrated through 0.45 m μ microfilters. The microfilters were incubated for 3 days at 30°C on PCA-agar plates after which the number of colony forming units (CFU) per filter was estimated. In addition,

samples of 5 ml dialysate were obtained in pyrogen free tubes and directly frozen at -20°C for LAL testing. LPS was measured with a chromogenic Limulus assay (Coatest endotoxin, Kabivitrum, Stockholm, Sweden) according to the manufacturer's instructions. The detection limit was 10 pg/ml. In addition, data of routine cultures were used to estimate the bacterial content of RO-water.

Before dialysis as well as at $t=30, 60$ and 120 minutes, arterial afferent blood was obtained for the estimation of the plasma levels of Bactericidal/Permeability Increasing Factor (BPI) and the soluble TNF receptor p75 (sTNF-Rp75). Blood was withdrawn in ice-chilled tubes (EDTA) and immediately centrifuged. Plasma was obtained without disturbing the buffy-coat and stored at -20°C .

BPI was determined by sandwich ELISA using a human rBPI specific Mab 4E3, as described previously¹⁸. The sTNF-Rp75 receptor was determined by an enzyme-linked immunological biological assay (ELIBA) using a mAb utr4¹⁹. The detection limits for BPI and sTNF-Rp75 were respectively 100 pg/ml and 400 pg/ml.

Arterial blood pressure (BP) and heart rate (HR) were measured as described in chapter 3 while forearm vascular resistance (FVR) and venous tone (VT) were measured as described in chapter 2. The hand circulation was occluded from one minute before as well as during flow measurements. All hemodynamic measurements were performed before (pre-treatment) and every 20 minutes during dialysis.

Luminal blood temperatures were measured in the arterial and venous lines with needle thermometers (Hewlett-Packard 78214C temperature monitor). Connectors with a built-in adaptor were placed in the arterial and venous lines under sterile conditions at a distance of about 15 centimetres from the patient. There was no direct contact between blood and thermometer. Changes in relative blood volume (BV) were measured continuously and non-invasively by means of an optical reflection method which operates at a wavelength of 950 nm. The optical sensor was clipped on the arterial line (Haemoguard 2000, Sanofi Sante)²⁰. The baseline value was obtained after two minutes of extracorporeal circulation at a blood flow of 250 ml/min without ultrafiltration to exclude the influence of saline (recirculation) present in the extracorporeal circuit at the start of dialysis.

Before as well as after two hours UF+HD, blood was withdrawn for the estimation of: sodium, potassium, total CO_2 , urea, osmolality (Wescor, vapour pressure osmometer) and colloid osmotic pressure (COP). In addition, blood was withdrawn in ice-chilled tubes and directly centrifuged at a temperature of 4°C for the estimation of plasma catecholamines (HPLC, fluorescence detection).

Statistical analysis

Changes in hemodynamic parameters within each treatment as well as differences between treatments were analyzed by repeated measurements MANOVA (SPSS-PC version). If the sphericity of the variance-covariance matrix appeared to be violated, degrees of freedom in the univariate MANOVA tests were corrected by the Greenhouse-Geisser epsilon to avoid Type 1 error in testing the F-ratio. All laboratory parameters were analyzed by Friedmans ANOVA and if appropriate, by the Wilcoxon's signed-rank test.

Results

Samples obtained prior to S UF+HD showed no bacterial growth in culture with the exception of two samples. 4 colony forming units (CFU) were counted in both samples, identified as respectively *Acinetobacter* and *Bacillus* species. All dialysate samples were positive in culture in NS UF+HD, identified as mainly *Pseudomonas* species. Due to confluence of colonies it was not possible to count the number of CFU in NS UF+HD reliably. All LAL tests were negative in both treatment modalities. RO-water contained 201 ± 209 CFU per ml ($n=30$).

The immunological data are presented in figure 4.1a+b. Although not statistically significant, BPI tended to increase during NS UF+HD ($p=0.063$ at $t=60$ min). In S UF+HD, BPI remained unchanged. The sTNF-Rp75 increased significantly in both treatments ($p<0.01$). There were no significant differences in BPI or sTNF-Rp75 between the two treatment modalities.

Figure 4.1a Plasma BPI during NS UF+HD (—) and S UF+HD (---)

Figure 4.1b Plasma sTNF-Rp75 during NS UF+HD (—) and S UF+HD (- - -);
 *= $p < 0.01$ as compared to pre-treatment level

The predialysis weights in NS UF+HD and S UF+HD were respectively 69.1 ± 6.6 kg and 69.1 ± 6.5 kg (ns). The hemodynamic data are presented in table 4.1 and figure 4.2a+b. Systolic BP decreased significantly in NS UF+HD (-11 ± 9 mmHg, $p < 0.05$) while diastolic and mean arterial BP remained unchanged. There were no significant changes in arterial BP during S UF+HD. Heart rate increased significantly during both NS UF+HD and S UF+HD with respectively 6 ± 8 ($p < 0.05$) and 6 ± 7 ($p < 0.01$) beats/min. Concerning vascular reactivity, FVR did not change significantly during both treatments. VT decreased significantly during NS UF+HD and S UF+HD with respectively -2.4 ± 3.6 ($p < 0.05$) and -3.1 ± 3.5 ($p < 0.01$) mmHg/ml/100ml. There were no significant differences in hemodynamic parameters comparing NS-UF+HD with S UF+HD.

Table 4.1 Hemodynamic data

	NS UF+HD		S UF+HD		changes
	pre	post	pre	post	NS vs S
Systolic BP	151 (16)	141 (11) ^a	149 (18)	144 (14)	ns
Diastolic BP	84 (8)	85 (9)	82 (10)	84 (11)	ns
Mean Arterial BP	111 (13)	108 (9)	110 (16)	105 (12)	ns
Heart Rate	72 (11)	78 (12) ^b	74 (13)	80 (14) ^a	ns
FVR	2769 (1300)	3440 (979)	2819 (1503)	2993 (700)	ns
VT	17.5 (5.5)	15.2 (3.9) ^b	17.5 (3.9)	14.4 (2.9) ^a	ns

BP (mmHg); HR (beats/min); FVR (mmHg/ml/100ml/sec); VT (mmHg/ml/100ml); post vs pre,
^a= $p < 0.01$, ^b= $p < 0.05$.

Figure 4.2a FVR during NS UF+HD (—) and S UF+HD (- - -)

Figure 4.2b VT during NS UF+HD (—) and S UF+HD (- - -); *= $p < 0.05$
#= $p < 0.01$ as compared to pre-treatment level

The laboratory data are presented in table 4.2. Changes in the different laboratory parameters were comparable between NS UF+HD and S-UF+HD. Plasma adrenaline remained unchanged during both treatment modalities. During NS-UF+HD, plasma noradrenaline increased significantly ($p=0.050$) while during S-UF+HD the increase did not reach statistical significance ($p=0.059$). Blood volume decreased respectively to 84 ± 6 and $86 \pm 7\%$ (ns).

Table 4.2 Laboratory data

	NS UF+HD		S UF+HD		changes NS vs S
	pre	post	pre	post	
Sodium	137.7 (2.5)	140.2 (1.6) ^a	138.1 (2.1)	139.4 (1.8) ^b	ns
Potassium	5.7 (0.6)	4.2 (0.4) ^a	5.6 (0.5)	4.1 (0.2) ^a	ns
total CO ₂	21.7 (2.4)	24.3 (2.1) ^a	21.7 (3.0)	24.3 (1.8) ^a	ns
Urea	22.7 (4.9)	11.3 (3.4) ^a	23.4 (2.8)	11.6 (2.2) ^a	ns
Osmolality	300.7 (10.1)	289.6 (8.4) ^a	298.4 (7.1)	287.6 (7.8) ^a	ns
COP	3.4 (0.3)	4.2 (0.5) ^a	3.4 (0.3)	4.1 (0.6) ^a	ns
Noradrenaline	1.84 (0.62)	2.15 (0.58) ^b	1.74 (0.76)	1.99 (0.57)	ns
Adrenaline	0.18 (0.08)	0.18 (0.12)	0.19 (0.08)	0.17 (0.09)	ns

Sodium, Potassium, total CO₂ and Urea (mmol/l); Osmolality (mOsm/kg); COP (kPa); Catecholamines (nmol/l); post vs pre, ^a=p < 0.01 and ^b=p < 0.05

The blood temperatures are presented in figure 4.3a+b. The extracorporeal blood temperatures were comparable with no significant changes during either treatment.

Figure 4.3a Arterial (—) and venous (- - -) blood temperature during NS UF+HD

Figure 4.3b Arterial (—) and venous (---) blood temperature during S UF+HD

Discussion

To study the role of bacteriological contamination of dialysate in the hemodynamic response during dialysis it was investigated whether vascular reactivity during UF+HD could be improved by the use of sterile dialysate. The main conclusion of our results is that both arteriolar and venous reactivity during UF+HD are not improved when sterile dialysate is used.

Ultrapure dialysate was prepared by ultrafiltration of dialysate through two polyamide membranes. The physicochemical characteristics of this hydrophilic membrane have been shown to completely remove pyrogenic, cytokine-inducing substances from dialysate^{21,22}. In agreement, only two out of ten cultures were positive in S UF+HD with a very low count of CFU while all LAL tests were negative. Identification as *Acinetobacter* and *Bacillus* species of these two positive samples is highly suggestive for secondary contamination. Whereas it was not possible to count the number of CFU reliably, samples obtained in NS-UF+HD were all positive in culture with *Pseudomonas* species being the most dominant genus, present in all samples. Based on the number of CFU in RO-water, at least 200 CFU per ml can be expected in dialysate. Although all LAL tests were negative in NS UF+HD, presence of pyrogens other than LPS or lipid A that might pass the dialysis membrane such as LPS fragments, peptidoglycans and muramylpeptides is not excluded¹⁰. A possibility that is excluded in S UF+HD since monocytes in culture are not stimulated by contaminated dialysate after

filtration of dialysate through polyamide membranes²². Moreover, exposition to pyrogens in NS UF+HD is further suggested by a trend towards an increase in BPI which remained unchanged during S UF+HD. In agreement, Schindler et al. have found an increase in BPI during dialysis with either a cuprophane or polysulfone dialyzer. The microbiological contamination of dialysate in their study varied between 20-150 CFU per ml²³.

Since the sTNF-Rp75 increased significantly during both treatment modalities stimuli other than pyrogens such as the dialysis membrane could play a role in the release of the sTNF-Rp75 from mononuclear and endothelial cells during dialysis.

Hemodynamics were measured under strictly standardised conditions with comparable pre-dialysis weights and changes in relative blood volume. In addition, changes in laboratory parameters as well as in extracorporeal blood temperature were also comparable. Only in NS UF+HD, the decrease in systolic blood pressure reached statistical significance. However, there were no significant differences in either arterial blood pressure, heart rate or vascular reactivity between NS UF+HD and S UF+HD. Despite the use of sterile dialysate UF+HD was still associated with a decrease in VT while FVR remained unchanged. In addition, HR increased. In previous studies comparable hemodynamic responses have been observed in UF+HD using non-sterile dialysate which contrasts with a clear increase in peripheral vascular tone as observed during hemofiltration and isolated ultrafiltration^{1,2,6}. Consequently, our results show that moderate dialysate contamination is not responsible for the impaired vascular response during UF+HD as compared to hemofiltration and isolated ultrafiltration. We can however not completely exclude the possibility that high levels of dialysate contamination might lead to a further impairment of the vascular response in UF+HD.

References

- 1 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
- 2 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
- 3 Daugirdas JT: Dialysis hypotension: A hemodynamic analysis. *Kidney Int* 1991; 39: 233-246.
- 4 Henrich WL, Woodard TD, Blachley JD, Gomez-Sanchez C, Pettinger W, Cronin RE: Role of osmolality in blood pressure stability after dialysis and ultrafiltration. *Kidney Int* 1980; 18: 480-488.
- 5 Maggiore Q, Pizzarelli F, Sisca S, Zoccali C, Parlongo S, Nicolò F, Creazzo G: Blood temperature and vascular stability during hemodialysis and hemofiltration. *Trans Am Soc Artif Intern Organs* 1982; 28: 523-527.
- 6 van Kuijk WHM, Luik AJ, de Leeuw PW, van Hooff JP, Nieman FHM, Habets HML, Leunissen KML: Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences. *Nephrol Dial Transplant* 1995; (in press).
- 7 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
- 8 Schindler S, Lonneman G, Shaldon S, Koch KM, Dinarello CA: Transcription, not synthesis, of interleukin-1 and tumor necrosis factor by complement. *Kidney Int* 1990; 37: 85-93.
- 9 Colton CK, Ward RA, Shaldon S: Scientific basis for assessment of biocompatibility in extracorporeal blood treatment. *Nephrol Dial Transplant* 1994; 9: 11-17.
- 10 Lonnemann G: Dialysate bacteriological quality and the permeability of dialyzer membranes to pyrogens. *Kidney Int* 1993; 43: S195-S200.
- 11 Laude-Sharp M, Caroff M, Simard L, Pusineri C, Kazatchkine MD, Haeffner-Cavaillon N: Induction of IL-1 during hemodialysis: transmembrane passage of intact endotoxins. *Kidney Int* 1990; 38: 1089-1094.
- 12 Herbelin A, Nguyen AT, Zingraff J, Ureña P, Descamps-Latscha B: Influence of uremia and hemodialysis on circulating interleukin-1 and tumor necrosis factor α . *Kidney Int* 1990; 37: 116-125.
- 13 Tetta C, Camussi G, Turello E, Salomone M, Aimo G, Priolo G, Segoloni G, Vercellone A: Production of cytokines in hemodialysis. *Blood Purif* 1990; 8: 337-346.
- 14 Bingel M, Lonnemann G, Koch KM, Dinarello CA, Shaldon S: Plasma interleukin-1 during hemodialysis: the influence of dialysis membranes. *Nephron* 1988; 50: 273-276.

-
- 15 Haeffner-Cavaillon N, Cavaillon J-M, Ciancioni C, Bacle F, Delons S, Kazatchkine MD: In vivo induction of interleukin-1 during hemodialysis. *Kidney Int* 1989; 35: 1212-1218.
 - 16 Beasley D, Brenner BM: Role of nitric oxide in hemodialysis hypotension. *Kidney Int* 1992; 42: S96-S100.
 - 17 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
 - 18 Dentener MA, Francot GJM, Smit FT, Froon AHM, Pennings H-J, Wouters EFM, Buurman WA: Presence of bactericidal/permeability-increasing protein in disease: detection by ELISA. *J Infect Dis* 1995; 171: 739-743.
 - 19 Leeuwenberg JFM, Mat O, Abramowicz D, Gastaldello R, Tielemans C, Buurman WA: Increased plasma levels of soluble tumor necrosis factor-receptors in uraemic patients: effects of dialysis, type of membrane, and anticoagulation method. *Nephrol Dial Transplant* 1994; 9: 1125-1129.
 - 20 de Vries JPPM, Olthof CG, Viosser V, Kouw PM, van Es A, Donker AJM, de Vries PMJM: Continuous measurement of blood volume during hemodialysis by an optical method. *ASAIO Journal* 1992; 38: M181-M185.
 - 21 Lonnemann G, Mahiout A, Schindler R, Colton CK: Pyrogen retention by the polyamide membranes. *Contrib Nephrol* 1992; 96: 47-63.
 - 22 Di Felice A, Cappelli G, Facchini F, Tetta C, Cornia F, Aimo G, Lusvardi E: Ultrafiltration and endotoxin removal from dialysis fluids. *Kidney Int* 1993; 43: S201-S204.
 - 23 Schindler R, Marra MN, McKelligon BM, Lonnemann G, Schulzeck P, Schulze M, Oppermann M, Shaldon S: Plasma levels of bactericidal/permeability-increasing protein (BPI) and lipopolysaccharide-binding protein (LBP) during hemodialysis. *Clin Nephrol* 1993; 40: 346-351.
-

Chapter 5

Influence of changes in ionized calcium on cardiovascular reactivity during hemodialysis

WHM van Kuijk, AW Mulder, CH Peels, GA Harff,
KML Leunissen

(submitted for publication)

Abstract

Background

In order to prevent hypercalcemia due to the treatment of secondary hyperparathyroidism it is advocated to use low-calcium dialysate. However, as calcium ions play a pivotal role in both myocardial and vascular smooth muscle contraction, lowering the dialysate calcium concentration might result in a further impairment of the cardiovascular response during dialysis.

Methods

Therefore, arterial blood pressure, forearm vascular resistance (FVR) and venous tone (VT) (strain-gauge plethysmography) as well as cardiac dimensions and output (echocardiography) were measured in 10 hemodynamically stable dialysis patients (ejection fraction >30%) during two standardised sessions of three hours combined ultrafiltration-hemodialysis (UF+HD) at two different dialysate calcium concentrations: 1.25 and 1.75 mmol/l.

Results

High-calcium UF+HD resulted in a significant increase in plasma ionized calcium ($+0.19 \pm 0.11$ mmol/l; $p < 0.01$) while ionized calcium remained unchanged during low-calcium UF+HD (-0.02 ± 0.07 mmol/l). As a result, systolic, diastolic and mean arterial blood pressure were respectively 14 ± 10 , 5 ± 7 and 9 ± 9 mmHg higher during high-calcium UF+HD as compared to low-calcium UF+HD ($p < 0.05$). There were no significant differences in both FVR and VT between the two treatments. During both treatments FVR increased while VT decreased. In addition, there were no differences in calculated systemic vascular resistance. However, with comparable end-diastolic dimensions, stroke volume (-18 ± 13 ml) and cardiac output (-1.3 ± 1.5 l/min) decreased significantly ($p < 0.05$) only during low-calcium UF+HD.

Conclusions

We conclude that even in hemodynamically stable patients changes in plasma ionized calcium are an important determinant of the blood pressure response during dialysis therapy. Whereas peripheral vascular reactivity is unaffected by changes in ionized calcium, myocardial contractility is improved with higher dialysate calcium concentrations.

Introduction

Despite a growing understanding of the underlying pathophysiology, hemodynamic instability with symptomatic hypotension is still one of the most important complications occurring during dialysis therapy¹. In addition to the patient's cardiovascular status and changes in intravascular volume, hemodynamic stability is also related to the cardiovascular response during dialysis therapy. The importance of the latter factor is illustrated by the fact that on the basis of a more physiological increase in peripheral vascular tone, hemodynamic stability is better during hemofiltration as compared to combined ultrafiltration-hemodialysis^{2,3}.

With calcium ions playing a pivotal role in the contractile process of both vascular smooth muscle cells and cardiac myocytes as well as in the release of catecholamines from the synaptic cleft, changes in ionized calcium might play an important role in the cardiovascular response during dialysis. Different authors have indeed found a correlation between changes in ionized calcium and blood pressure during dialysis, with differences in myocardial contractility being the most consistent finding⁴⁻⁸. However, it is unknown whether vascular reactivity is also related to changes in ionized calcium. Knowledge about the role of changes in ionized calcium in both blood pressure and cardiovascular responses seems to be especially important as it is advocated to use lower dialysate calcium concentrations in order to prevent or treat hypercalcemia due to the treatment of secondary hyperparathyroidism.

In the present study therefore, blood pressure and both vascular and cardiac reactivity were measured during combined ultrafiltration-hemodialysis at two different dialysate calcium concentrations in the first place to study the influence of changes in plasma ionized calcium on vascular reactivity and secondly to investigate the clinical importance with respect to arterial blood pressure of lowering the dialysate calcium concentration.

Subjects and methods

Subjects and dialysis

After obtaining informed consent, two women and eight men on intermittent hemodialysis were included with a mean age of 47 years (range 25-73) and a mean time on hemodialysis of 64 months (range 6-276). All patients were hemodynamically stable patients who rarely suffered from intradialytic hypotension. Exclusion criteria were severe coronary (NYHA II or more) or valvular heart disease, compromised left ventricular function (ejection fraction 30% or less) and diabetes mellitus. Patients were studied on the regular day of his or her dialysis

schedule. By adjusting fluid intake, patients were instructed to achieve a pre-dialysis weight between 2.0 and 3.0 kilograms above their dry-weight that was estimated by echography of the inferior caval vein⁹. All vasoactive medication was stopped 48 hours before the study while patients were not allowed to smoke or drink caffeine containing beverages during 12 hours preceding to the study.

With each patient serving as his or her own control, measurements were performed during two standardised sessions of three hours combined ultrafiltration-hemodialysis that only differed in the dialysate calcium concentration: 1.25 and 1.75 mmol/l. Dialysis was performed with a Monitral-S dialysis monitor (Hospal, France) using hemophane membranes (GFS-16, Gambro, Sweden) in all treatments. The composition of dialysate was: sodium 141 mmol/l, potassium 2.0 mmol/l, calcium 1.25 or 1.75 mmol/l, bicarbonate 34 mmol/l, acetate 3 mmol/l, magnesium 0.5 mmol/l and chloride 108 mmol/l. The dialysate temperature was 37.5°C. Blood flow was 250 ml/min and dialysate flow was 500 ml/min.

Methods

Forearm vascular resistance (FVR) and venous tone (VT) were measured as described in chapter 2. The hand circulation was occluded from one minute before as well as during flow measurements. Arterial blood pressure (BP) and heart rate (HR) were measured as described in chapter 3. Measurements were performed before dialysis and every 30 minutes during dialysis.

Before as well as immediately after dialysis, an echocardiogram was obtained (Hewlett-Packard, Sonos 1000, Andover, MA, USA) with the patient in the left lateral decubitus position using a 2.5 mHz phased array transducer for both imaging and Doppler studies. Left ventricular end-systolic (LVESD, mm) and end-diastolic (LVEDD, mm) dimensions were measured on M-mode and two-dimensional echo. The diameter of the left ventricular outflow tract (LVOT) was measured at the aortic annulus in the parasternal long axis view. The LVOT area was calculated assuming circular geometry [$LVOT \text{ area} = 3.14 (D/2)^2$]¹⁰. Using pulsed wave Doppler, flow was measured in the LVOT area at the annular level just proximal from the point where aortic leaflet artifacts were observed¹¹. By measuring the area under the curve, the mean systolic velocity integral (SVI) was calculated from 3 consecutive cycles. Left ventricular stroke volume (SV, ml) was calculated from multiplying the LVOT area with SVI. Cardiac output (CO, l/min) was calculated from SV and heart rate. Systemic vascular resistance (SVR, dyne/sec/cm⁻⁵) was calculated from CO and MAP ($SVR = MAP/CO \times 80$). Both temporal and intra-observer variability of flow measurements using the non-invasive Doppler method have been shown to be sufficiently small for the method to be useful in serial hemodynamic measurements^{10,12}.

Changes in relative blood volume (BV) were measured continuously and non-invasively by means of an optical reflection method which operates at a wavelength of 950 nm. The optical sensor was clipped on the arterial line (Haemoguard 2000, Sanofi Sante, The Netherlands)¹³. The baseline value was obtained after two minutes of extracorporeal circulation at a blood flow of 250 ml/min without ultrafiltration to exclude the influence of saline (recirculation) present in the blood lines at the start of dialysis.

Before as well as after three hours UF+HD, blood was withdrawn for the estimation of: sodium, potassium, total CO₂, urea, osmolality (Osmomat 30, Gonotec, Berlin, Germany) and colloid osmotic pressure. In addition, blood was withdrawn in ice-chilled tubes and directly centrifuged at a temperature of 4°C for the estimation of plasma catecholamines (HPLC, fluorescence detection). Plasma ionized calcium was estimated at pre-dialysis as well as every hour during dialysis (Stat profile 6, NOVA biomedical, Waltham, MA, U.S.A.).

Statistical analysis

Changes in hemodynamic parameters within each treatment as well as differences between treatments were analyzed by repeated measurements MANOVA (SPSS-PC version). If the sphericity of the variance-covariance matrix appeared to be violated, degrees of freedom in the univariate MANOVA tests were corrected by the Greenhouse-Geisser epsilon to avoid Type 1 error in testing the F-ratio. All laboratory parameters were analyzed by Friedmans ANOVA and if appropriate, by the Wilcoxon's signed-rank test. All values are expressed as mean $\pm$ SD.

Results

The pre-dialysis weights in low- and high-calcium UF+HD were respectively 70.7 ± 10.7 and 70.8 ± 10.6 kg ($p=0.13$). The mean UF-rate was 0.79 ± 0.10 l/hr in both treatment modalities. The laboratory data are shown in table 5.1 and figure 5.1. Plasma ionized calcium remained unchanged during low-calcium UF+HD (-0.02 ± 0.07 mmol/l) whereas it increased significantly during high-calcium UF+HD ($+0.19 \pm 0.11$ mmol/l). As a result, ionized calcium was significantly higher during the latter treatment as compared to low-calcium UF+HD. All other laboratory parameters as well as relative blood volume changed comparably during both treatments. Plasma adrenaline remained also unchanged during both treatment modalities. However, plasma noradrenaline increased significantly only during high-calcium UF+HD. As compared to low-calcium UF+HD ($+0.37 \pm 0.74$ nmol/l) the increase in plasma noradrenaline was significantly larger during

high-calcium UF+HD ($+0.70 \pm 0.53$ nmol/l). Relative BV decreased to 86 ± 4 and $85 \pm 5\%$ in respectively low- and high-calcium UF+HD (ns).

Table 5.1 Laboratory data

	Calcium 1.25 mmol/l		Calcium 1.75 mmol/l		changes
	pre	post	pre	post	1.25 vs 1.75
ionized Calcium	1.13 (0.15)	1.11 (0.11)	1.13 (0.16)	1.33 (0.15) ^a	<0.001
Sodium	137.3 (1.5)	138.2 (2.4)	137.0 (3.4)	136.2 (3.7)	ns
Potassium	5.0 (0.8)	3.5 (0.3) ^a	5.1 (1.0)	3.6 (0.3) ^a	ns
total CO ₂	23.3 (2.9)	28.3 (1.8) ^a	23.1 (3.5)	28.2 (1.7) ^a	ns
Urea	24.2 (5.0)	10.0 (2.7) ^a	25.5 (5.8)	10.4 (3.0) ^a	ns
Osmolality	309.0 (5.4)	291.9 (8.1) ^a	308.1 (12.7)	292.1 (9.6) ^a	ns
COP	3.4 (0.6)	4.2 (0.6) ^a	3.3 (0.4)	4.2 (0.6) ^a	ns
Noradrenaline	1.84 (0.61)	2.22 (0.81)	1.71 (0.60)	2.41 (0.78) ^a	0.037
Adrenaline	0.23 (0.12)	0.43 (0.70)	0.18 (0.07)	0.28 (0.19)	ns

ionized Calcium, Sodium, Potassium, total CO₂ and Urea (mmol/l); Osmolality (mOsm/kg); COP (kPa); Catecholamines (nmol/l); post vs pre, ^a= $p < 0.01$.

Figure 5.1 Plasma ionized calcium during low (—) and high (---) calcium UF+HD; *= $p < 0.01$ as compared to pre-treatment level

The data on blood pressure and vascular reactivity are shown in table 5.2 and figure 5.2a-c. During low-calcium UF+HD, both systolic and mean arterial BP decreased significantly while there were no significant changes in the different BP parameters during high-calcium UF+HD. Moreover, both systolic and diastolic as

well as mean arterial BP were significantly lower during low-calcium UF+HD as compared to high-calcium UF+HD. After 3 hours UF+HD, the differences were respectively 14 ± 10 , 5 ± 7 and 9 ± 9 mmHg. HR increased significantly and comparably during both low- and high-calcium UF+HD. Concerning vascular reactivity, there were no significant differences in both FVR and VT between low- and high-calcium UF+HD. During both treatments, FVR increased significantly whereas VT decreased significantly.

Table 5.2 Hemodynamic data: blood pressure and vascular reactivity

	Calcium 1.25 mmol/l		Calcium 1.75 mmol/l		changes 1.25 vs 1.75
	pre	post	pre	post	
Systolic BP	139 (20)	121 (18) ^a	140 (26)	135 (19)	0.010
Diastolic BP	8 (12)	77 (13)	81 (12)	82 (11)	0.022
Mean Arterial BP	105 (16)	96 (18) ^b	105 (19)	105 (16)	0.021
Heart Rate	69 (7)	79 (10) ^a	69 (10)	77 (13) ^a	ns
FVR	1626 (729)	2480 (1205) ^a	1555 (826)	2365 (1226) ^b	ns
VT	16.9 (4.6)	14.7 (5.0) ^b	16.6 (4.5)	14.3 (4.0) ^b	ns

BP (mmHg), HR (beats/min), FVR (mmHg/ml/100ml/sec), VT (mmHg/ml/100ml); post vs pre, ^a= $p < 0.01$, ^b= $p < 0.05$.

Figure 5.2a Arterial BP during low (—) and high (---) calcium UF+HD

*= $p < 0.01$, #= $p < 0.05$ as compared to pre-treatment level

Figure 5.2b FVR during low (—) and high (---) calcium UF+HD; *= $p < 0.01$, #= $p < 0.05$ as compared to pre-treatment level

Figure 5.2c VT during low (—) and high (---) calcium UF+HD; *= $p < 0.05$ as compared to pre-treatment level

The echocardiographic data are shown in table 5.3. With a comparable pre-dialysis LVEDD in both high- and low-calcium UF+HD there were no significant changes in left ventricular end-diastolic and end-systolic dimensions during both treatment modalities. Both SV (-18 ± 13 ml) and CO (-1.3 ± 1.5 l/min) decreased significantly during low-calcium UF+HD whereas during high-calcium UF+HD the decrease in SV (-8 ± 11 ml) and CO (-0.7 ± 1.2 l/hr) did not reach statistical significance. There were no significant changes in calculated systemic vascular resistance during both treatments.

Table 5.3 Hemodynamic data: echocardiography before and after UF+HD

	Calcium 1.25		Calcium 1.75		changes 1.25 vs 1.75
	pre	post	pre	post	
LVEDD	48 (6)	46 (6)	48 (6)	46 (8)	ns
LVESD	33 (4)	31 (6)	31 (5)	30 (5)	ns
Heart Rate	72 (10)	72 (9)	76 (13)	76 (12)	ns
Stroke Volume	92 (22)	73 (25) ^a	87 (20)	79 (15)	ns
Cardiac Output	6.5 (1.5)	5.3 (1.7) ^a	6.5 (1.4)	5.9 (0.4)	ns
Cardiac Index	3.5 (0.7)	2.8 (1.0) ^a	3.5 (0.5)	3.2 (0.2)	ns
Systemic Vasc Res	1392 (424)	1674 (796)	1385 (420)	1478 (266)	ns

LVEDD and LVESD (mm); HR (beats/minute); SV (ml); CO (l/min); CI (l/min/m²); SVR (dyne/sec/cm⁵); post vs pre, ^a = $p < 0.05$.

Discussion

Arterial blood pressure and different cardiovascular parameters were measured during low (1.25) and high (1.75) calcium UF+HD to study if, in addition to myocardial contractility, vascular reactivity is also influenced by changes in ionized calcium. In addition, the effect of lowering the dialysate calcium concentration on intradialytic blood pressure was assessed. Our results show that even in hemodynamically stable patients, ionized calcium is an important determinant of the blood pressure response during UF+HD. Moreover, it is shown that this relation is exclusively based on changes in myocardial contractility.

Plasma ionized calcium increased significantly and was significantly higher during high-calcium UF+HD. With a dialysate calcium concentration of 1.25 mmol/l plasma ionized calcium remained more or less unchanged. In high-calcium UF+HD, the increase in plasma ionized calcium was largest during the first hour after which it reached a plateau. When comparing our data with previous studies, the data show that dialysate calcium concentrations of ± 1.35 or higher are associated with a net positive calcium balance^{4,8,14-16}.

With all other dialysis parameters standardised, exactly comparable pre-dialysis weights as well as changes in relative blood volume, systolic, diastolic and mean arterial blood pressure were all substantially and significantly higher during high-calcium UF+HD as compared to low-calcium UF+HD. Only during the latter treatment, both systolic and mean arterial blood pressure decreased significantly. So, even in hemodynamically stable patients selected for their good cardiac function, changes in plasma ionized calcium are a very important determinant of the blood pressure response during dialysis therapy. In agreement with these observations, most authors also found a clinically important positive

correlation between changes in plasma ionized calcium and arterial blood pressure during dialysis therapy^{4,7,8}. At variance, Sherman et al. found only a small but significant decrease in mean arterial BP during low-calcium UF+HD (1.25 mmol/l) as compared to high-calcium UF+HD (1.75 mmol/l). With a maximum mean reduction in mean arterial BP of 4.6 mmHg they concluded that dialysate calcium could be reduced if necessary with only a clinically minor effect on intradialytic blood pressure¹⁷. A relatively low UF-rate of ± 0.56 l/hr and differences concerning their methodology might partly explain these findings. However, the bulk of evidence points to an important role of changes in ionized calcium on the blood pressure response during dialysis therapy which should be taken into consideration when lowering the dialysate calcium concentration in treating hypercalcemia in dialysis patients.

A main new finding in the present study is the fact that the peripheral vascular response during UF+HD is not directly influenced by differences in plasma ionized calcium. During both low- and high-calcium UF+HD, FVR increased whereas VT decreased with no significant differences between the two treatments. Moreover, calculated SVR was also comparable. However, both SV and CO decreased significantly only during low-calcium UF+HD while changes in SV and CO were smaller and insignificant during high-calcium UF+HD. So, with comparable changes in relative blood volume and comparable end-diastolic dimensions, indicative of a comparable pre-load, blood pressure was maintained at a higher level during high-calcium UF+HD as a result of an improvement in myocardial contractility only and not on the basis of differences in peripheral vascular resistance or venous tone. In agreement, in isovolemic dialysis as well as in dialysis combined with ultrafiltration different authors also found a positive correlation between changes in plasma ionized calcium and left ventricular contractility^{4,6}. Moreover, both Fellner et al. and Lang et al. also found no significant differences in calculated SVR after dialysis at three different dialysate calcium concentrations^{4,6}. However, their conclusions are based on post-dialytic echocardiographies only.

Whereas plasma adrenaline remained unchanged during both low- and high-calcium UF+HD, plasma noradrenaline increased significantly only during the latter treatment. The difference between the two treatments was statistically significant. These results suggest that also during dialysis the release of noradrenaline from sympathetic nerve endings is calcium dependent¹⁸. In healthy volunteers and patients with a mildly impaired renal function, Marone et al. found that infusion of calcium gluconate (15 mg/kg body weight) resulted in a small but significant increase in plasma noradrenaline after three hours. Interesting and in contrast with our results, the accompanying increase in systolic and diastolic blood pressure in their study was based on an increase in peripheral resistance whereas

cardiac output remained unchanged¹⁹. Therefore, the finding of higher level of plasma noradrenaline with no improvement in the peripheral vascular response could be interpreted as a further confirmation that vascular reactivity is impaired during UF+HD. The present data do not allow us to conclude whether this is due to either a post-receptor defect²⁰ or the presence of some disturbing factor such as net heat load²¹ or the effects of dialysate derived contaminants²².

In conclusion, lowering the dialysate calcium concentration from 1.75 to 1.25 mmol/l has a clinically important negative effect on the blood pressure response during combined ultrafiltration-hemodialysis. This relation is based on differences in myocardial contractility only and not on differences in peripheral vascular tone. Especially in cardiac compromised patients, this could be of clinical importance.

References

- 1 Daugirdas JT: Dialysis hypotension: A hemodynamic analysis. *Kidney Int* 1991; 39: 233-246.
- 2 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
- 3 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
- 4 Fellner SK, Lang RM, Neumann A, Spencer KT, Bushinsky DA, Borow KM: Physiological mechanisms for calcium-induced changes in systemic arterial pressure in stable dialysis patients. *Hypert* 1989; 13: 213-218.
- 5 Henrich WL, Hunt JM, Nixon JV: Increased ionized calcium and left ventricular contractility during hemodialysis. *N Engl J Med* 1984; 310: 19-23.
- 6 Lang RM, Fellner SK, Neumann A, Bushinsky DA, Borow KM: Left ventricular contractility varies directly with blood ionized calcium. *Ann Intern Med* 1988; 108: 524-529.
- 7 Leunissen KML, van den Berg BW, van Hooff JP: Ionized calcium plays a pivotal role in controlling blood pressure during haemodialysis. *Blood Purif* 1989; 7: 233-239.
- 8 Maynard JC, Cruz C, Kleerekoper M, Levin NW: Blood pressure response to changes in serum ionized calcium during hemodialysis. *Ann Intern Med* 1986; 104: 358-361.
- 9 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
- 10 Ihlen H, Endresen K, Myreng Y, Myhre E: Reproducibility of cardiac stroke volume estimated by doppler echocardiography. *Am J Cardiol* 1987; 59: 975-978.
- 11 Dittmann H, Voelker W, Karsch K-R, Siepel L: Influence of sampling site and flow area on cardiac output measurements by doppler echocardiography. *J Am Coll Cardiol* 1987; 10: 818-823.
- 12 Robson SC, Boys RJ, Hunter S: Doppler echocardiographic estimation of cardiac output: analysis of temporal variability. *Eur Heart J* 1988; 9: 313-318.
- 13 de Vries JPPM, Olthof CG, Viosser V, Kouw PM, van Es A, Donker AJM, de Vries PMJM: Continuous measurement of blood volume during hemodialysis by an optical method. *ASAIO Journal* 1992; 38: M181-M185.
- 14 Goldsmith RS, Furszyfer J, Johnson WJ, Beeler GW, Taylor WF: Calcium flux during hemodialysis. *Nephron* 1978; 20: 132-140.
- 15 Hou SH, Zhao J, Ellman CF, Hu J, Griffin Z, Spiegel DM, Bourdeau JE: Calcium and phosphorus fluxes during hemodialysis with low calcium dialysate. *Am J Kidney Dis* 1991; 18: 217-224.

-
- 16 Argilès A, Kerr PG, Canaud B, Flavier JL, Mion C: Calcium kinetics and the long-term effects of lowering dialysate calcium concentration. *Kidney Int* 1993; 43: 630-640.
 - 17 Sherman RA, Bialy GB, Gazinsky B, Bernholc AS, Eisinger RP: The effect of dialysate calcium levels on blood pressure during hemodialysis. *Am J Kidney Dis* 1986; 8: 244-247.
 - 18 Rubin RP: Neurotransmitter substances and hormones. *Pharmacol Rev* 1970; 22: 389-428.
 - 19 Marone C, Beretta-Piccoli C, Weidmann P: Acute hypercalcemic hypertension in man: role of hemodynamics, catecholamines and renin. *Kidney Int* 1980;20:92-96.
 - 20 Brodde O-E, Daul A: Alpha- and Beta-adrenoreceptor changes in patients on maintenance hemodialysis. *Contrib Nephrol* 1984; 41: 99-107.
 - 21 van Kuijk WHM, Luik AJ, de Leeuw PW, van Hooff JP, Nieman FHM, Habets HML, Leunissen KML: Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences. *Nephrol Dial Transplant* 1995;(in press).
 - 22 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
-

Chapter 6

Critical role of the extracorporeal blood temperature in the hemodynamic response during hemofiltration

WHM van Kuijk, D Hillion, C Savoiu, KML Leunissen

(submitted for publication)

Abstract

Background

Impaired vascular reactivity during combined ultrafiltration-hemodialysis (UF+HD) as compared to hemofiltration (HF) remains a rather enigmatic problem of which the causes are still not well understood. Although a number of factors have been claimed to be responsible most recent studies point to a major role of the extracorporeal blood temperature which is normally lower during HF as compared to UF+HD. However, previous studies in which hemodynamics were studied during UF+HD and HF in relation to the extracorporeal blood temperature are limited by the use of acetate in UF+HD while measurements were often confined to blood pressure and heart rate.

Methods

Therefore, we measured arterial blood pressure and forearm vascular resistance (FVR) and venous tone (VT) (strain-gauge plethysmography) in 11 hemodialysis patients during three hours UF+HD (37.5°C) and pre-dilution HF (39.0°C=warm HF) resulting in equivalent extracorporeal blood temperatures. In addition, patients were also studied during cold HF at an infusate temperature of 36.0°C. UF+HD and HF were matched with respect to the dialysate and infusate composition (bicarbonate), biocompatibility factors and small molecule clearance.

Results

At equivalent temperatures, UF+HD and HF were associated with a comparable vascular and blood pressure response. Only cold HF was associated with a significant increase in FVR. In addition, FVR and VT as well as arterial blood pressure were all significantly higher during cold HF compared to both UF+HD and warm HF.

Conclusions

These results indicate that disparity in vascular reactivity between UF+HD and HF is primarily related to differences in the extracorporeal blood temperature.

Introduction

With an increasing number of elderly, cardiovascular compromised patients on dialysis therapy hemodynamic instability with symptomatic hypotension will remain an important clinical problem. Since their clinical application, convective dialysis treatments such as hemofiltration are considered to be the most optimal treatment modality for hemodynamically unstable patients as blood pressure stability is better during these treatments compared to diffusive treatments such as combined ultrafiltration-hemodialysis¹. In previous studies it has been shown that these differences in blood pressure stability are related to the fact that peripheral vascular resistance increases adequately during hemofiltration while peripheral vascular resistance remains more or less unchanged during combined ultrafiltration-hemodialysis^{2,3}. In addition, we found that venous reactivity is also impaired during the latter therapy⁴. Although hemofiltration seems to be the most optimal treatment for cardiovascular compromised patients its clinical application is still limited due to the higher cost and limitations in clearance with higher hematocrits as a result of the use of erythropoietin. If we are able to unravel the mechanisms responsible for disparity in vascular reactivity between convective and diffusive dialysis treatments it should be possible to improve hemodynamic stability during the latter treatment modality.

Traditionally, the larger decrease in plasma osmolality during combined ultrafiltration-hemodialysis as a result of a larger small molecule clearance was thought to play a pivotal role in the impaired vascular response⁵. Urea supplemented dialysate, however, did not improve the hemodynamic response in dogs and men^{6,7}. Moreover, we found a comparable vascular response during high- and low-sodium dialysis⁸. Henderson et al. suggested that the impaired vascular response during combined ultrafiltration-hemodialysis might be related to bio-incompatibility factors related to the dialysis membrane or to the presence of non-sterile dialysate⁹. However, vascular reactivity did not improve when using sterile dialysate while Aakhus et al. measured comparable central hemodynamics using membranes of low and high biocompatibility^{10,11}. Several studies have pointed to the importance of the extracorporeal blood temperature¹²⁻¹⁴ which is normally lower in hemofiltration as a result of the lower infusate flow and the larger extracorporeal circuit. However, results from comparative studies are confounded by the use of acetate as a single buffer in combined ultrafiltration-hemodialysis while hemodynamics and changes in blood volume were not always monitored. In addition, hemodialysis and hemofiltration performed in post-dilution mode were not matched for small molecule clearance and biocompatibility^{15,16}.

Therefore, we measured arterial blood pressure as well as vascular reactivity under strictly standardised conditions and in relation to the extracorporeal blood

temperature comparing combined ultrafiltration-hemodialysis with pre-dilution hemofiltration. In both treatments bicarbonate was used as a buffer.

Subjects and methods

Subjects and dialysis

After obtaining informed consent, four women and seven men on intermittent hemodialysis were included with a mean age of 43 years (range 26-58) and a mean time on dialysis of 65 months (range 5-252). All patients were hemodynamically stable patients who rarely suffered from intradialytic hypotension. The following exclusion criteria were used: (1) severe coronary (NYHA II or more) or valvular heart disease, (2) compromised left ventricular function (ejection fraction 30% or less) and (3) diabetes mellitus. By adjusting fluid intake patients were instructed to achieve a pre-dialysis weight in the range from 2.0 to 3.0 kilograms above their dry-weight that was estimated by echography of the inferior caval vein¹⁷. Vasoactive medication was stopped 48 hours before the study while patients were not allowed to drink caffeine containing beverages during 12 hours preceding the study. Recirculation was less than 10% in the studied population.

With each patient serving as his or her own control, measurements were performed during three consecutive dialysis treatments in randomised order. Blood flow was set at 300 ml/min in all treatments while the same ultrafiltration rate was used within the same patient. One session consisted of three hours combined ultrafiltration-hemodialysis (UF+HD) using a polyamide dialyser (Polyflux 160, Gambro, Sweden), a dialysate temperature of 37.5°C and a dialysate flow of 500 ml/min. The other two sessions consisted of three hours pre-dilution hemofiltration (HF) using a polyamide hemofilter (FH 88 H, Gambro, Sweden) and an infusate flow of 350 ml/min to obtain an almost comparable small molecule clearance as in UF+HD. In one session the infusate temperature was set at 39.0°C (warm HF) to obtain the same blood temperature in the efferent, venous line as during UF+HD. On the basis of former experiments, the infusate temperature was set at 36.0°C in cold hemofiltration (cold HF) to obtain a comparable extracorporeal blood temperature as in post-dilution HF.

UF+HD and HF were performed with an AK-100 Ultra monitor (Gambro, Sweden) to prepare on-line sterile infusate and dialysate. The composition of both dialysate and infusate was: sodium 141 mmol/l, potassium 2.0 mmol/l, calcium 1.75 mmol/l, bicarbonate 34 mmol/l, acetate 3 mmol/l, magnesium 0.5 mmol/l and chloride 108 mmol/l.

Methods

The luminal blood temperature was measured in the arterial and venous line with needle thermometers (Hewlett-Packard 78214C temperature monitor). Connectors with a built-in adaptor were placed in the arterial and venous line under sterile conditions at a distance of about 15 centimetres from the patient. There was no direct contact between blood and thermometer.

Forearm vascular resistance (FVR) and venous tone (VT) were measured as described in chapter 2. The hand circulation was occluded from one minute before as well as during flow measurements. Arterial blood pressure (BP) and heart rate (HR) were measured as described in chapter 3. All hemodynamic measurements were performed before and every 30 minutes during dialysis.

Changes in relative blood volume (BV) were measured continuously and non-invasively by means of an optical reflection method which operates at a wavelength of 950 nm. The optical sensor was clipped on the arterial line (Haemoguard 2000, Sanofi Sante, The Netherlands)¹⁸. The baseline value was obtained after two minutes of extracorporeal circulation at a blood flow of 250 ml/min without ultrafiltration to exclude the influence of saline (recirculation) present in the blood lines at the start of dialysis.

Both before and after dialysis, blood was withdrawn for the estimation of: sodium, potassium, ionized calcium (NOVA 8, NOVA Biomedical, Waltham, MA, USA) total CO₂, urea, osmolality (Wescor, vapour pressure monitor) and colloid osmotic pressure. In addition, blood was withdrawn in ice-chilled tubes and directly centrifuged at a temperature of 4°C for the estimation of plasma catecholamines (HPLC, fluorescence detection).

Statistical analysis

Changes in hemodynamic parameters within each treatment as well as differences between treatments were analyzed by repeated measurements MANOVA (SPSS-PC version). If the sphericity of the variance-covariance matrix appeared to be violated, degrees of freedom in the univariate MANOVA tests were corrected by the Greenhouse-Geisser epsilon to avoid Type 1 error in testing the F-ratio. All laboratory parameters were analyzed by Friedmans ANOVA and if appropriate, by the Wilcoxon's signed-rank test. All values are expressed as mean $\pm$ SD.

Results

The pre-dialysis weights in UF+HD, warm HF and cold HF were respectively 69.2 ± 10.8 , 69.3 ± 11.1 and 69.5 ± 10.9 kg (ns) while the mean ultrafiltration rate was 0.92 ± 0.12 l/hr. The blood temperatures are shown in figure 6.1a-c. UF+HD and warm HF resulted in an almost comparable arterial and venous blood temperature. After three hours, the arterial blood temperature increased by $0.4 \pm 0.2^\circ\text{C}$ in both treatments ($p < 0.01$). In addition, the venous blood temperature increased by 0.2 ± 0.2 ($p < 0.05$) and $0.3 \pm 0.2^\circ\text{C}$ ($p < 0.01$) respectively. In cold-HF, blood cooled down during extracorporeal circulation by 1.2°C while after three hours the arterial blood temperature had increased by $0.1 \pm 0.2^\circ\text{C}$ ($p < 0.05$) without significant changes in the venous blood temperature.

Figure 6.1a Arterial (—) and venous (- - -) blood temperatures in UF+HD
*= $p < 0.01$ #= $p < 0.05$ as compared to $t=0$ min.

Figure 6.1b Arterial (—) and venous (- - -) blood temperatures in warm HF
*= $p < 0.01$ as compared to $t=0$ min.

Figure 6.1c Arterial (—) and venous (---) blood temperatures in cold HF

*= $p < 0.05$ as compared to $t=0$ min.

The laboratory data are shown in table 6.1. All laboratory parameters changed comparably during the different treatment modalities. The decrease in plasma Potassium, Urea and Osmolality and the increase in total CO_2 tended to be smaller during both warm and cold HF as compared to UF+HD but these differences did not reach statistical significance. In addition, differences in the urea reduction ratio (URR) were not statistically significant. There were no significant changes in plasma adrenaline or noradrenaline. Relative blood volume decreased significantly ($p < 0.01$) and comparably to 86 ± 6 , 84 ± 6 and $85 \pm 6\%$ during respectively UF+HD, warm HF and cold HF.

The hemodynamic data are presented in table 6.2 and figure 6.2a-c. UF+HD and warm HF were associated with comparable vascular and blood pressure responses with no significant differences between the two treatments. In both, FVR tended to increase during the first hour ($p > 0.05$) after which it stabilised while VT decreased significantly. In contrast, Cold HF was associated with a significant increase in FVR, especially during the first two hours while VT remained unchanged. As compared to both UF+HD and warm HF, FVR ($p < 0.01$) and VT ($p < 0.05$) were significantly higher during cold HF. There were no significant changes in arterial blood pressure during UF+HD. During warm HF, only systolic blood pressure decreased significantly while heart rate increased significantly during both UF+HD and warm HF. In cold HF, all three blood pressure parameters increased significantly while heart rate decreased. Mean arterial BP was significantly higher as compared to both UF+HD and warm HF ($p < 0.05$). Differences in systolic ($p < 0.05$) and diastolic BP ($p < 0.05$) reached statistical significance only between warm and cold HF.

Table 6.1 Laboratory data

	UF+HD 37.5°C		HF 39.0°C		HF 36.0°C	
	pre	post	pre	post	pre	post
Sodium	139.0 (2.1)	140.1 (2.1) ^a	137.4 (2.5)	139.6 (1.9) ^a	138.2 (3.0)	139.6 (2.1) ^a
Potassium	4.7 (0.6)	3.5 (0.4) ^a	4.7 (0.5)	3.8 (0.4) ^a	4.8 (0.7)	3.7 (0.4) ^a
ionized Calcium	1.17 (0.13)	1.40 (0.08) ^a	1.14 (0.13)	1.37 (0.07) ^a	1.14 (0.11)	1.34 (0.07) ^a
total CO ₂	22.7 (2.8)	25.6 (1.4) ^a	23.1 (3.2)	24.8 (1.5) ^b	23.0 (3.0)	24.9 (1.6)
Urea	25.0 (7.3)	9.7 (3.5) ^a	25.9 (6.5)	11.9 (3.4) ^a	23.9 (5.5)	11.8 (2.8) ^a
URR		59.5 (15.9)		53.5 (10.4)		53.5 (6.8)
Osmolality	295.3 (7.7)	283.2 (5.6) ^a	297.4 (7.3)	288.7 (6.1) ^a	296.0 (7.1)	286.4 (4.5) ^b
COP	2.9 (0.3)	4.4 (0.4) ^a	2.9 (0.3)	4.4 (0.6) ^a	2.9 (0.2)	4.3 (0.3) ^a
Noradrenaline	1.27 (0.47)	1.58 (0.51)	1.10 (0.87)	1.46 (0.52)	1.24 (0.46)	1.69 (0.85)
Adrenaline	0.12 (0.11)	0.13 (0.10)	0.15 (0.12)	0.17 (0.16)	0.13 (0.07)	0.16 (0.15)

Sodium, Potassium, ionized Calcium, total CO₂ and Urea (mmol/l); URR (%); Osmolality (mOsm/kg); COP (kPa); Catecholamines (nmol/l); post vs pre; ^a = $p < 0.01$, ^b = $p < 0.05$

Table 6.2 Hemodynamic data

	UF+HD 37.5°C		HF 39.0°C		HF 36.0°C	
	pre	post	pre	post	pre	post
Systolic BP	129 (17)	126 (25)	133 (20)	125 (18) ^a	131 (20)	137 (26) ^a
Diastolic BP	77 (9)	80 (10)	80 (11)	82 (9)	79 (10)	85 (9) ^a
Mean Arterial BP	97 (12)	98 (17)	101 (15)	100 (12)	98 (15)	106 (15) ^a
Heart Rate	74 (7)	78 (16) ^b	70 (6)	75 (14) ^b	72 (7)	68 (12) ^a
FVR	1524 (606)	2186 (2318)	1704 (664)	2366 (1282)	1588 (602)	5110 (2688) ^a
VT	18.6 (5.2)	15.7 (5.0) ^a	19.2 (3.3)	15.2 (2.8) ^a	17.9 (5.0)	18.0 (4.8)

BP (mmHg); HR (beats/min); FVR (mmHg/ml/100ml/sec); VT (mmHg/ml/100ml); post vs pre; ^a = $p < 0.01$, ^b = $p < 0.05$

Figure 6.2a Arterial BP during UF+HD (—), warm HF (---) and cold HF (···); * = $p < 0.01$ as compared to pre-treatment level

Figure 6.2b FVR during UF+HD (—), warm HF (---) and cold HF (···); * = $p < 0.01$ as compared to pre-treatment level

Figure 6.2c VT during UF+HD (—), warm HF (---) and cold HF (···); * = $p < 0.01$ as compared to pre-treatment level

Discussion

Our study is the first to show that at equivalent extracorporeal blood temperatures the vascular and blood pressure responses are comparable between UF+HD and HF. Only during extracorporeal blood cooling, HF is associated with an increase in peripheral resistance and a more stable venous tone resulting in a higher blood pressure.

To match UF+HD and HF with respect to the extracorporeal blood temperature the infusate temperature was increased to 39.0°C in pre-dilution HF which resulted in an almost comparable temperature of blood in the arterial and venous line. In contrast, HF at an infusate temperature of 36.0°C was associated with cooling of blood in the extracorporeal circuit. The significant increase in the arterial blood temperature after three hours UF+HD and warm HF indicates that the net thermal balance was positive in both treatments which is in agreement with previous studies^{19,20}. Although very little, the arterial blood temperature increased significantly also during cold HF. This suggests that despite extracorporeal blood cooling the patient's body temperature increased which might be related to either dialysis induced catabolism²¹ or to a decrease in heat loss as a result of baroreceptor mediated vasoconstriction of the skin vasculature.

At comparable pre-dialysis weights and changes in intravascular volume and with UF+HD and HF matched for the buffer substrate, small molecule clearances and biocompatibility factors, only cold HF was associated with a physiological hemodynamic response to compensate for the decrease in circulating blood volume. FVR and VT were significantly different during cold HF compared to both UF+HD and warm HF without significant differences between the latter two treatments. These results indicate that the well described differences in hemodynamics between UF+HD and HF are primarily related to thermal differences. The curves in FVR suggest that even UF+HD and warm HF were associated with an increase in peripheral vascular tone during the first hour. Thereafter this increase seems to be counteracted most probably as a result of the net positive thermal balance. In agreement, Baldamus et al. also found an increase in TPR during the first two hours of bicarbonate dialysis after which it tended to decrease during the following two hours (dialysate temperature is not given)³. In contrast, the cold stimulus during cold HF initiated an instantaneous increase in peripheral resistance as well as in mean arterial pressure of which the former continued to increase during the second hour.

Our results are at variance with results obtained by Maggiore et al. and Fox et al. who concluded that differences in the extracorporeal blood temperature are only partially responsible for disparity in hemodynamic stability. However, Maggiore et al. only measured blood pressure and heart rate while acetate was used as a single

buffer in UF+HD while using lactate in HF¹⁵. The vasodilatory capacities of acetate are well described. In the study of Fox et al. differences in the extracorporeal blood temperature are not excluded, as in their study temperature measurements were confined to measurements of core body temperature while the same temperature was used for dialysate and substitution fluid. In addition, acetate was used in both UF+HD and HF¹⁶.

Although plasma catecholamines only weakly reflect overall sympathetic activity²² our data on plasma catecholamines do not confirm that improved vascular reactivity in cold HF is the result of an increase in sympathetic tone^{13,23}. More direct measurements of sympathetic activity are necessary to be more conclusive on this subject²⁴. In addition, it is unknown whether the effects of temperature on vascular tone are more generalised or remain confined to the skin vascular bed^{25,26}.

In conclusion, the present study shows that disparity in vascular reactivity between UF+HD and HF is primarily related to differences in the extracorporeal blood temperature which indicates that it should be possible to obtain a comparable hemodynamic stability in UF+HD as in HF by lowering the dialysate temperature. Future studies are necessary to determine the optimal dialysate and infusate temperature and to investigate the possible adverse effects of cold dialysis on the incidence of cardiac arrhythmias and tissue oxygenation.

References

- 1 Quellhorst E, Dohr B, Schuenemann B: Hemofiltration: treatment of renal failure by ultrafiltration and substitution. *J Dial* 1977; 1: 529-543.
- 2 Hampl H, Paeppler H, Unger V, Fischer C, Resa I, Kessel M: Hemodynamic changes during hemodialysis, sequential ultrafiltration, and hemofiltration. *Kidney Int* 1980; 18: S83-S88.
- 3 Baldamus CA, Ernst W, Frei U, Koch KM: Sympathetic and hemodynamic response to volume removal during different forms of renal replacement therapy. *Nephron* 1982; 31: 324-332.
- 4 Kooman JP, Gladziwa U, Böcker G, van Bortel LMAB, van Hooff JP, Leunissen KML: Role of the venous system in hemodynamics during ultrafiltration and bicarbonate dialysis. *Kidney Int* 1992; 42: 718-726.
- 5 Henrich WL, Woodard TD, Blachley JD, Gomez-Sanchez C, Pettinger W, Cronin RE: Role of osmolality in blood pressure stability after dialysis and ultrafiltration. *Kidney Int* 1980; 18: 480-488.
- 6 Wehle B, Asaba H, Castenfors J, Gunnarsson B, Bergström J: Influence of dialysate composition on cardiovascular function in isovolemic haemodialysis. *Proc EDTA* 1981; 18: 153-159.
- 7 Daugirdas JT, Ing TS, Chen W-T, Vestal RE, Hano JE, Al-Kudsi RR: Ultrafiltration hemodynamics in conscious, uremic dogs: Effect of a decreasing plasma urea level. *Mineral Electrolyte Metab* 1985; 11: 79-84.
- 8 van Kuijk WHM, Wirtz JJM, Grave W, de Heer F, van Hooff JP, de Leeuw PW, Leunissen KML: The effect of high and low sodium dialysate on vascular reactivity during hemodialysis (abstract). *Nephrol Dial Transplant* 1995; 10: 1028.
- 9 Henderson LW, Koch KM, Dinarello CA, Shaldon S: Hemodialysis hypotension: The interleukin-1 hypothesis. *Blood Purif* 1983; 1: 3-8.
- 10 van Kuijk WHM, Buurman WA, Gerlag PGG, Leunissen KML: Is vascular reactivity improved by using ultrapure dialysate ? (abstract). *Nephrol Dial Transplant* 1995; 10: 1028.
- 11 Aakhus S, Bjoernstad K, Jorstad S: Systemic cardiovascular response in hemodialysis without and with ultrafiltration with membranes of high and low biocompatibility. *Blood Purif* 1995; 13: 229-240.
- 12 Coli U, Landini S, Lucatello S, Fracasso A, Morachiello P, Righetto F, Scanferla F, Onesti G, Bazzato G: Cold as cardiovascular stabilizing factor in hemodialysis: hemodynamic evaluation. *Trans Am Soc Artif Intern Organs* 1983; 29: 71-75.
- 13 Mahida BH, Dumlér F, Zasuwa G, Fleig G, Levin NW: Effect of cooler dialysate on serum catecholamines and blood pressure stability. *Trans Am Soc Artif Intern Organs* 1983; 24: 384-389.
- 14 Jost CMT, Agarwal R, Khair-El-Din T, Grayburn PA, Victor RG, Henrich WL: Effects of cooler temperature dialysate on hemodynamic stability in "problem" dialysis patients. *Kidney Int* 1993; 44: 606-612.

-
- 15 Maggiore Q, Pizzarelli F, Sisca S, Zoccalli C, Parlongo S, Nicolò F, Creazzo G: Blood temperature and vascular stability during hemodialysis and hemofiltration. *Trans Am Soc Artif Intern Organs* 1982; 28: 523-527.
 - 16 Fox SD, Henderson LW: Cardiovascular response during hemodialysis and hemofiltration: thermal, membrane, and catecholamine influences. *Blood Purif* 1993; 11: 224-236.
 - 17 Leunissen KML, Kouw P, Kooman JP, Cheriex EC, de Vries PMJM, Donker AJM, van Hooff JP: New techniques to determine fluid status in hemodialyzed patients. *Kidney Int* 1993; 43: S50-S56.
 - 18 de Vries JPPM, Olthof CG, Viosser V, Kouw PM, van Es A, Donker AJM, de Vries PMJM: Continuous measurement of blood volume during hemodialysis by an optical method. *ASAIO Journal* 1992; 38: M181-M185.
 - 19 Lindholm T, Thysell H, Yamamoto Y, Forsberg B, Gullberg CA: Temperature and vascular stability in hemodialysis. *Nephron* 1985; 39: 130-133.
 - 20 van Kuijk WHM, Luik AJ, de Leeuw PW, van Hooff JP, Nieman FHM, Habets HML, Leunissen KML: Vascular reactivity during hemodialysis and isolated ultrafiltration: thermal influences. *Nephrol Dial Transplant* 1995; (in press).
 - 21 Gutierrez A, Alvestrand A, Bergström J, Beving H, Lantz B, Henderson LW: Biocompatibility of hemodialysis membranes: A study in healthy subjects. *Blood Purif* 1994; 12: 95-105.
 - 22 Esler MD, Hasking GJ, Willett IR, Leonard PW, Jennings GL: Noradrenaline release and sympathetic nervous system activity. *J Hypert* 1985; 3: 117-129.
 - 23 Bergström J: Catecholamines and control of blood pressure during hemodialysis and hemofiltration. *Kidney Int* 1988; 34: S110-S114.
 - 24 Converse RL, Jacobsen TN, Jost CMT, Toto RD, Grayburn PA, Obregon TM, Fouad-Tarazi F, Victor RG: Paradoxical withdrawal of reflex vasoconstriction as a cause of hemodialysis-induced hypotension. *J Clin Invest* 1992; 90: 1657-1665.
 - 25 Rowell BR: Cardiovascular adjustment to thermal stress; in Shepherd JT, Abboud FM (ed.): *Handbook of Physiology*, sect 2, Cardiovascular system, vol 3, Peripheral circulation and organ blood flow. Bethesda, American Physiological Society, 1983, 3, pp. 967-1023
 - 26 Vanhoutte PM: Physical factors of regulation; in Bohr DF, Somlyo AP (ed.): *Handbook of Physiology*, sect 2, Cardiovascular System, vol 2, Vascular smooth muscle. Bethesda, American Physiological Society, 1980, 2, pp. 443-474
-

Chapter 7

**Summary and concluding
remarks**

Summary

Hemodynamic instability with symptomatic hypotension is still one of the main problems in dialysis therapy. Therefore it is of the utmost importance to improve dialysis with respect to intradialytic blood pressure regulation especially since the number of elderly, cardiovascular compromised patients on dialysis therapy is still increasing. In this respect, it is well established that blood pressure stability is better when fluid is withdrawn during either isolated ultrafiltration or hemofiltration compared to combined ultrafiltration-hemodialysis. In **chapter 1** the literature is reviewed on the possible role of plasma volume preservation and cardiovascular compensatory mechanisms in differences in hemodynamic stability among different dialysis strategies. Both mechanisms play a central role in blood pressure regulation during dialysis therapy. From the bulk of evidence it can be concluded that disparity in hemodynamic stability between different dialysis strategies is primarily related to differences in vascular reactivity. Isolated ultrafiltration and hemofiltration are associated with a physiological increase in peripheral vascular tone when circulating blood volume decreases whereas both arteriolar and venous reactivity are impaired during combined ultrafiltration-hemodialysis. However, the mechanism responsible for these differences in vascular reactivity is still not well understood. The present studies were designed to investigate pathophysiologic mechanisms responsible for disparity in vascular reactivity between different dialysis strategies. In all studies, forearm vascular reactivity was measured during dialysis therapy (bicarbonate) in cardiovascular stable patients using strain-gauge plethysmography.

The extracorporeal blood temperature has been shown to be an important determinant of the blood pressure response during dialysis therapy. In **chapter 2** the role of differences in extracorporeal blood temperature in the vascular response during isolated ultrafiltration and combined ultrafiltration-hemodialysis was studied (**protocol A**). During combined ultrafiltration-hemodialysis measurements were performed at two different dialysate temperatures: 37.5 and 35.0°C. Isolated ultrafiltration was associated with cooling of blood during extracorporeal circulation (-1.7°C). In contrast, extracorporeal blood cooling was compensated by heat transfer from dialysate (37.5°C) to blood during combined ultrafiltration-hemodialysis which resulted in a significant increase of the arterial blood temperature after four hours treatment. At a dialysate temperature of 35.0°C extracorporeal blood temperatures were comparable between isolated ultrafiltration and combined ultrafiltration-hemodialysis. The hemodynamic data confirm that both arteriolar and venous reactivity are impaired during combined ultrafiltration-hemodialysis: in contrast to isolated ultrafiltration vascular resistance remained unchanged while venous tone even decreased, resulting in secondary tachycardia.

Compared to warm dialysis, venous tone was significantly higher while heart rate was significantly lower during cold dialysis while vascular resistance was not significantly different. Therefore, we conclude that impaired vascular reactivity during combined ultrafiltration-hemodialysis compared to isolated ultrafiltration is only partially explained on the basis of differences in extracorporeal blood temperature which suggests that additional factors should play a role. In **protocol B** blood pressure was measured during both warm and cold combined ultrafiltration-hemodialysis as well as during 22 post-dialysis hours to investigate whether cold dialysis is associated with rebound hypotension. Mean arterial blood pressure was significantly higher during cold dialysis compared to warm dialysis. However, in the post-dialytic period, blood pressure was comparable between warm and cold dialysis while none of the patients experienced rebound hypotension. Therefore, lowering the dialysate temperature seems to be a safe method to improve blood pressure stability without inducing post-dialytic hypotension.

Traditionally, a decrease in plasma osmolality is considered to play an important role in the impaired vascular response during combined ultrafiltration-hemodialysis. In this respect, it is well established that hemodynamic stability can be improved by the use of high dialysate sodium concentrations. Moreover, a relation has been suggested between hemodynamic instability during low-sodium dialysis and an increase in plasma prostaglandin E_2 . In **chapter 3** therefore, vascular reactivity as well as changes in plasma prostaglandin E_2 were measured during combined ultrafiltration-hemodialysis using high (144 mmol/l) and low (134 mmol/l) sodium dialysate. High-sodium dialysis resulted in a significantly higher plasma sodium level. Continuous measurements of relative blood volume confirmed results obtained in previous studies showing a significantly larger decrease in intravascular volume during low-sodium dialysis. Prostaglandin E_2 increased significantly only during low-sodium dialysis. However, there were no significant differences in vascular reactivity between high- and low-sodium dialysis. In addition, we found no relation between changes in prostaglandin E_2 and changes in arterial blood pressure or vascular tone. Consequently, better hemodynamic stability in high-sodium dialysis is only related to a better preservation of plasma volume and not to an improvement in vascular reactivity. Moreover, the data indicate that impaired vascular reactivity during combined ultrafiltration-hemodialysis is not related to changes in plasma osmolality.

Theoretically, hemodynamic instability during combined ultrafiltration-hemodialysis could be related to release of cytokines by monocytes in the presence of dialysate derived contaminants as suggested by Henderson and colleagues. Better hemodynamic stability during isolated ultrafiltration and hemofiltration would then result from the absence of dialysate and contaminants and the use of sterile substitution fluid in hemofiltration. In **chapter 4**, however, we found no

differences in vascular reactivity during combined ultrafiltration-hemodialysis with either non-sterile and sterile dialysate. Despite use of pyrogen free dialysate, combined ultrafiltration-hemodialysis was still associated with a decrease in venous tone while vascular resistance remained unchanged. Although we cannot exclude that high levels of dialysate contamination may further impair vascular reactivity during combined ultrafiltration-hemodialysis, our results show that disparity in vascular reactivity between different dialysis strategies is not primarily related to differences in pyrogenic exposition.

Since calcium ions play a pivotal role in the contraction of vascular smooth muscle cells and cardiac myocytes, changes in plasma ionized calcium might be of importance with respect to the hemodynamic response during dialysis therapy. As a matter of fact, a relation has been found in previous studies between changes in ionized calcium and the blood pressure response during dialysis therapy. In this respect, an increase in myocardial contractility with higher dialysate calcium concentrations has been a consistent finding. However, it is unknown whether vascular reactivity also depends on changes in ionized calcium. In **chapter 5** both vascular and myocardial reactivity were studied during combined ultrafiltration-hemodialysis with high (1.75 mmol/l) and low-calcium (1.25 mmol/l) dialysate. Plasma ionized calcium increased significantly only during high-calcium dialysis which resulted in a significantly higher blood pressure as compared to low-calcium dialysis. Although we found no significant differences in both arteriolar and venous reactivity between the two treatment modalities, cardiac output decreased significantly only during low-calcium dialysis. Consequently, while confirming the pivotal role of changes in plasma ionized calcium in the blood pressure response during dialysis therapy our data indicate that this relation is based on differences in myocardial contractility only.

According to the results of the first four studies, the extracorporeal blood temperature appears to be the most important factor with respect to the vascular response during dialysis therapy. In **chapter 6** therefore, vascular reactivity was measured in relation to the extracorporeal blood temperature comparing combined ultrafiltration-hemodialysis with hemofiltration. To obtain a comparable extracorporeal blood temperature in hemofiltration just as in combined ultrafiltration-hemodialysis (37.5°C), the infusate temperature was set at 39.0°C. In addition, measurements were also performed during hemofiltration at an infusate temperature of 36.0°C. Treatments were matched with respect to the buffer substrate (bicarbonate), small molecule clearance and biocompatibility. The results show that at equivalent extracorporeal blood temperatures, combined ultrafiltration-hemodialysis and hemofiltration are associated with a comparable hemodynamic response. Vascular resistance increased significantly only during cold hemofiltration while venous tone remained unchanged. In contrast, both warm

hemofiltration and combined ultrafiltration-hemodialysis were associated with a similar decrease in venous tone while vascular resistance remained unchanged. As a result, mean arterial blood pressure was significantly higher in cold hemofiltration compared to both combined ultrafiltration-hemodialysis and warm hemofiltration.

Concluding remarks

In the present thesis the pathophysiology of disparity in vascular reactivity between different dialysis strategies was investigated. In general it can be concluded that vascular reactivity during dialysis therapy is especially influenced by changes in the extracorporeal blood temperature. The results obtained in chapter 6 even indicate that compared to hemofiltration, impaired vascular reactivity during combined ultrafiltration-hemodialysis is primarily related to differences in extracorporeal blood temperature. The normal hemodynamic response to a decrease in circulating blood volume seems to be counteracted by an increase in core body temperature during combined ultrafiltration-hemodialysis, resulting in peripheral vasodilation. This suggests that it should be possible to obtain comparable hemodynamic stability in combined ultrafiltration-hemodialysis like in hemofiltration by lowering the dialysate temperature. To investigate this hypothesis, the blood pressure response and the incidence of hypotensive episodes should be studied in a prospective clinical study in cardiovascular compromised patients during hemofiltration and (cold) combined ultrafiltration-hemodialysis at equivalent extracorporeal blood temperatures. In addition, studies are necessary to investigate the possible adverse effects of cold dialysis.

Chapter 8

Samenvatting

Samenvatting

Hemodynamische instabiliteit met symptomatische hypotensie is nog steeds een van de belangrijkste dialyse gerelateerde complicaties. Met een toenemend aantal oudere, cardiovasculair gecompromitteerde dialyse patiënten is het van groot belang hemodialyse als therapie te optimaliseren met betrekking tot de intradialytische bloeddrukregulatie. Ten aanzien hiervan is het algemeen bekend dat het onttrekken van vocht beter wordt getolereerd door de patiënt wanneer dit plaats vindt tijdens geïsoleerde ultrafiltratie of hemofiltratie dan tijdens gecombineerde ultrafiltratie-hemodialyse. De laatste therapie wordt ook wel klassieke hemodialyse genoemd. In **hoofdstuk 1** wordt een overzicht van de literatuur gegeven betreffende plasma volume preservatie en cardiovasculaire compensatie mechanismen in relatie tot de verschillende vormen van nierfunctievervangende therapie. Beide mechanismen spelen een centrale rol in de bloeddrukregulatie tijdens hemodialyse. Geconcludeerd kan worden dat verschillen in hemodynamische stabiliteit tussen verschillende vormen van nierfunctievervangende therapie primair zijn gerelateerd aan verschillen in vasculaire reactiviteit. Zowel geïsoleerde ultrafiltratie als hemofiltratie zijn geassocieerd met een normale fysiologische toename in perifere vaattonus indien het circulerend bloedvolume daalt. In tegenstelling hiermee is zowel de arteriële als de veneuze reactiviteit gestoord tijdens gecombineerde ultrafiltratie-hemodialyse. Het is echter nog steeds onduidelijk welke mechanismen hiervoor verantwoordelijk zijn. In de huidige studies werd daarom de pathofysiologie van verschillen in vasculaire reactiviteit tijdens verschillende vormen van nierfunctievervangende therapie bestudeerd. Middels veneuze occlusie-plethysmografie werd de vasculaire reactiviteit van de onderarm gemeten bij cardiovasculair stabiele patiënten tijdens hemodialyse in relatie tot een aantal factoren.

Het is bekend dat de extracorporele bloedtemperatuur een belangrijke rol speelt in de bloeddrukregulatie tijdens hemodialyse. In **hoofdstuk 2** werd de rol van verschillen in de extracorporele bloedtemperatuur in de vasculaire reacties tijdens geïsoleerde ultrafiltratie en gecombineerde ultrafiltratie-hemodialyse bestudeerd (**protocol A**). Tijdens gecombineerde ultrafiltratie-hemodialyse werd gebruik gemaakt van twee verschillende dialysaattemperaturen. Tijdens geïsoleerde ultrafiltratie koelde het bloed tijdens extracorporele circulatie af met 1.7°C. Tijdens gecombineerde ultrafiltratie-hemodialyse met een dialysaattemperatuur van 37.5°C werd afkoeling van bloed voorkomen door verwarming van bloed door het warme dialysaat. Als gevolg hiervan nam de arteriële bloedtemperatuur tijdens 4 uur behandeling significant toe. Bij een dialysaattemperatuur van 35.0°C was de extracorporele bloedtemperatuur tijdens gecombineerde ultrafiltratie-hemodialyse exact gelijk aan geïsoleerde ultrafiltratie. De hemodynamische gegevens bevestigen

dat zowel de arteriële als veneuze reactiviteit gestoord is tijdens gecombineerde ultrafiltratie-hemodialyse: in tegenstelling tot geïsoleerde ultrafiltratie bleef de vasculaire weerstand onveranderd terwijl de veneuze tonus zelfs daalde, hetgeen resulteerde in een secundaire toename van de hartfrequentie. Tijdens koude gecombineerde ultrafiltratie-hemodialyse was alleen de veneuze tonus significant hoger en de hartfrequentie significant lager dan tijdens warme gecombineerde ultrafiltratie-hemodialyse. Dit impliceert dat de verminderde vasculaire reactiviteit tijdens gecombineerde ultrafiltratie-hemodialyse in vergelijking met geïsoleerde ultrafiltratie slechts ten dele wordt verklaard door verschillen in de extracorporele bloedtemperatuur. Dit suggereert dat additionele factoren hierin een rol moeten spelen. In **protocol B** werden de effecten van koude dialyse op de post-dialytische bloeddruk bestudeerd door zowel tijdens als 22 uur na warme en koude gecombineerde ultrafiltratie-hemodialyse de bloeddruk te meten. Koude hemodialyse was alleen tijdens de behandeling geassocieerd met een significant hogere bloeddruk. In de post-dialytische periode bestonden er geen verschillen in bloeddruk tussen warme en koude hemodialyse. Post-dialytische hypotensie na koude dialyse trad evenmin op. Het verlagen van de dialysaattemperatuur lijkt daarom een veilige methode om de hemodynamische stabiliteit tijdens gecombineerde ultrafiltratie-hemodialyse te doen toenemen.

Een daling in plasma osmolaliteit wordt traditioneel gezien als een belangrijke factor in de gestoorde hemodynamische reactiviteit tijdens gecombineerde ultrafiltratie-hemodialyse. Het is bekend dat de hemodynamische stabiliteit tijdens hemodialyse kan worden verbeterd door gebruik te maken van een hogere dialysaatsodiumconcentratie. Tevens wordt er een relatie gesuggereerd tussen hemodynamische instabiliteit bij gebruik van lage dialysaatsodiumconcentraties en een stijging van het plasma prostaglandine E_2 . In **hoofdstuk 3** werd daarom zowel de vasculaire reactiviteit als veranderingen in prostaglandine E_2 bestudeerd tijdens gecombineerde ultrafiltratie-hemodialyse bij zowel een hoge (144 mmol/l) als een lage (134 mmol/l) dialysaatsodiumconcentratie. Hoog natrium dialyse resulteerde in een significant hogere plasma natriumconcentratie. Continue meting van veranderingen in bloedvolume toonde een grotere afname in het intravasculaire volume tijdens laag natrium dialyse hetgeen resultaten van eerdere studies bevestigde. Prostaglandine E_2 nam alleen tijdens laag natrium dialyse significant toe. De vasculaire reactiviteit was echter volledig vergelijkbaar tussen hoog en laag natrium dialyse. Tevens werd geen relatie gevonden tussen veranderingen in plasma prostaglandine E_2 en veranderingen in arteriële bloeddruk of vasculaire tonus. Deze resultaten tonen aan dat de verbeterde hemodynamische stabiliteit bij gebruik van hogere dialysaatsodiumconcentraties wordt bepaald door een betere preservatie van plasma volume en niet door een verbetering van de vasculaire reactiviteit. Verder duiden zij erop dat de verminderde vasculaire reactiviteit tijdens gecombineerde

ultrafiltratie-hemodialyse niet is gerelateerd aan veranderingen in plasma osmolaliteit.

In aanwezigheid van niet steriel dialysaat zou hemodynamische instabiliteit tijdens gecombineerde ultrafiltratie-hemodialyse gerelateerd kunnen zijn aan de productie van cytokinen door monocysten zoals wordt gesuggereerd door Henderson en anderen. Betere hemodynamische stabiliteit tijdens geïsoleerde ultrafiltratie en hemofiltratie zou dan het gevolg zijn van de afwezigheid van dialysaat (en pyrogenen) en het gebruik van een steriel substitutievloeistof tijdens hemofiltratie. In **hoofdstuk 4** konden wij echter geen verschil aantonen in vasculaire reactiviteit tijdens gecombineerde ultrafiltratie-hemodialyse tussen gebruik van niet-steriel en steriel dialysaat. Gecombineerde ultrafiltratie-hemodialyse was nog steeds geassocieerd met een afname in veneuze tonus en een onveranderde perifere weerstand ondanks gebruik van dialysaat vrij van pyrogenen. Deze resultaten tonen aan dat verschillen in vasculaire reactiviteit tussen verschillende vormen van nierfunctievervangende therapie niet primair zijn gerelateerd aan verschillen in expositie aan pyrogenen. Het is echter niet uitgesloten dat expositie aan een hogere graad van pyrogenen zou kunnen leiden tot een verdere verstoring van de vasculaire reactiviteit tijdens gecombineerde ultrafiltratie-hemodialyse.

Aangezien calcium ionen een belangrijke rol spelen in de contractie van zowel vasculaire gladde spiercellen als van cardiale myocyten kunnen veranderingen in plasma geïoniseerd calcium van grote invloed zijn op de hemodynamische reactie tijdens hemodialyse. In eerdere studies werd reeds een relatie gevonden tussen veranderingen in geïoniseerd calcium en de bloeddruk tijdens hemodialyse. Een toename in myocardiale contractiliteit bij gebruik van hogere dialysaat calcium concentraties is hierbij meerdere malen aangetoond. Onbekend is echter of de perifere vasculaire reactiviteit tevens wordt beïnvloed door veranderingen in geïoniseerd calcium. In **hoofdstuk 5** werd daarom zowel de vasculaire als myocardiale reactiviteit gemeten tijdens gecombineerde ultrafiltratie-hemodialyse met een lage (1.25 mmol/l) en een hoge (1.75 mmol/l) dialysaatcalcium-concentratie. Het plasma geïoniseerd calcium nam alleen tijdens hoog calcium dialyse significant toe hetgeen resulteerde in een significant hogere bloeddruk. Terwijl de cardiale output alleen tijdens laag calcium dialyse significant afnam werden geen verschillen in perifere weerstand en veneuze tonus gevonden tussen laag en hoog calcium dialyse. Deze resultaten tonen aan dat de belangrijke rol die veranderingen in geïoniseerd calcium spelen in de bloeddrukregulatie tijdens dialyse primair zijn gerelateerd aan veranderingen in myocardiale contractiliteit.

Op basis van de resultaten van de eerste vier studies kan geconcludeerd worden dat de extracorporele bloedtemperatuur een belangrijke rol speelt in de vasculaire reactiviteit tijdens hemodialyse. In **hoofdstuk 6** werd daarom de vasculaire reactiviteit bestudeerd in relatie tot de extracorporele bloedtemperatuur tijdens

zowel gecombineerde ultrafiltratie-hemodialyse als hemofiltratie. Om een vergelijkbare bloedtemperatuur te verkrijgen tijdens hemofiltratie als tijdens gecombineerde ultrafiltratie-hemodialyse (37.5°C) werd de temperatuur van de substitutievloeistof verhoogd naar 39.0°C. Dezelfde metingen werden tevens verricht tijdens hemofiltratie met een temperatuur van de substitutievloeistof van 36.0°C. Beide behandelingen waren gematched met betrekking tot de gebruikte buffer (bicarbonaat), klein moleculaire klaring en biocompatibiliteit. Bij een exact gelijke extracorporele bloedtemperatuur waren gecombineerde ultrafiltratie-hemodialyse en hemofiltratie geassocieerd met een vergelijkbare hemodynamische reactiviteit. Alleen tijdens koude hemofiltratie bleef de veneuze tonus onveranderd terwijl de perifere weerstand significant toenam. Tijdens zowel warme hemofiltratie als gecombineerde ultrafiltratie-hemodialyse nam de veneuze tonus significant af terwijl de perifere weerstand onveranderd bleef. Een en ander resulteerde in een significant hogere bloeddruk tijdens koude hemofiltratie in vergelijking met zowel warme hemofiltratie als gecombineerde ultrafiltratie-hemodialyse.

Conclusie

In het voorliggende proefschrift werd de pathofysiologie van verschillen in vasculaire reactiviteit tijdens verschillende vormen van nierfunctie vervangende therapie bestudeerd. In het algemeen kan worden geconcludeerd dat de vasculaire reactiviteit tijdens dialyse met name wordt beïnvloed door de extracorporele bloedtemperatuur. De resultaten van hoofdstuk 6 laten zelfs zien dat verschillen in vasculaire reactiviteit tussen gecombineerde ultrafiltratie-hemodialyse en hemofiltratie primair worden bepaald door verschillen in de extracorporele bloedtemperatuur. Tijdens gecombineerde ultrafiltratie-hemodialyse wordt de normale fysiologische hemodynamische reactie op een afname in circulerend bloedvolume verstoord door een toename in lichaamstemperatuur, resulterend in perifere vasodilatatie. Dit suggereert dat het mogelijk moet zijn om een vergelijkbare hemodynamische stabiliteit te verkrijgen tijdens gecombineerde ultrafiltratie-hemodialyse als tijdens hemofiltratie door verlaging van de dialysaattemperatuur. Het is van belang om deze hypothese te bestuderen in een prospectieve meer klinische georiënteerde studie. Hierin zouden hemodynamische parameters en de incidentie van hypotensieve episodes gemeten moeten worden bij cardiovasculair gecompromitteerde patiënten tijdens hemofiltratie en (koude) gecombineerde ultrafiltratie-hemodialyse bij een vergelijkbare extracorporele bloedtemperatuur. Tevens lijkt het noodzakelijk om de mogelijk nadelige gevolgen van koude dialyse te bestuderen.

Dankwoord

Dankwoord

Het verrichten van klinisch-wetenschappelijk onderzoek is voor een belangrijk deel gebaseerd op een goede samenwerking met meerdere mensen. Zo heb ook ik de afgelopen jaren velerlei mensen mogen ontmoeten, waarvan eenieder op zijn of haar wijze aan de totstandkoming van dit proefschrift heeft bijgedragen. Zonder woord van dank is de afsluiting van drie jaar onderzoek niet compleet.

Door het praktisch-klinische karakter van de verschillende studies was de medewerking van patiënten onontbeerlijk. Met veel geduld ondergingen zij de verschillende metingen, waarbij 'stil zitten' een vaak uitgesproken verzoek mijnerzijds was. Ik wil hen graag hartelijk danken voor het getoonde vertrouwen en hun bereidwilligheid om aan de studies deel te nemen.

De meeste studies zijn uitgevoerd op de afdeling hemodialyse van het Academisch Ziekenhuis Maastricht. Deze afdeling fungeerde de afgelopen jaren voor mij als een soort thuisbasis. Ik kan terugkijken op een bijzonder prettige samenwerking met het voltallige personeel. Een dag op de dialyse begon meestal met een kopje koffie en een praatje met Marleen Dresens. Verder stond Pierre Waterval altijd open voor een nieuw verzoek mijnerzijds. Ook in tijden wanneer de ruimte binnen de afdeling bijzonder beperkt was. Voor vragen of problemen van technische aard kon ik altijd bij Piet Claessens terecht. De door eenieder getoonde interesse heb ik altijd bijzonder op prijs gesteld.

Dr. F. de Heer (Maasland ziekenhuis, Sittard) was bereid om mij een aantal van zijn patiënten voor het onderzoek toe te vertrouwen. Verder waren Dr. W. Grave en Dr. J.J.J.M. Wirtz (St.Laurentius ziekenhuis, Roermond) alsook Dr. P.G.G. Gerlag en Dr. A.W.L. van den Wall Bake (St.Joseph ziekenhuis, Veldhoven) en Dr. A.W. Mulder en Dr. S.J. Hoorntje (Catharina ziekenhuis, Eindhoven) bereid de deuren van hun afdeling voor mij te openen. In Veldhoven was de medewerking van de afdeling medische microbiologie onder leiding van Dr. M. Haverkorn van bijzonder belang. Dhr. A. van Dorst moet ooit gedacht hebben dat aan de stroom van te analyseren dialysaatmonsters geen einde meer zou komen. In Eindhoven werd het echocardiografisch onderzoek verricht door medewerkers van de afdeling Hartfunctie onder leiding van Dr. C.H. Peels. Het verrichten van onderzoek elders gaf aan de verschillende studies een extra dimensie. Dankzij de medewerking van de verschillende internist-nefrologen, vele verpleegkundigen, technici, laboranten en secretaresses was het mogelijk om in vaak korte tijd een schat aan informatie te verkrijgen. I also would like to thank Dr. D. Hillion and his co-workers (CHI Poissy, France) for their French hospitality and cooperation in performing the last study. Dr. Hillion, it was a great pleasure for me to share your experience with convective dialysis treatments.

In Maastricht werd het onderzoek vanuit verschillende hoeken ondersteund. Centraal in alle studies stond de plethysmograaf van Janssen Scientific Instruments. Luc van Bortel maakte de beschikking hierover zowel binnen als buitenshuis mogelijk. De medewerkers van het hypertensie laboratorium en het laboratorium klinische farmacologie stonden garant voor een nauwkeurige en snelle bepaling van catecholaminen. Het laboratorium voor klinische chemie verzorgde de vele klinisch chemische bepalingen waarbij

Ing. H.M.L. Habets een altijd coöperatief aanspreekpunt vormde. Dr. P.P.C.A. Menheere was bereid zich te verdiepen in de bepaling van prostaglandinen hetgeen geen sinecure bleek te zijn. Op het laboratorium van Dr. W.A. Buurman en in het bijzonder door Gaby Francot werden de immunologische bepalingen verricht. De uitvoering en verslaglegging waren altijd bijzonder accuraat. Dr. F.H.M. Nieman verzorgde in de beginfase de statistische bewerking van de verkregen gegevens en gaf mij in korte tijd een spoedcursus statistiek. Fred, het feit dat ik toch nog regelmatig bij je moest aankloppen geeft aan dat MANOVA voor mij nog steeds een aantal geheimen heeft. Bij de uiteindelijke afronding van mijn proefschrift waren zowel de positieve inzet als ervaring van Tiny Wouters voor mij van bijzondere betekenis. Om met Karel te spreken, bij Tiny brandt nog het heilig vuur.

Zonder uitzondering wil ik alle leden van de Werkgroep Nefrologie (klein maar fijn) graag bedanken voor de bijzonder prettige samenwerking van de afgelopen jaren. Wat mij betreft zullen onze paden elkaar de komende jaren nog meerdere malen kruisen.

The financial support of the Department of Research of Gambro AB, Sweden, coordinated by Dr. I. Ledebø, is also greatly acknowledged. Dear Ingrid, at many meetings and congresses we have been able to discuss the progress of our mutual project. Your critical and constructive comments have always been of great help.

Tevens wil ik Gambro Nederland danken voor hun logistieke ondersteuning bij in het bijzonder mijn exercitie in Veldhoven.

De leden van de beoordelingscommissie, Prof.Dr. P.J.E.H.M. Kitslaar, Prof.Dr. M.P. van Dieijen-Visser en Prof.Dr. H.A.J. Struyker Boudier wil ik danken voor de kritische beoordeling van mijn proefschrift, waarin een voor hen niet alledaags onderwerp centraal staat.

I also would like to thank Prof.Dr. C. Baldamus and Prof.Dr. J. Bergström for their willingness to critically review my manuscript. The present thesis is in fact based on their scientific work.

Mijn promotor, Prof.Dr. J.P. van Hooff wil ik graag danken voor zijn constructieve bijdrage aan de afronding van mijn proefschrift. Zijn intelligente en praktische benadering van problemen werkte vaak bijzonder verhelderend.

Een bijzonder woord van dank gaat uit naar mijn co-promotor, Dr. K.M.L. Leunissen. Beste Karel, ondanks een zich voor jou opstapelend aantal functies en taken heb ik de afgelopen jaren nooit vergeefs een beroep op je gedaan. Zowel je immer opbouwende kritiek als je heldere ideeën en enthousiasme waren voor mij van bijzondere betekenis en vormden een drijvende kracht achter de totstandkoming van dit proefschrift. Gelukkig beperkte jou begeleiding zich niet tot het onderzoek alleen. Met name dankzij jou was de achterliggende periode voor mij op velerlei gebied bijzonder leerzaam. Ik wil je hiervoor hartelijk danken.

Tot slot wil ik mijn ouders danken, voor de mogelijkheden die zij mij altijd hebben geboden. Bouwen begint bij een goed fundament.

Curriculum Vitae

Curriculum Vitae

Willibrord van Kuijk

16-11-1965	Geboren te Waalwijk
1978-1984	atheneum-β, Dr. Mollercollege te Waalwijk
1984-1991	studie Geneeskunde aan de Rijksuniversiteit Limburg te Maastricht
1991-1992	vervulling militaire dienstplicht als eerste luitenant arts, Koninklijke Militaire School te Weert
1992-1995	assistent-in-opleiding aan de Rijksuniversiteit Limburg te Maastricht (begeleider: Dr. K.M.L. Leunissen)
1995-1996	arts-assistent Intensive Care, Catharina Ziekenhuis te Eindhoven
01-10-1997	aanvang opleiding tot internist, Catharina Ziekenhuis te Eindhoven (opleider: Dr. W.P.M. Breed)