Reliable assessment of physical activity in disease: an update on activity monitors

Citation for published version (APA):

Westerterp, K. R. (2014). Reliable assessment of physical activity in disease: an update on activity monitors. *Current Opinion in Clinical Nutrition and Metabolic Care*, 17(5), 401-406.
<https://doi.org/10.1097/MCO.0000000000000080>

Document status and date:

Published: 01/09/2014

DOI:

[10.1097/MCO.0000000000000080](https://doi.org/10.1097/MCO.0000000000000080)

Document Version:

Publisher's PDF, also known as Version of record

Document license:

Taverne

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.umlib.nl/taverne-license

Take down policy

If you believe that this document breaches copyright please contact us at:

repository@maastrichtuniversity.nl

providing details and we will investigate your claim.

Reliable assessment of physical activity in disease: an update on activity monitors

Klaas R. Westerterp

Purpose of review

Physical activity is the main determinant of variability of individual energy expenditure. Reliable assessment of the activity level of an individual provides information on energy requirement and vital health. Here, evidence is presented on the validity of methods to assess physical activity as applied in health and disease.

Recent findings

Improvement of technology has resulted in a growing number of physical activity monitors to evaluate the activity level of an individual. Outcome measures vary from raw data on body acceleration and posture to activity levels derived from proprietary algorithms based on body acceleration combined with sensor information on additional activity-induced physiological responses. Data interpretation is limited by evidence-based studies on the value of measured parameters for the assessment of physical activity.

Summary

The optimal tool for reliable assessment of physical activity is an accelerometer providing a valid measure of body movement. Overall validity is derived from studies using doubly labelled water-assessed activity-induced energy as a reference. Thus, out of the large range of activity monitors, three came out with a better validity. The ultimate tool is a single unobtrusive device allowing valid and long-term monitoring of activity with regard to type and intensity.

Keywords

accelerometers, body movement, doubly labelled water, energy expenditure, physical activity

INTRODUCTION

Physical activity is an important determinant of health and well-being. On the contrary, physical activity is negatively affected by disease. Thus, assessment of physical activity is important for understanding relations between physical activity and health and for determining the effectiveness of interventions. As stated by a recent guide to the assessment of physical activity, 'deleterious health consequences of physical inactivity are vast, and they are of paramount clinical and research importance. Risk identification, benchmarks, efficacy, and evaluation of physical activity behaviour change initiatives for clinicians and researchers all require a clear understanding of how to assess physical activity' [1].

Methods for the assessment of physical activity have evolved from behavioural observation and self-report through interviews, questionnaires, and diaries, to physiological markers including heart rate, calorimetry, and motion sensors. The doubly labelled water method has become the gold standard for the validation of methods of assessing physical activity

behaviour [2]. Physical activity is defined as 'any bodily movements produced by skeletal muscles that result in energy expenditure' [3]. Doubly labelled water allows measuring total energy expenditure (TEE) in unrestrained individuals over a time interval of 1–4 weeks [4]. In combination with a measurement of resting energy expenditure (REE), activity-induced energy expenditure (AEE) can be calculated as $AEE = 0.9 \text{ TEE} - \text{REE}$. The calculation assumes that diet-induced energy expenditure is a constant fraction of 10% of TEE in individuals consuming a mixed diet and eating according to what they need [5]. Alternatively, physical activity can be calculated as

Department of Human Biology, Maastricht University Medical Centre, Maastricht, The Netherlands

Correspondence to Klaas R. Westerterp, Department of Human Biology, Maastricht University Medical Centre, PO Box 616, 6200 MD Maastricht, The Netherlands. Tel: 31 43 3881628; fax: 31 43 3670976; e-mail: k.westerterp@maastrichtuniversity.nl

Curr Opin Clin Nutr Metab Care 2014, 17:401–406

DOI:10.1097/MCO.0000000000000080

KEY POINTS

- Accelerometry is the most promising method to objectively assess physical activity behaviour as a reflection of metabolic health or determinant of metabolic risk.
- The outcome of activity monitors is best evaluated when signals are accessible as raw data.
- Overall validity of activity monitors is judged with doubly labelled water-assessed activity energy expenditure or PAI as a reference.
- Out of the large range of activity monitors, three came out with a better validity.
- The ultimate activity monitor is a single unobtrusive device allowing valid and long-term monitoring of activity with regard to type and intensity.

the physical activity index (PAI) = TEE/REE. The PAI is a figure to compare physical activity, as derived from energy expenditure, between individuals of different size [6]. Overall reliability of a method to assess physical activity can be judged with doubly labelled water-assessed AEE or PAI as a reference.

Validation studies of methods for the assessment of physical activity with doubly labelled water as a criterion method have shown accelerometry to be superior to self-report and heart rate monitoring. With self-report, individuals easily overestimate or underestimate activity intensity and the time spent in activities. Heart rate is affected by more factors than physical activity and data conversion to physical activity needs individual calibration against oxygen consumption. Thus, heart rate monitoring remains a proxy measure for physical activity.

Accelerometry is the most promising method to objectively assess physical activity behaviour as a reflection of metabolic health or determinant of metabolic risk. Critical output parameters include time inactive, moderately active, and vigorously active. The purpose of the current review is to evaluate recent evidence on the validity of monitors to assess physical activity behaviour as applied in health and disease.

ACTIVITY MONITORS: PRACTICAL ASPECTS AND OUTCOME PARAMETERS

There is a large range of activity monitors from pedometers and load transducers, for walking activity, to accelerometers and multiple sensor systems, to monitor total physical activity. A review on the validity of accelerometers and multiple sensor systems in health and disease as published from 2000 to 2012 already included 40 different devices [7]. Aspects presented here are wearing comfort including size and placement, and outcome including

data processing and storage capacity; validity is discussed in the next section.

Wearing comfort of a monitor for habitual physical activity should be such that it does not interfere with physical activity behaviour. One should forget about wearing a monitor. Modern technology allowed impressive miniaturization without loss of performance. We started in 1985 with an accelerometer weighing 350 g, including the unit for data storage, with the battery life of 7 days. Our current accelerometer weighs 12.5 g, has a battery life of 3 weeks, and can store data for up to 22 weeks [8]. Placement remains an issue. Theoretically, a monitor for the detection of total physical activity should be located near the centre of mass of the body since movements of the centre of mass reflect movements of the total body. Alternatively, monitors should be attached to the locations of the body where they would register the most activity, that is, the legs during walking and cycling. Thus, multiple accelerometers are used to classify activity modes. The ultimate monitor classifies activity modes with algorithms based on signals at one location, where sensors at more than one location reduce wearing comfort. Smart phones have a potential as activity monitor because people wear them everyday [9].

Outcome parameters of activity monitors are a function of specifications such as sampling frequency and dynamic range of the acceleration signal and of data transformation. Sampling frequency and dynamic range determine whether a monitor successfully filters accelerations from active movement and can be easily checked from the output during passive transport. Transformation of raw data is performed to increase storage capacity and thus observation time, and to translate the acceleration signal into physical activity outcomes such as activity time, type, and intensity including energy expenditure. Critical aspects are the inclusion of additional, not necessarily, activity-related parameters in the transformation process and in the use of proprietary equations. Outcome is best evaluated when signals are accessible as raw data [10]. Outcome should not be a function of proprietary equations where agreement between subsequent versions can be large [11].

VALIDATION

Activity monitors are designed to assess physical activity behaviour with a focus on time pattern and intensity. As stated in the introduction, overall validity is judged with doubly labelled water-assessed AEE or PAI as a reference. Then, the optimal observation time is 1 or 2 weeks. This is the optimal interval for valid doubly labelled water assessment of AEE or PAI. Observation for 1 or 2 weeks of free-

living activity is likely to include the full range of physical activity behaviour of an individual. Laboratory validations during standardized activities, mainly walking and running, are discouraged. The walking component of physical activity is better detectable than other activities of daily living and thus not representative for overall validity.

Two reviews evaluated the performance of 18 different activity monitors against doubly labelled water, as published until 2012, resulting in a large heterogeneity in outcome [7[■],10[■]]. Here, the main outcomes and conclusions are presented for validations of 11 monitors as published more recently (Table 1). Most monitors explained less than 50% of the variation in AEE or PAI. Only three out of the 11 tested, Actigraph GT3X, Dynaport, and TracmorD, came out better. Actiheart performed well in one study [15], whereas four other studies concluded that the performance was poor [13,14,16,17]. In only one study, the SenseWear armband explained more than 50% of the variation in AEE [22], whereas three other studies concluded poor performance as well.

APPLICATION IN HEALTH AND DISEASE

Of the three monitors selected above, Actigraph GT3X (ActiGraph, LLC., Pensacola, FL) is most widely

applied. In 2003, Actigraph 7164, an earlier uniaxial model, was already introduced in the 'National Health and Nutrition Examination Survey' in the USA. Aggregated data resulted in decision rules to facilitate conclusions on physical activity from accelerometer output [26]. The size of the Actigraph GT3X, in combination with a memory and battery capacity for 1-month data collection, facilitates longitudinal assessment of physical activity in relation to disease parameters [27[■]].

The Dynaport monitor (DynaPort MiniMod, McRoberts, The Hague, The Netherlands) is typically applied for the assessment of movement quality. In addition to overall physical activity, the monitor allows accurate assessment of postural stability and postural changes [28,29].

TracmorD (DirectLife, Philips Consumer Lifestyle, Amsterdam, The Netherlands) is applied in health and disease research and as a consumer product, in combination with a personal website and e-coach. Studies of obese individuals allowed the assessment of physical activity in relation to weight loss and weight maintenance [30,31]. A web-based intervention study showed the potential to increase physical activity, thus improving metabolic health [32[■]].

In free-living individuals, physical activity shows large variation within a day and between

Table 1. Summary of doubly labelled water evaluation studies of activity monitors as published since 2012, with number of individuals (*n*), explained variation of activity-induced energy expenditure (*R*²), and main conclusion

Monitor	<i>n</i>	<i>R</i> ²	Conclusion
Actigraph GT3X	39	70	Better-to-best performance out of six monitors tested [12 [■]]
Actiheart	18	17	Poor performance in adolescents [13]
	12	41	Not valid and considerable equipment failure [14]
	35	70	Good performance when individually calibrated [15]
	19	42	Underestimation of physical activity [16]
	75	0–6	No or little explanation of variation in physical activity [17]
Actiwatch	40	18	Moderate performance out of six monitors tested [12 [■]]
DynaPort	40	53	Better-to-best performance out of six monitors tested [12 [■]]
3DNX	155	6	Moderate explanation of variance in physical activity [18]
	49	18	Similar performance to uniaxial accelerometer [19]
EW4800P	16	20	No precise measurement of physical activity [20]
IDEEA	14	36	Capable of identifying activity, wiring limits use in daily life [21]
Lifecorder	40	10	Poor performance out of six monitors tested [12 [■]]
RT3	35	22	Reasonable in lean and poor performance in obese individuals [15]
	39	7	Poor performance out of six monitors tested [12 [■]]
SWA	73	8	Poor performance out of six monitors tested [12 [■]]
	19	38	Underestimation of physical activity [16]
	28	60	Considerable individual bias [22]
	62	–	Validity limited by large individual variation [11]
	14	16	No significant association with physical activity [23]
TracmorD	30	50	Moderate-to-strong validity in children [24]
	36	58	Valid in overweight and obese individuals [25]

FIGURE 1. Physical activity of a typical individual monitored for 1 year with TracmorD, integrated over intervals of (a) month, (b) week, (c) day, and (d) hour.

days. The variation is illustrated by activity data of a typical individual, as monitored with TracmorD and integrated over intervals of a month, week, day, and hour (Fig. 1). Thus, accelerometer output is usually evaluated over intervals of at least one or more weeks.

DISCUSSION

Out of the large range of activity monitors, three came out with a better validity as evaluated under free-living conditions with doubly labelled water-derived activity energy expenditure as a reference. The final choice of a valid monitor depends on further specifications including memory and battery capacity, additional options for data processing

including movement quality, and options for user feedback and coaching.

Activity monitors with better validity explained 50–70% of the variation in AEE or PAI, leaving at least 30% of the variation unexplained. Further improvement could be reached with monitors allowing the assessment of activity type. The relationship between activity counts and energy expenditure varies according to activity type [33]. Identification of activity types improved the assessment of energy expenditure compared with activity counts only [34,35].

Interpreting physical activity outcomes based on activity monitors requires adjustment for body size when comparing individuals of different size. Obese individuals are more sedentary than

individuals of normal weight while AEE might be similar [36]. Excess weight increases energy cost of movement and decreases movement economy. A weight correction factor will be higher for ambulatory activities and lower for sedentary activities. Activity energy expenditure is commonly corrected for differences in body weight by dividing AEE (J/min) by the square root of body weight [35]. A standard way of comparing outcomes of an activity monitor between individuals of different size is by converting the outcomes into a PAI value.

Activity monitors are often validated under laboratory conditions during standardized activities, mainly walking and running. The walking component of physical activity is better detectable than other activities of daily living and thus validity during walking is not representative for overall validity. Additionally, in relation to health, light-intensity activities might be as important as high-intensity activities. Light-intensity activity breaks decrease the well known risk of thrombosis associated with excessive sitting [37]. To detect uninterrupted bouts of inactivity, an activity monitor should detect minimal body movement to prevent mistaking uninterrupted bouts of inactivity for nonwear time [38].

CONCLUSION

Accelerometry is the most promising method to objectively assess physical activity behaviour as a reflection of metabolic health or determinant of metabolic risk. The validity of accelerometer-based activity monitors is judged with doubly labelled water-assessed activity energy expenditure or PAI as a reference, where the optimal observation time is 1 or 2 weeks. The outcome of an activity monitor should not be a function of proprietary equations where agreement between subsequent versions can be large. Most monitors explained less than 50% of the variation in activity energy expenditure or PAI. Only three out of the 11 tested, Actigraph GT3X, Dynaport, and TracmorD, came out better. The final choice of a valid monitor depends on further specifications including memory and battery capacity, additional options for data processing including movement quality, and options for user feedback and coaching.

Acknowledgements

None.

Conflicts of interest

There are no conflicts of interest.

REFERENCES AND RECOMMENDED READING

Papers of particular interest, published within the period of review, have been highlighted as:

- of special interest
- of outstanding interest

1. Strath SJ, Kaminsky LA, Ainsworth BE, *et al.* Guide to the assessment of physical activity: clinical and research applications; a scientific statement from the American Heart Association. *Circulation* 2013; 128:2259–2279.
- A recent review stating the rationale, key concepts, and options for the assessment of physical activity in a clinical setting.
2. Bonomi AG, Westerterp KR. Advances in physical activity monitoring and lifestyle interventions in obesity: a review. *Int J Obes* 2012; 36:167–177.
3. Butte NF, Ekelund U, Westerterp KR. Assessing physical activity using wearable monitors: measures of physical activity. *Med Sci Sports Exerc* 2012; 44:S5–S12.
4. Westerterp KR. Control of energy expenditure in humans. In: Weickert MA, editor. *Obesity*. Endotext.org; 2013.
5. Westerterp KR. Physical activity and physical activity induced energy expenditure in humans: measurement, determinants and effects. *Frontiers Physiol* 2013; 4:90.
6. FAO/WHO/UNU. Human energy requirements. FAO food and nutrition report series 1. Rome, Italy; 2004.
7. Van Remoortel H, Giavedoni S, Raste Y, *et al.* Validity of activity monitors in health and chronic disease: a systematic review. *Int J Beh Nutr Phys Act* 2012; 9:84.
- A systematic review identifying available activity monitors that have been validated in healthy adults and patients with chronic disease.
8. Bonomi AG, Plasqui G, Goris AH, Westerterp KR. Estimation of free-living energy expenditure using a novel activity monitor designed to minimize obtrusiveness. *Obesity* 2010; 18:1845–1851.
9. Antos SA, Albert MV, Kording KP. Hand, belt, pocket or bag: physical activity tracking with mobile phones. *J Neurosci Methods* 2013. [Epub ahead of print]. doi: 10.1016/j.jneumeth.2013.09.015.
10. Plasqui G, Bonomi AG, Westerterp KR. Daily physical activity assessment with accelerometers: new insights and validation studies. *Obes Rev* 2013; 14:451–462.
- A review on the validity of accelerometers using doubly labelled water as the reference.
11. Slinde F, Bertz F, Winkfist L, *et al.* Energy expenditure by multisensory armband in overweight and obese lactating women validated by doubly labelled water. *Obesity* 2013; 21:2231–2235.
12. Rabinovich RA, Louvaris Z, Raste Y, *et al.* Validity of physical activity monitors during daily life in patients with COPD. *Eur Resp J* 2013; 42:1205–1215.
- A comparative evaluation of the validity and usability of six activity monitors in patients with chronic obstructive pulmonary disease.
13. Campbell N, Prapavessis H, Gray C, *et al.* The actiheart in adolescents: a doubly labelled water validation. *Pediatr Exerc Sci* 2012; 24:589–602.
14. Santos DA, Silva AM, Matias CN, *et al.* Validity of combined heart rate and motion sensor for the measurement of free-living energy expenditure in very active individuals. *J Sci Med Sport* 2014; 17:387–393.
15. Villars C, Bergouignan A, Dugas J, *et al.* Validity of combining heart rate and uniaxial acceleration to measure free-living physical activity energy expenditure in young men. *J Appl Physiol* 2012; 113:1763–1771.
16. Farooqi N, Slinde F, Häglin L, Sandström T. Validation of sensewear armband and actiheart monitors for assessment of daily energy expenditure in free living women with chronic obstructive pulmonary disease. *Physiol Rep* 2013; 1:e00150.
17. Henriksson H, Forsum E, Löf M. Evaluation of Actiheart and a 7 day activity diary for estimating free-living total and activity energy expenditure using criterion methods in 1.5- and 3-year-old children. *Br J Nutr* 2014; 111:1830–1840.
18. Horner F, Bilzon JL, Rayson M, *et al.* Development of an accelerometer-based multivariate model to predict free-living energy expenditure in a large military cohort. *J Sports Sci* 2013; 31:354–360.
19. Ojiambo R, Konstabel K, Veidebaum T, *et al.* Validity of hip-mounted uniaxial accelerometry with heart-rate monitoring vs triaxial accelerometry in the assessment of free-living energy expenditure in young children: the IDEFICS validation study. *J Appl Physiol* 2012; 113:1520–1536.
20. Ishikawa-Takata K, Kaneko K, Koizumi K, Ito C. Comparison of physical activity energy expenditure in Japanese adolescents by EW4800P triaxial accelerometry and the doubly labelled water method. *Br J Nutr* 2013; 110:1347–1355.
21. Wybrow S, Ritz P, Horgan GW, Stubbs RJ. An evaluation of the IDEEA activity monitor for estimating energy expenditure. *Br J Nutr* 2013; 109:173–183.
22. Calabré MA, Stewart JM, Welk GJ. Validation of pattern-recognition monitors in children using doubly labeled water. *Med Sci Sports Exerc* 2013; 45:1313–1322.
23. Tanhoffer RA, Tanhoffer AIP, Raymond J, *et al.* Comparison of methods to assess energy expenditure and physical activity in people with spinal cord injury. *J Spinal Cord Med* 2012; 35:35–45.

24. Sijtsma A, Schierbeek H, Goris AH, *et al.* Validation of TracmorD triaxial accelerometer to assess physical activity in preschool children. *Obesity* 2013; 21:1877–1883.
25. Valenti G, Camps S, Verhoef S, *et al.* Validating measures of free-living physical activity in overweight and obese subjects using an accelerometer. *Int J Obes* 2013. [Epub ahead of print]. doi: 10.1038/ijo.2013.195.
26. Tudor-Locke C, Camhi SM, Troiano RP. A catalog of rules, variables, and definitions applied to accelerometer data in the national health and nutrition examination survey, 2003–2006. *Prev Chronic Dis* 2012; 9: E113.
27. Ehsan M, Khan R, Wakefield D, *et al.* A longitudinal study evaluating the effect of exacerbations on physical activity in patients with chronic obstructive pulmonary disease. *Ann Am Thorac Soc* 2013; 10:559–564.
- A longitudinal study with a valid activity monitor, showing a reduction of physical activity early during exacerbation and persisting for some weeks after recovery.
28. Kapoula Z, Gaertner C, Yang Q, *et al.* Vergence and standing balance in subjects with idiopathic bilateral loss of vestibular function. *PLoS One* 2013; 8:e66652.
29. Van Lummel RC, Ainsworth E, Lindemann U, *et al.* Automated approach for quantifying the repeated sit-to-stand using one body fixed sensor in young and older adults. *Gait Posture* 2013; 38:153–156.
30. Bonomi AG, Soenen S, Goris AHC, Westerterp KR. Weight loss induced changes in physical activity and activity energy expenditure in obese subjects before and after energy restriction. *PLoS One* 2013; 8: e59642.
31. Camps SG, Verhoef SP, Westerterp KR. Weight loss-induced reduction in physical activity recovers during weight maintenance. *Am J Clin Nutr* 2013; 98:917–923.
32. Wijsman CA, Westendorp RGJ, Verhagen EALM, *et al.* Effects of a web-based intervention on physical activity and metabolism in older adults: randomized controlled trial. *J Med Internet Res* 2013; 15:e233.
- One of the first studies showing the effectivity of a web-based physical activity intervention, based on daily physical activity as measured with a valid activity monitor.
33. Crouter SE, Churilla JR, Bassett Jr DR. Estimating energy expenditure using accelerometers. *Eur J Appl Physiol* 2006; 98:601–612.
34. Bonomi AG, Plasqui G, Goris AHC, Westerterp KR. Improving the assessment of daily energy expenditure by identifying types of physical activity using a single accelerometer. *J Appl Physiol* 2009; 107:655–661.
35. Van Hees VT, van Lummel RC, Westerterp KR. Estimating activity-related energy expenditure under sedentary conditions using a tri-axial seismic accelerometer. *Obesity* 2009; 17:1287–1292.
36. Ekelund U, Aman J, Yngve A, *et al.* Physical activity but not energy expenditure is reduced in obese adolescents: a case-control study. *Am J Clin Nutr* 2002; 76:935–941.
37. Howard BJ, Fraser SF, Sethi P, *et al.* Impact on hemostatic parameters of interrupting sitting with intermittent activity. *Med Sci Sports Exerc* 2013; 45:1285–1291.
38. Lee I-M, Shiroma EJ. Using accelerometers to measure physical activity in large-scale epidemiological studies: issues and challenges. *Br J Sports Med* 2014; 48:197–201.