

Advanced methods for quality assurance and dose distribution improvement for high dose rate brachytherapy

Citation for published version (APA):

Bellezzo, M. (2020). *Advanced methods for quality assurance and dose distribution improvement for high dose rate brachytherapy*. Gildeprint en Universitaire Pers Maastricht. <https://doi.org/10.26481/dis.20200611mb>

Document status and date:

Published: 01/01/2020

DOI:

[10.26481/dis.20200611mb](https://doi.org/10.26481/dis.20200611mb)

Document Version:

Publisher's PDF, also known as Version of record

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.umlib.nl/taverne-license

Take down policy

If you believe that this document breaches copyright please contact us at:

repository@maastrichtuniversity.nl

providing details and we will investigate your claim.

Summary

Brachytherapy is a method of radiotherapy that consists on placing a radioactive source in close contact or inside the tumor, resulting in a steep gradient on the dose distribution, which helps protecting organs at risk but also makes the technique prone to errors in case of wrong positioning of the radioactive source. There are several stages involved in a treatment workflow, from afterloader commissioning to treatment delivery and verification. In this thesis, several stages of a brachytherapy treatment were evaluated as part of an effort to improve accuracy and precision of dose delivery.

The viability to 3D tracking a brachytherapy source using an imaging panel was shown possible by applying a 2D gaussian fit to the panel response to locate the source and the distance between the source and the panel could be determined considering the gaussian shape. The same principle was adopted in a quality assurance system for brachytherapy applicators that uses an ^{192}Ir source to acquire a projection of the applicator into the imaging panel and then sends the source inside the applicator to track its movement with submillimeter precision.

The main focus of this thesis, however, was to develop an applicator for rectal cancer capable of generating a dose distribution similar to those generated by 50 kVp x-ray devices, but using HDR ^{192}Ir sources. The applicator was successfully modeled by using the inverse square law to achieve a steep dose falloff by bringing the ^{192}Ir source at several points in close contact to the treatment surface.

Finally, a new afterloading system was thoroughly evaluated using multiple quality assurance devices, high speed camera and imaging panel to verify current limitations in brachytherapy dose delivery due to precision and accuracy of dwell times, dwell positions and transit dose effects.

Keywords: brachytherapy, Monte Carlo, dosimetry, commissioning