

Individual optimisation of contrast media application and radiation dose in computed tomographic angiography

Citation for published version (APA):

Kok, M. (2016). *Individual optimisation of contrast media application and radiation dose in computed tomographic angiography: from phantom to patient*. [Doctoral Thesis, Maastricht University]. Maastricht University. <https://doi.org/10.26481/dis.20161118mk>

Document status and date:

Published: 01/01/2016

DOI:

[10.26481/dis.20161118mk](https://doi.org/10.26481/dis.20161118mk)

Document Version:

Publisher's PDF, also known as Version of record

Please check the document version of this publication:

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

www.umlib.nl/taverne-license

Take down policy

If you believe that this document breaches copyright please contact us at:

repository@maastrichtuniversity.nl

providing details and we will investigate your claim.

9.1

Summary

This thesis addressed several topics on individual optimisation of contrast media (CM) volume and radiation dose for different CT angiography (CTA) examinations. The overall focus of this research was to maintain diagnostic image quality in CTA while using iodinated CM and radiation dose in a more efficient way for the individual patient.

In **Chapter 2** the relationship between CM concentration and viscosity was described. The results show that the viscosity of CM lower with lower concentrations of iodine (240mg/ml and 300mg/ml) was significantly lower compared to CM with higher concentrations of iodine (370mg/ml and 400mg/ml). Furthermore, pre-heating CM to body temperature drastically decreased the viscosity of CM. In addition, the influence of different viscosity levels was investigated in circulation phantom, where CM was injection at different temperatures (20 – 37°C). From these experiments, we could conclude that the injection pressure significantly decreased while using lower concentrated CM and, moreover, when injecting these concentrations at body temperature. Therefore, standardised pre-heating should be a prerequisite for clinical CM administration.

In **Chapter 3** the potential for automated kVp-selection to reduce the radiation dose during CTA using different concentrated CM – normalised to an identical iodine delivery rate (IDR; [g/s]) – was explored using a circulation phantom. The results showed that automated tube current modulation can result in radiation dose reduction up to 53% and automated tube voltage selection up to 77%, if human subjects approximate the dimensions and therefore the attenuation values of the phantom used in the experiment. Furthermore, Dose reduction was – as expected – independent of the CM concentration applied at normalised IDR.

In **Chapter 4** the influence of scan parameters – especially tube voltage [kVp] settings – on intravascular enhancement in coronary CTA (CCTA) was investigated. A circulation phantom was used to systematically investigate how IDR and CM volume could be adapted to a particular kVp setting in order to remain optimal enhancement levels in the coronary arteries. These primary results were then also tested in sixty patients referred for CCTA. In a circulation phantom, lower kVp settings allowed a substantial reduction in CM volume - up to 56% at 70kVp - whilst maintaining diagnostically sufficient attenuation within the target vessels. Furthermore, initial results in patients confirmed this finding at 100kVp using 12% reduction of CM volume in CCTA.

In **Chapter 5** the effect on image quality when using body weight adapted injections was studied in patients referred for CT pulmonary angiography (CTPA). The use of individualised CM protocols provided diagnostic and robust enhancement in emergency CTPA, as well as a substantial CM volume reduction in lower weight patients compared to a fixed CM protocol. Moreover, greater consistency of vascular enhancement

values – indicating a more reliable protocol – was observed throughout the patients who received body weight adapted injections, whereas the scans for patients who received fixed injections showed a steady decline in attenuation with increasing body weight.

In **Chapter 6** the role of individual adaptation of both injection and scan protocols based on patient's body mass index (BMI) was elaborated in patients referred for pre-TAVI CT examination. As these patients frequently suffer from an impaired renal function, the possibility of reducing CM volume while using lower kVp settings was investigated. The use of low kVp (80kVp and 70kVp) scan protocols allowed for substantial reduction in CM volume as compared to common high volume injection protocols for pre-TAVI CT examinations. Sufficient image quality was maintained for the evaluation of the aortic root and peripheral access site despite a CM volume reduction of 34-67%.

In **Chapter 7** the possibility of reducing radiation dose exposure and CM volume was studied in patients referred for the evaluation of the aorta, as aortic CTA is associated with frequent follow-up over time, which implies high cumulative radiation dose and iodinated CM volume for these patients. It turned out that individualised CTA protocols based on the BMI of the patient are favourable for CTA of the whole aorta in routine clinical practice. Iterative reconstruction algorithms resulted in 23-57% less radiation in combination with 55-63% less contrast volume when compared to standard CT protocols.